

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052815\
 Data File : BM001434.D
 Acq On : 28 May 2015 19:04
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 29 03:40:43 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM052815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 28 15:41:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2 S	2-Fluorophenol	1.000	1.024	-2.4	98	0.00
3 S	Phenol-d6	1.000	1.018	-1.8	100	0.00
4	bis(2-Chloroethyl)ether	1.000	1.089	-8.9	103	0.00
5 P	n-Nitroso-di-n-propylamine	1.000	0.985	1.5	98	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	100	0.00
7 S	Nitrobenzene-d5	1.000	0.998	0.2	100	0.00
8	bis(2-Chloroethoxy)methane	1.000	0.944	5.6	99	0.01
9 I	Acenaphthene-d10	5.000	5.000	0.0	102	0.00
10 S	2,4,6-Tribromophenol	1.000	0.934	6.6	101	0.01
11 S	2-Fluorobiphenyl	1.000	1.023	-2.3	102	0.00
12	2,6-Dinitrotoluene	1.000	0.915	8.5	105	0.00
13	2,4-Dinitrotoluene	1.000	0.880	12.0	105	0.00
14	4-Chlorophenyl-phenylether	1.000	1.026	-2.6	100	0.00
15 I	Phenanthrene-d10	5.000	5.000	0.0	107	0.01
16	4-Bromophenyl-phenylether	1.000	0.975	2.5	101	0.01
17 I	Chrysene-d12	5.000	5.000	0.0	105	0.00
18 S	Terphenyl-d14	1.000	1.035	-3.5	107	0.00
19	Benzo(a)anthracene	1.000	0.976	2.4	104	0.00
20	3,3'-Dichlorobenzidine	1.000	0.925	7.5	102	0.00
21 I	Perylene-d12	5.000	5.000	0.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0