

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM052815\
 Data File : BM001441.D
 Acq On : 29 May 2015 00:30
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: May 29 04:15:06 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM052815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 03:45:53 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	0.00
2 S	2-Fluorophenol	1.000	1.035	-3.5	107	0.00
3 S	Phenol-d6	1.000	1.089	-8.9	116	-0.03
4	bis(2-Chloroethyl)ether	1.000	0.989	1.1	101	0.00
5 P	n-Nitroso-di-n-propylamine	1.000	1.075	-7.5	117	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	107	0.00
7 S	Nitrobenzene-d5	1.000	1.008	-0.8	108	0.00
8	bis(2-Chloroethoxy)methane	1.000	1.013	-1.3	115	0.00
9 I	Acenaphthene-d10	5.000	5.000	0.0	106	0.00
10 S	2,4,6-Tribromophenol	1.000	1.004	-0.4	114	0.00
11 S	2-Fluorobiphenyl	1.000	1.033	-3.3	106	0.00
12	2,6-Dinitrotoluene	1.000	0.940	6.0	112	0.00
13	2,4-Dinitrotoluene	1.000	0.909	9.1	113	0.00
14	4-Chlorophenyl-phenylether	1.000	1.028	-2.8	104	0.00
15 I	Phenanthrene-d10	5.000	5.000	0.0	105	0.01
16	4-Bromophenyl-phenylether	1.000	1.037	-3.7	105	0.00
17 I	Chrysene-d12	5.000	5.000	0.0	103	0.00
18 S	Terphenyl-d14	1.000	1.031	-3.1	105	0.00
19	Benzo(a)anthracene	1.000	1.039	-3.9	109	-0.02
20	3,3'-Dichlorobenzidine	1.000	1.083	-8.3	125	-0.02
21 I	Perylene-d12	5.000	5.000	0.0	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0