

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052820\
 Data File : BM026155.D
 Acq On : 28 May 2020 12:26
 Operator : MA/SJ
 Sample : PB129135BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK35

Quant Time: May 28 13:06:56 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 28 11:12:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	120977	20.00	ng/ul	0.00
18) Naphthalene-d8	10.24	136	492076	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.13	164	284245	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.87	188	626515	20.00	ng/ul	0.00
78) Chrysene-d12	21.07	240	678882	20.00	ng/ul	0.00
86) Perylene-d12	23.19	264	735807	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.06	96	25682	6.40	ng/uL	0.00
5) Phenol-d5	6.68	99	413516	37.68	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.83	67	277772	36.59	ng/ul	0.00
9) 2-Chlorophenol-d4	7.02	132	315967	36.78	ng/ul	0.00
13) 4-Methylphenol-d8	8.20	113	314006	37.97	ng/ul	0.00
19) Nitrobenzene-d5	8.62	128	144255	35.96	ng/ul	0.00
22) 2-Nitrophenol-d4	9.34	143	143249	36.10	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.88	165	280307	35.18	ng/ul	0.00
29) 4-Chloroaniline-d4	10.39	131	395297	48.92	ng/ul	0.00
44) Dimethylphthalate-d6	13.55	166	789437	34.63	ng/ul	0.00
47) Acenaphthylene-d8	13.82	160	946422	35.05	ng/ul	0.00
52) 4-Nitrophenol-d4	14.35	143	117965	29.22	ng/ul	0.00
58) Fluorene-d10	15.13	176	693723	34.99	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.26	200	73565	20.14	ng/ul	0.00
71) Anthracene-d10	16.97	188	1066695	34.55	ng/ul	0.00
79) Pyrene-d10	19.27	212	1209821	34.69	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.06	264	1382628	35.30	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM052820\
Data File : BM026155.D
Acq On : 28 May 2020 12:26
Operator : MA/SJ
Sample : PB129135BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK35

Quant Time: May 28 13:06:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM051320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 28 11:12:05 2020
Response via : Initial Calibration

