

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052824\
 Data File : BM045772.D
 Acq On : 28 May 2024 22:15
 Operator : MA/JU
 Sample : SSTD3.244
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2006

Quant Time: May 28 23:52:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM052824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 28 23:49:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.485	152	6924	0.400	ng/ul	0.00
4) Naphthalene-d8	10.255	136	20949	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.123	164	11043	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.871	188	21878	0.400	ng/ul	0.00
17) Chrysene-d12	21.058	240	14558	0.400	ng/ul	0.00
23) Perylene-d12	23.174	264	15856	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.063	96	27685	3.177	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.092	88	27855	3.335	ng/ul#	79
14) Pentachlorophenol	16.563	266	4346	0.904	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052824\
Data File : BM045772.D
Acq On : 28 May 2024 22:15
Operator : MA/JU
Sample : SSTD3.244
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2006

Quant Time: May 28 23:52:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM052824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 28 23:49:11 2024
Response via : Initial Calibration

