




Data Path : Z:\HPCHEM1\BNA M\DATA\BM053017\
Data File : BM010313.D
Acq On : 30 May 2017 16:00
Operator : SJ/MA
Sample : I3162-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Quant Time: May 31 06:01:58 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM053017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 31 06:00:47 2017
Response via : Initial Calibration



TIC: BM010313.D

(20) Perylene-d12 (I)

23.663min (-0.153) 0.40ng/ul

response 111542

Ion	Exp%	Act%
264.00	100	100
260.00	31.30	23.58#
265.00	132.10	23.24#
0.00	0.00	0.00