


Data Path : Z:\HPCHEM1\BNA_M\Data\BM053017\
 Data File : BM010301.D
 Acq On : 30 May 2017 08:28
 Operator : SJ/MA
 Sample : DFTPP02
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP02

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM053017.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 31 08:43:19 2017



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.3	237461	PASS
68	69	0.00	2	2.0	6883	PASS
69	198	0.00	100	38.9	351489	PASS
70	69	0.00	2	0.5	1655	PASS
127	198	10	80	39.8	360128	PASS
197	198	0.00	2	0.6	5498	PASS
198	198	100	100	100.0	904512	PASS
199	198	5	9	6.6	60050	PASS
275	198	10	60	30.4	274944	PASS
365	198	1	100	5.3	47725	PASS
441	443	0.01	100	86.8	98240	PASS
442	198	50	100	67.9	614400	PASS
443	442	15	24	18.4	113173	PASS

m/z	Abundance
37.00	2401.0
38.00	6214.0
39.10	17528.0
40.10	1515.0
41.30	749.0
42.90	629.0
43.20	598.0
43.80	310.0
45.00	1577.0
49.20	3219.0
50.10	94752.0
51.10	305152.0
52.10	14506.0
53.10	1135.0
55.10	2691.0
56.10	11198.0
57.00	21480.0
58.00	1074.0
58.70	786.0
59.80	628.0
61.00	4365.0
62.20	6497.0
63.10	16536.0
63.90	2154.0
65.10	4794.0
66.00	744.0
67.00	853.0
68.00	7959.0
69.00	438656.0
70.00	1804.0
71.20	490.0
73.10	10036.0
74.10	59560.0
75.00	89664.0
76.10	32744.0
77.10	521024.0
78.10	31600.0
79.10	45240.0
80.00	29800.0
81.00	37896.0
82.10	10465.0
83.00	6321.0
84.10	854.0
85.00	8205.0
86.00	13142.0
87.10	6074.0
87.90	1456.0
89.00	1173.0
91.00	6801.0
92.00	10610.0
93.00	88464.0
94.00	5552.0
94.90	436.0
96.00	1933.0
97.10	1175.0
98.10	70864.0
99.00	48400.0
100.00	3248.0
101.10	17512.0
102.20	1434.0
103.00	6972.0
104.00	15610.0
105.10	14661.0
106.10	4464.0
107.10	168448.0
108.10	22592.0
109.00	4509.0
110.00	240384.0
111.10	39800.0
112.10	5739.0
113.20	2408.0
114.80	1039.0
116.10	10580.0
117.10	239552.0
118.00	16456.0
119.00	1504.0
120.10	2717.0
120.90	706.0
122.00	11004.0
123.00	18912.0
124.00	8774.0
125.10	7409.0
126.20	2683.0
127.10	421056.0
128.10	32912.0
129.10	201344.0
130.10	18048.0
131.10	5356.0
132.10	931.0
133.10	1760.0
134.00	8841.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

135.00	19840.0
136.00	7973.0
137.10	11711.0
138.10	1879.0
138.90	1106.0
140.00	2663.0
141.00	32768.0
142.00	8810.0
143.10	8069.0
143.90	1853.0
144.90	1722.0
146.20	4980.0
147.00	19488.0
148.00	52128.0
149.00	9156.0
150.10	2642.0
151.00	4139.0
152.10	1843.0
153.10	10158.0
154.10	6605.0
155.10	19616.0
156.10	24528.0
157.00	3477.0
158.00	7857.0
159.00	5079.0
160.00	8924.0
161.00	14564.0
162.10	4470.0
162.80	1016.0
163.90	2202.0
165.10	12488.0
166.10	12380.0
167.10	86880.0
168.10	44248.0
169.00	7413.0
170.00	3510.0
171.10	2327.0
172.00	5819.0
173.10	7538.0
174.10	15244.0
175.10	23656.0
176.10	5111.0
177.10	14728.0
178.00	4158.0
179.00	52328.0
180.00	35072.0
181.10	14435.0
182.10	2447.0
182.90	1238.0
184.00	3317.0
185.10	26592.0
186.10	169792.0
187.10	47344.0
188.00	5898.0
189.00	14675.0
189.90	2229.0
190.20	2162.0
191.00	7064.0
192.00	17480.0
193.10	19960.0
194.10	3251.0
195.10	1860.0
196.00	27256.0
197.10	5874.0
198.00	1026560.0
199.00	64608.0
200.10	7707.0
201.60	4951.0
203.00	11552.0
204.00	50464.0
205.10	81904.0
206.10	322112.0
207.10	44896.0
208.00	14938.0
209.00	7238.0
210.20	5010.0
211.10	16568.0
212.10	2120.0
213.00	707.0
213.90	1047.0
215.00	4989.0
216.00	10710.0
217.00	95224.0
218.00	12620.0
219.00	1160.0
219.80	1606.0
221.00	36256.0
222.00	11430.0
223.00	20304.0
224.10	187520.0
225.10	48088.0
226.10	7019.0
227.10	95200.0
228.10	14473.0
229.10	19584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

230.00	2927.0
231.10	5598.0
232.00	1747.0
233.00	1979.0
234.10	4950.0
235.10	5510.0
236.00	5072.0
237.00	7134.0
237.90	1262.0
239.00	5382.0
240.00	3446.0
241.00	4660.0
242.00	10724.0
243.10	10625.0
244.10	137600.0
245.10	21448.0
246.00	36144.0
247.00	6303.0
248.00	1495.0
249.10	5377.0
250.00	865.0
251.10	1343.0
252.00	2392.0
253.00	3708.0
254.20	4403.0
255.10	697664.0
256.10	105456.0
257.10	9912.0
258.00	56648.0
259.10	10919.0
260.10	1574.0
261.00	1718.0
264.00	1342.0
265.00	23208.0
266.10	3062.0
267.20	1008.0
267.90	568.0
269.10	367.0
270.00	881.0
271.10	1766.0
272.10	2260.0
273.10	22472.0
274.00	54560.0
275.10	277120.0
276.10	40712.0
277.00	30672.0
278.10	8805.0
279.00	2082.0
281.10	3004.0
282.00	2404.0
283.10	5837.0
284.00	2241.0
285.10	5322.0
286.00	1249.0
288.10	386.0
288.80	635.0
289.30	742.0
290.10	1143.0
291.10	1035.0
292.10	1107.0
293.00	7275.0
294.10	1443.0
295.00	1664.0
296.00	114384.0
297.10	14183.0
298.10	1416.0
299.00	400.0
301.00	1541.0
302.10	1649.0
303.10	9408.0
304.10	4050.0
304.90	328.0
305.40	356.0
306.10	665.0
307.10	747.0
308.10	1309.0
309.10	971.0
310.20	1300.0
311.10	583.0
313.00	467.0
314.00	4362.0
315.10	12914.0
316.10	4236.0
317.00	2030.0
321.10	2036.0
323.10	25936.0
324.20	4080.0
325.00	379.0
326.00	1282.0
327.00	6398.0
328.00	2929.0
328.90	506.0
330.00	410.0
332.10	3105.0
333.20	2258.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

334.10	16592.0
335.10	3375.0
336.20	1099.0
341.10	3834.0
342.00	746.0
345.20	311.0
345.90	5612.0
347.10	1653.0
347.90	397.0
349.90	362.0
351.10	945.0
352.10	7065.0
353.10	5719.0
354.20	5923.0
355.00	2490.0
356.30	558.0
358.80	621.0
359.20	441.0
364.00	697.0
365.10	44544.0
366.10	6995.0
367.20	358.0
370.20	778.0
371.20	1434.0
372.10	12491.0
373.10	2421.0
374.10	803.0
376.90	385.0
383.20	3364.0
384.20	1026.0
385.00	352.0
386.00	359.0
387.10	486.0
390.10	1299.0
391.10	1094.0
392.10	1074.0
393.60	302.0
398.60	411.0
401.00	918.0
402.00	3603.0
402.90	5343.0
404.20	1575.0
405.00	898.0
414.80	749.0
421.10	4405.0
422.10	4910.0
423.10	30448.0
424.10	10727.0
425.20	2209.0
437.70	325.0
439.30	944.0
439.70	1028.0
441.10	79800.0
442.10	536000.0
443.10	99536.0
444.00	8707.0
445.00	1125.0
446.90	374.0
460.00	367.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
36.10	577.0
37.10	2207.0
38.10	4216.0
39.10	16560.0
40.00	1085.0
41.10	590.0
43.00	1682.0
44.00	410.0
45.00	1345.0
48.20	600.0
49.20	2852.0
50.10	88448.0
51.10	272320.0
52.10	12197.0
55.10	2783.0
56.00	9074.0
57.10	16392.0
58.00	1433.0
58.70	504.0
60.00	897.0
61.20	5157.0
62.00	5029.0
63.00	16440.0
64.20	2298.0
65.10	6943.0
66.00	560.0
66.70	471.0
67.20	1110.0
68.10	8836.0
69.00	423744.0
70.10	2863.0
71.10	1227.0
72.00	372.0
73.10	8253.0
74.10	59264.0
75.10	75784.0
76.10	26928.0
77.10	483520.0
78.10	35240.0
79.10	44128.0
80.10	28624.0
81.10	40048.0
82.00	8675.0
83.00	7070.0
84.00	1891.0
85.00	6507.0
86.10	11327.0
87.10	3570.0
87.90	903.0
88.20	889.0
89.30	431.0
91.00	8858.0
92.10	11453.0
93.10	84080.0
94.10	4424.0
95.10	1251.0
96.00	1872.0
97.20	2003.0
98.00	69920.0
99.00	41704.0
100.00	4437.0
101.00	20608.0
101.90	938.0
103.00	5765.0
104.10	12400.0
105.00	14566.0
106.10	4817.0
107.10	161920.0
108.10	23440.0
109.00	3708.0
110.00	242688.0
111.10	41800.0
112.10	3246.0
113.00	2708.0
113.70	514.0
115.10	1278.0
116.10	8963.0
117.10	239296.0
118.10	17200.0
119.00	2685.0
120.10	3777.0
121.00	1150.0
122.00	12177.0
123.10	16608.0
124.00	8202.0
125.10	6521.0
126.00	1487.0
127.10	442112.0
128.10	33584.0
129.00	215296.0
130.10	17512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

131.00	4070.0
132.00	2012.0
133.00	1436.0
134.10	6979.0
135.00	23384.0
136.10	7735.0
137.10	10240.0
138.20	2859.0
139.00	2352.0
140.10	1782.0
141.10	33120.0
142.10	10422.0
143.10	6059.0
144.20	2178.0
145.20	982.0
146.00	8096.0
147.10	19752.0
148.00	57048.0
149.10	11973.0
150.10	2101.0
151.20	4795.0
152.10	2154.0
153.10	9993.0
154.10	8006.0
155.10	16656.0
156.10	24960.0
156.90	3471.0
158.00	9121.0
159.00	3474.0
160.00	12842.0
161.10	16448.0
162.10	4072.0
163.00	1375.0
164.00	1720.0
164.90	12324.0
166.00	9880.0
167.10	84752.0
168.10	43808.0
169.00	5726.0
170.00	2479.0
171.00	3974.0
172.10	6906.0
173.10	8882.0
174.00	14503.0
175.00	28448.0
176.10	6763.0
177.10	12631.0
178.00	4902.0
179.00	55056.0
180.10	34280.0
181.10	15782.0
182.10	3096.0
182.90	1323.0
184.00	5347.0
185.10	25480.0
186.10	187136.0
187.10	55832.0
188.10	5855.0
189.00	17072.0
190.00	2822.0
191.00	5236.0
192.10	21888.0
193.10	26072.0
194.10	3937.0
195.20	1416.0
196.00	29688.0
196.90	6133.0
198.00	1102848.0
199.00	75184.0
200.10	7008.0
201.50	2665.0
202.10	2195.0
203.00	14963.0
204.10	54856.0
205.10	89168.0
206.10	351168.0
207.10	54216.0
208.00	17912.0
209.00	6818.0
210.10	5291.0
211.10	15979.0
212.80	1152.0
213.30	582.0
213.90	762.0
215.10	5208.0
216.00	10791.0
217.00	111128.0
218.10	14107.0
219.20	1318.0
220.00	1367.0
221.10	36216.0
221.90	15984.0
223.10	24752.0
224.10	210560.0
225.10	53384.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

226.00	6999.0
227.10	112536.0
228.00	17056.0
229.10	24080.0
230.10	3015.0
231.10	8520.0
232.20	1231.0
233.10	2208.0
234.00	8090.0
235.10	6559.0
236.00	5482.0
237.00	9584.0
238.10	1961.0
239.00	5032.0
240.10	4270.0
241.10	6428.0
242.00	12113.0
243.10	13142.0
244.10	158080.0
245.10	20176.0
246.00	45240.0
247.00	8471.0
248.00	2234.0
249.10	7198.0
250.10	925.0
251.10	1933.0
252.00	2838.0
253.00	3137.0
254.10	7679.0
255.10	816832.0
256.10	118784.0
257.10	10077.0
258.00	65616.0
259.10	11111.0
260.20	976.0
261.10	1132.0
262.00	548.0
262.90	531.0
264.10	1118.0
265.10	26536.0
266.00	4483.0
266.90	1058.0
268.00	1014.0
269.00	501.0
270.20	1156.0
271.00	2432.0
272.00	1920.0
273.00	24920.0
274.10	71016.0
275.10	348352.0
276.00	42560.0
277.00	40920.0
278.10	7880.0
278.90	888.0
281.20	1972.0
282.10	1812.0
283.10	4652.0
284.10	2231.0
285.10	7491.0
286.00	1147.0
287.30	334.0
288.10	603.0
289.00	1122.0
289.80	897.0
291.10	1007.0
291.90	1098.0
293.10	7544.0
294.10	2132.0
295.10	3619.0
296.00	144384.0
297.00	17904.0
298.20	1035.0
301.10	2218.0
302.00	2022.0
303.10	12943.0
304.10	3391.0
305.00	646.0
308.00	1555.0
309.00	2000.0
310.10	1357.0
311.00	481.0
312.90	601.0
314.10	6436.0
315.10	16720.0
316.00	7324.0
316.90	1416.0
320.10	566.0
321.20	3387.0
321.90	1970.0
323.20	30848.0
324.20	6537.0
326.00	1064.0
327.10	6144.0
328.10	3797.0
328.80	637.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

329.20	669.0
331.10	645.0
332.20	2027.0
333.00	3753.0
334.10	23488.0
335.10	6986.0
336.00	1053.0
337.70	367.0
338.90	465.0
339.30	400.0
340.30	467.0
341.10	5098.0
342.00	1202.0
343.30	549.0
344.80	310.0
346.10	8520.0
347.00	1476.0
348.00	673.0
349.90	498.0
351.10	1195.0
352.10	10085.0
353.00	7004.0
354.20	8622.0
355.10	3402.0
356.20	648.0
357.00	451.0
358.10	774.0
359.20	624.0
363.20	342.0
363.70	360.0
364.20	533.0
365.10	57808.0
366.10	9225.0
367.00	1275.0
369.00	669.0
370.20	871.0
371.10	3298.0
372.10	15105.0
373.10	4059.0
374.20	565.0
377.00	507.0
383.10	4762.0
383.90	1742.0
385.00	735.0
387.20	428.0
390.10	2409.0
390.90	1859.0
391.90	1789.0
393.10	629.0
396.40	506.0
397.00	714.0
400.80	1269.0
402.10	6247.0
403.10	6403.0
404.20	2302.0
405.10	510.0
415.10	727.0
419.70	303.0
421.00	6296.0
422.20	8256.0
423.10	36488.0
424.10	9979.0
425.10	1462.0
426.20	610.0
429.30	336.0
431.40	352.0
437.80	643.0
438.40	759.0
439.60	1030.0
441.10	127304.0
442.10	768064.0
443.10	142016.0
444.10	14115.0
445.00	696.0
476.80	496.0
532.00	474.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
37.10	985.0
38.20	2574.0
39.10	7421.0
40.00	472.0
40.80	486.0
41.30	327.0
43.10	1385.0
44.10	435.0
45.20	555.0
49.10	1558.0
50.10	41440.0
51.10	134912.0
52.10	5698.0
52.90	641.0
55.10	1326.0
56.10	4141.0
57.10	7440.0
57.90	630.0
58.90	518.0
59.80	504.0
60.90	2179.0
62.10	3024.0
63.00	5337.0
64.00	1301.0
65.10	2976.0
66.80	899.0
68.00	3854.0
69.00	194048.0
70.00	1510.0
71.20	543.0
73.10	7201.0
74.10	29712.0
75.10	37832.0
76.10	13743.0
77.10	239040.0
78.20	15002.0
79.10	20480.0
80.00	13151.0
81.10	20208.0
82.10	3944.0
83.10	3849.0
84.00	584.0
85.10	4361.0
86.10	7632.0
87.10	2691.0
87.90	791.0
89.00	812.0
91.00	2664.0
92.10	5561.0
93.00	41152.0
94.10	2490.0
94.90	871.0
96.20	1977.0
97.00	1217.0
98.00	35336.0
99.10	20112.0
100.20	1672.0
101.00	9152.0
102.00	703.0
103.00	3342.0
104.00	7780.0
105.00	7418.0
106.10	2679.0
107.10	78200.0
108.00	11138.0
109.10	2678.0
110.00	114936.0
111.00	22416.0
112.20	2437.0
113.00	1201.0
114.80	332.0
115.30	632.0
116.10	6332.0
117.10	122272.0
118.10	7765.0
119.20	443.0
120.20	1771.0
121.20	846.0
122.10	4300.0
123.10	9480.0
124.00	5899.0
125.00	4963.0
126.20	2318.0
127.10	217216.0
128.10	17192.0
129.00	102440.0
130.10	8841.0
131.00	2090.0
131.90	1081.0
133.00	990.0
134.10	4092.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

135.10	10615.0
136.10	4726.0
137.10	3978.0
137.80	610.0
139.30	452.0
139.90	2336.0
141.10	16728.0
142.00	4865.0
143.00	3331.0
143.80	1080.0
144.70	495.0
146.00	3831.0
147.10	9306.0
148.00	29840.0
149.00	6133.0
150.00	1971.0
151.10	2532.0
151.80	1739.0
153.10	4892.0
154.10	2961.0
155.00	11216.0
156.00	12449.0
157.20	2148.0
158.00	3439.0
159.10	2812.0
160.10	6920.0
161.00	6701.0
162.00	2001.0
163.10	926.0
164.10	894.0
165.00	5769.0
165.80	3870.0
167.10	46728.0
168.10	26168.0
169.10	2422.0
170.00	1185.0
170.80	2516.0
172.10	3114.0
173.00	3725.0
174.10	7966.0
175.10	15423.0
176.00	2759.0
177.10	9594.0
178.00	2201.0
179.00	28456.0
180.10	18088.0
181.10	9098.0
182.00	1364.0
183.20	674.0
184.10	2604.0
185.00	14843.0
186.10	104760.0
187.00	26168.0
188.10	2891.0
189.00	8681.0
189.90	2607.0
191.00	4029.0
192.00	10108.0
193.10	10862.0
194.10	3145.0
194.80	911.0
195.20	633.0
196.00	14598.0
197.10	4489.0
198.00	584128.0
199.00	40360.0
200.00	3694.0
201.40	1832.0
202.30	884.0
203.10	6147.0
204.00	32216.0
205.00	48696.0
206.10	186944.0
207.10	26024.0
208.00	11297.0
209.10	4830.0
210.20	3959.0
211.00	9381.0
212.10	2002.0
213.10	1228.0
214.20	738.0
215.00	3559.0
216.10	6528.0
217.00	57384.0
218.10	10394.0
219.00	1350.0
219.90	1327.0
221.00	17680.0
222.00	8137.0
223.10	14397.0
224.10	115360.0
225.10	30928.0
226.10	3446.0
227.00	62016.0
228.10	7657.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

229.00	12292.0
230.10	1073.0
231.00	3605.0
232.00	542.0
233.00	872.0
234.00	4530.0
235.10	4602.0
236.10	4569.0
237.00	5403.0
238.10	1510.0
239.00	2487.0
240.10	2020.0
241.10	2441.0
242.10	6688.0
243.10	7286.0
244.10	91880.0
245.00	12680.0
246.00	24888.0
247.00	5313.0
248.00	1073.0
249.10	2482.0
250.30	367.0
251.10	815.0
252.00	712.0
252.20	708.0
253.10	2877.0
254.10	5572.0
255.10	483264.0
256.00	67344.0
256.90	5037.0
258.10	36512.0
259.10	6495.0
259.90	413.0
260.30	746.0
260.90	833.0
264.10	1236.0
265.10	16233.0
266.00	2864.0
267.10	1075.0
268.00	1334.0
270.20	819.0
271.00	2213.0
271.90	1680.0
273.00	15351.0
274.10	38656.0
275.10	199360.0
276.10	21824.0
277.00	25936.0
278.00	4569.0
278.90	1379.0
281.10	3001.0
282.00	1561.0
283.10	2140.0
284.10	1620.0
285.10	3347.0
286.10	586.0
288.10	300.0
289.10	2630.0
289.80	421.0
290.30	617.0
290.90	493.0
291.90	766.0
293.00	5442.0
294.20	1411.0
295.00	2333.0
296.00	91136.0
297.00	12752.0
298.00	507.0
301.10	1464.0
302.20	1663.0
303.10	7769.0
304.10	3337.0
304.90	646.0
306.00	331.0
307.90	1250.0
309.00	1168.0
310.00	862.0
310.90	832.0
313.30	528.0
314.10	3647.0
315.00	10027.0
316.20	3746.0
317.00	1196.0
320.20	306.0
321.20	1959.0
321.80	1025.0
323.10	20624.0
324.00	4523.0
325.10	561.0
326.00	855.0
327.00	5561.0
328.10	1776.0
329.10	469.0
330.00	430.0
332.00	1486.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

333.10	3565.0
334.10	15051.0
335.10	3368.0
336.00	607.0
339.20	986.0
340.10	646.0
341.20	2548.0
343.20	877.0
344.10	422.0
346.10	6903.0
347.00	1364.0
351.00	402.0
352.00	5203.0
353.20	3980.0
354.10	5955.0
355.10	2518.0
357.50	666.0
359.20	476.0
359.80	336.0
360.90	778.0
363.40	390.0
364.10	492.0
365.10	40824.0
366.00	4471.0
366.90	304.0
367.30	416.0
370.20	1233.0
371.20	1220.0
372.10	9941.0
373.00	2557.0
373.90	309.0
379.20	658.0
383.20	3208.0
384.00	797.0
385.00	780.0
389.90	1786.0
391.20	738.0
392.00	591.0
398.00	338.0
401.10	970.0
402.00	3749.0
403.00	7192.0
404.20	2111.0
405.30	309.0
410.20	605.0
415.10	531.0
421.10	5682.0
422.20	3037.0
423.10	33296.0
424.10	7968.0
425.00	1183.0
426.20	304.0
431.30	388.0
434.60	345.0
435.00	377.0
436.70	474.0
437.50	322.0
438.50	593.0
439.00	697.0
440.10	1511.0
441.10	87616.0
442.10	539136.0
443.10	97968.0
444.20	10805.0
444.90	354.0
490.10	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02

m/z	Abundance
60.00	392.0
69.10	660.0
69.90	404.0
70.40	397.0
70.90	783.0
73.00	3938.0
75.20	582.0
77.00	865.0
80.20	424.0
81.20	602.0
82.20	452.0
83.10	925.0
83.80	463.0
85.00	517.0
90.90	324.0
95.00	310.0
95.80	363.0
102.90	433.0
105.20	471.0
106.60	331.0
117.20	339.0
119.10	798.0
120.90	473.0
132.90	782.0
133.40	423.0
135.00	1269.0
137.00	387.0
145.00	451.0
147.00	1806.0
148.10	470.0
149.00	589.0
152.00	339.0
152.90	405.0
162.20	304.0
177.20	497.0
179.00	348.0
190.90	735.0
193.10	2156.0
207.20	3931.0
208.00	1878.0
209.10	823.0
209.70	327.0
210.20	310.0
213.10	349.0
221.00	820.0
222.10	513.0
223.00	357.0
235.60	354.0
249.20	464.0
252.90	505.0
267.10	1001.0
269.20	417.0
281.20	2231.0
282.10	783.0
283.00	1310.0
284.90	313.0
312.40	355.0
324.90	729.0
330.40	331.0
340.90	394.0
342.00	400.0
355.00	1132.0
429.00	424.0

Instrument :
BNA_M
ClientSampleId :
DFTPP02