

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM053115\
 Data File : BM001492.D
 Acq On : 31 May 2015 02:42
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: May 31 05:13:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM053115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun May 31 05:09:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	13582	5.00	ng	0.00
6) Naphthalene-d8	10.15	136	51688	5.00	ng	0.00
9) Acenaphthene-d10	14.07	164	26770	5.00	ng	0.00
15) Phenanthrene-d10	16.81	188	53615	5.00	ng	0.00
17) Chrysene-d12	21.02	240	45268	5.00	ng	0.00
21) Perylene-d12	23.11	264	39560	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 2-Fluorophenol	4.97	112	13164	5.84	ng	0.00
3) Phenol-d6	6.57	99	17803	7.40	ng	0.00
7) Nitrobenzene-d5	8.53	82	14402	5.15	ng	0.00
10) 2,4,6-Tribromophenol	15.57	330	5255	6.32	ng	0.00
11) 2-Fluorobiphenyl	12.67	172	38615	5.11	ng	0.00
18) Terphenyl-d14	19.47	244	28701	5.07	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) bis(2-Chloroethyl)ether	6.80	93	13196	5.24	ng	88
5) n-Nitroso-di-n-propylamine	8.18	70	11510	7.70	ng	# 95
8) bis(2-Chloroethoxy)methane	9.58	93	17790	6.32	ng	96
12) 2,6-Dinitrotoluene	13.65	165	7905	6.90	ng	90
13) 2,4-Dinitrotoluene	14.45	165	9764	7.38	ng	# 71
14) 4-Chlorophenyl-phenylether	15.11	204	19119	4.74	ng	99
16) 4-Bromophenyl-phenylether	16.02	248	11246	5.35	ng	99
19) Benzo(a)anthracene	21.01	228	54616	19.19	ng	97
20) 3,3'-Dichlorobenzidine	20.95	252	20535	6.83	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM053115\
Data File : BM001492.D
Acq On : 31 May 2015 02:42
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Quant Time: May 31 05:13:37 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM053115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun May 31 05:09:58 2015
Response via : Initial Calibration

