

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM053115\
 Data File : BM001499.D
 Acq On : 31 May 2015 06:54
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM053115

Quant Time: May 31 07:47:28 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM053115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun May 31 07:43:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2 S	2-Fluorophenol	0.984	0.985	-0.1	101	0.00
3 S	Phenol-d6	1.284	1.358	-5.8	103	0.00
4	bis(2-Chloroethyl)ether	0.990	1.074	-8.5	104	0.00
5 P	n-Nitroso-di-n-propylamine	0.809	0.754	6.8	100	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
7 S	Nitrobenzene-d5	0.270	0.258	4.4	106	-0.03
8	bis(2-Chloroethoxy)methane	0.323	0.316	2.2	102	0.00
9 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.01
10 S	2,4,6-Tribromophenol	0.177	0.162	8.5	99	0.00
11 S	2-Fluorobiphenyl	1.449	1.438	0.8	103	0.00
12	2,6-Dinitrotoluene	0.237	0.219	7.6	100	0.00
13	2,4-Dinitrotoluene	0.270	0.240	11.1	100	0.00
14	4-Chlorophenyl-phenylether	0.726	0.719	1.0	106	0.00
15 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
16	4-Bromophenyl-phenylether	0.208	0.201	3.4	102	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
18 S	Terphenyl-d14	0.669	0.652	2.5	102	0.00
19	Benzo(a)anthracene	1.363	1.360	0.2	109	-0.01
20	3,3'-Dichlorobenzidine	0.445	0.430	3.4	108	0.00
21 I	Perylene-d12	1.000	1.000	0.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0