

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM053115\
 Data File : BM001507.D
 Acq On : 31 May 2015 11:44
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 01 05:57:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM053115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun May 31 07:43:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2 S	2-Fluorophenol	1.000	0.989	1.1	92	0.00
3 S	Phenol-d6	1.000	1.011	-1.1	90	0.00
4	bis(2-Chloroethyl)ether	1.000	1.000	0.0	88	0.00
5 P	n-Nitroso-di-n-propylamine	1.000	0.919	8.1	91	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	92	0.00
7 S	Nitrobenzene-d5	1.000	0.955	4.5	94	0.00
8	bis(2-Chloroethoxy)methane	1.000	0.986	1.4	91	0.00
9 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
10 S	2,4,6-Tribromophenol	1.000	0.914	8.6	89	-0.01
11 S	2-Fluorobiphenyl	1.000	0.997	0.3	94	0.00
12	2,6-Dinitrotoluene	1.000	0.888	11.2	91	0.00
13	2,4-Dinitrotoluene	1.000	0.866	13.4	91	0.00
14	4-Chlorophenyl-phenylether	1.000	1.070	-7.0	103	0.00
15 I	Phenanthrene-d10	5.000	5.000	0.0	96	0.00
16	4-Bromophenyl-phenylether	1.000	0.963	3.7	95	-0.01
17 I	Chrysene-d12	5.000	5.000	0.0	97	-0.01
18 S	Terphenyl-d14	1.000	0.994	0.6	96	0.00
19	Benzo(a)anthracene	1.000	0.979	2.1	99	-0.01
20	3,3'-Dichlorobenzidine	1.000	1.003	-0.3	104	0.00
21 I	Perylene-d12	5.000	5.000	0.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0