

Quantitation Report (Qedit)




Data Path : Z:\HPCHEM1\BNA_M\Data\BM053117\
 Data File : BM010344.D
 Acq On : 01 Jun 2017 00:20
 Operator : SJ/MA
 Sample : I3164-08
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jun 01 02:12:43 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 01 01:28:02 2017
 Response via : Initial Calibration

Instrument :
 BNA_M
 ClientSampleId :
 JHSL4

Manual Integrations
 APPROVED

mohammad
 6/1/2017 3:01:47 PM



(29) 4-Chloroaniline-d4 (S)

10.893min (+0.000) 1.53ng/ul m

response 38583

Ion	Exp%	Act%
131.00	100	100
133.00	31.40	31.78
69.00	24.30	15.80#
0.00	0.00	0.00

SJ 06/02/17

Data Path : Z:\HPCHEM1\BNA_M\Data\BM053117\
 Data File : BM010344.D
 Acq On : 01 Jun 2017 00:20
 Operator : SJ/MA
 Sample : I3164-08
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 JHSL4

Quant Time: Jun 01 02:12:43 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 01 01:28:02 2017
 Response via : Initial Calibration

Manual Integrations
 APPROVED

mohammad
 6/1/2017 3:01:47 PM

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	412971	20.00	ng/ul	0.00
18) Naphthalene-d8	10.73	136	1496360	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.56	164	888529	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.30	188	1977357	20.00	ng/ul	0.00
75) Chrysene-d12	21.46	240	2559328	20.00	ng/ul	0.00
83) Perylene-d12	23.82	264	2625823	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.39	96	32774	3.26	ng/uL	0.00
5) Phenol-d5	7.10	99	575567	18.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.26	67	328001	18.65	ng/ul	0.00
9) 2-Chlorophenol-d4	7.46	132	484251	19.41	ng/ul	0.00
13) 4-Methylphenol-d8	8.65	113	472064	18.70	ng/ul	0.00
19) Nitrobenzene-d5	9.11	128	219348	20.01	ng/ul	0.00
22) 2-Nitrophenol-d4	9.82	143	263615	20.76	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.35	165	510510	19.57	ng/ul	0.00
29) 4-Chloroaniline-d4	10.89	131	38583m-)	1.53	ng/ul	0.00
43) Dimethylphthalate-d6	13.97	166	1607719	21.78	ng/ul	0.00
46) Acenaphthylene-d8	14.26	160	1798567	20.88	ng/ul	0.00
51) 4-Nitrophenol-d4	14.80	143	149761	13.56	ng/ul	0.00
57) Fluorene-d10	15.55	176	1330456	20.50	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.67	200	245460	17.53	ng/ul	0.00
70) Anthracene-d10	17.40	188	2001189	20.80	ng/ul	0.00
76) Pyrene-d10	19.69	212	2542778	24.03	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.66	264	2797974	22.89	ng/ul	0.00

SS 86/02/17

Target Compounds

					Qvalue
6) Phenol	7.13	94	131003	4.083 ng/ul	96
44) Dimethylphthalate	14.02	163	392763	5.410 ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM053117\
Data File : BM010344.D
Acq On : 01 Jun 2017 00:20
Operator : SJ/MA
Sample : I3164-08
Misc :
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jun 01 02:12:43 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 01 01:28:02 2017
Response via : Initial Calibration

Instrument :


BNA_M

ClientSampleId :

JHSL4

Manual Integrations
APPROVED

mohammad
6/1/2017 3:01:47 PM



Quantitation Report (Qedit)




Data Path : Z:\HPCHEM1\BNA_M\Data\BM053117\
 Data File : BM010344.D
 Acq On : 01 Jun 2017 00:20
 Operator : SJ/MA
 Sample : I3164-08
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Quant Time: Jun 09 15:10:04 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 09 15:09:36 2017
 Response via : Initial Calibration

Instrument :
 BNA_M
 ClientSampleId :
 JHSL4

Manual Integrations
 APPROVED

mohammad
 6/1/2017 3:01:47 PM



TIC: BM010344.D

(29) 4-Chloroaniline-d4 (S)
 11.022min (+0.142) 0.02ng/ul
 response 543

Ion	Exp%	Act%
131.00	100	100
133.00	31.40	37.39
69.00	24.30	20.18
0.00	0.00	0.00