

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM053123\
 Data File : BM040094.D
 Acq On : 01 Jun 2023 00:08
 Operator : CG/JU
 Sample : 02961-10
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0FQ2

Quant Time: Jun 01 00:43:36 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM053123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 31 18:20:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	6332	0.400	ng/ul	0.00
4) Naphthalene-d8	10.810	136	22851	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.631	164	11558	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.373	188	20792	0.400	ng/ul	0.00
17) Chrysene-d12	21.546	240	10606	0.400	ng/ul	0.00
23) Perylene-d12	23.980	264	10838	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.371	96	37546	4.824	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.393	152	11867	0.427	ng/ul	0.00
18) Fluoranthene-d10	19.397	212	17774	0.573	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM053123\
Data File : BM040094.D
Acq On : 01 Jun 2023 00:08
Operator : CG/JU
Sample : 02961-10
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
H0FQ2

Quant Time: Jun 01 00:43:36 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM053123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 31 18:20:24 2023
Response via : Initial Calibration

