

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060121\
 Data File : BM030236.D
 Acq On : 01 Jun 2021 14:02
 Operator : CG/JU
 Sample : SSTD1.676
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.6076

Quant Time: Jun 01 14:32:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 13:27:12 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	5915	0.40	ng/ul	0.00
4) Naphthalene-d8	10.66	136	20437	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.49	164	10850	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.22	188	21249	0.40	ng/ul	0.00
17) Chrysene-d12	21.38	240	14275	0.40	ng/ul	0.00
23) Perylene-d12	23.67	264	11469	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.33	96	8428	0.91	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.24	152	43783	1.49	ng/ul	0.00
18) Fluoranthene-d10	19.24	212	74510	1.66	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.37	88	9134	1.051	ng/ul#	83
5) Naphthalene	10.71	128	87203	1.563	ng/ul	99
7) 2-Methylnaphthalene	12.31	142	60899	1.636	ng/ul	99
8) 1-Methylnaphthalene	12.53	142	58143	1.592	ng/ul	99
10) Acenaphthylene	14.21	152	76715	1.706	ng/ul	99
11) Acenaphthene	14.55	153	63834	1.609	ng/ul	99
12) Fluorene	15.53	166	73928	1.750	ng/ul	99
14) Pentachlorophenol	16.87	266	8966	1.894	ng/ul	99
15) Phenanthrene	17.26	178	110369	1.707	ng/ul	99
16) Anthracene	17.35	178	98697	1.814	ng/ul	98
19) Fluoranthene	19.27	202	114379	1.977	ng/ul	99
20) Pyrene	19.63	202	111211	1.831	ng/ul	100
21) Benzo(a)anthracene	21.36	228	82279	1.902	ng/ul	99
22) Chrysene	21.42	228	87375	1.551	ng/ul	100
24) Benzo(b)fluoranthene	22.97	252	82772	1.977	ng/ul	88
25) Benzo(k)fluoranthene	23.02	252	80943	1.605	ng/ul#	89
26) Benzo(a)pyrene	23.56	252	66964	1.679	ng/ul#	86
27) Indeno(1,2,3-cd)pyrene	25.99	276	86651	1.647	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.00	278	70720	1.666	ng/ul	94
29) Benzo(g,h,i)perylene	26.70	276	74303	1.594	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060121\
Data File : BM030236.D
Acq On : 01 Jun 2021 14:02
Operator : CG/JU
Sample : SSTD1.676
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.6076

Quant Time: Jun 01 14:32:49 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 13:27:12 2021
Response via : Initial Calibration

