

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060121\
 Data File : BM030232.D
 Acq On : 01 Jun 2021 11:37
 Operator : CG/JU
 Sample : SST0.178
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.1078

Quant Time: Jun 01 13:29:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM060121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 01 13:27:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.867	152	6148	0.40	ng/ul	0.00
4) Naphthalene-d8	10.657	136	19433	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.485	164	9432	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.221	188	17642	0.40	ng/ul	0.00
17) Chrysene-d12	21.379	240	12993	0.40	ng/ul	0.00
23) Perylene-d12	23.665	264	12415	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.00	ng/ul	
6) 2-Methylnaphthalene-d10	12.241	152	2819	0.10	ng/ul	0.00
18) Fluoranthene-d10	19.237	212	4401	0.11	ng/ul	0.00
Target Compounds				Qvalue		
5) Naphthalene	10.707	128	5977	0.11	ng/ul	96
7) 2-Methylnaphthalene	12.317	142	4092	0.12	ng/ul	98
8) 1-Methylnaphthalene	12.533	142	3892	0.11	ng/ul	100
10) Acenaphthylene	14.208	152	4780	0.12	ng/ul	97
11) Acenaphthene	14.550	153	4167	0.12	ng/ul	98
12) Fluorene	15.532	166	4878	0.13	ng/ul	96
15) Phenanthrene	17.263	178	7065	0.13	ng/ul	96
16) Anthracene	17.353	178	5753	0.13	ng/ul	97
19) Fluoranthene	19.268	202	6778	0.13	ng/ul#	93
20) Pyrene	19.630	202	7162	0.13	ng/ul#	94
21) Benzo(a)anthracene	21.362	228	5529	0.14	ng/ul	98
22) Chrysene	21.413	228	5878	0.11	ng/ul	98
24) Benzo(b)fluoranthene	22.970	252	6516	0.14	ng/ul	68
25) Benzo(k)fluoranthene	23.016	252	5967	0.11	ng/ul#	69
26) Benzo(a)pyrene	23.561	252	5130	0.12	ng/ul#	56
27) Indeno(1,2,3-cd)pyrene	25.989	276	6877	0.12	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.999	278	5551	0.12	ng/ul#	83
29) Benzo(g,h,i)perylene	26.696	276	6008	0.12	ng/ul	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060121\
Data File : BM030232.D
Acq On : 01 Jun 2021 11:37
Operator : CG/JU
Sample : SST0.178
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.1078

Quant Time: Jun 01 13:29:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM060121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 01 13:27:12 2021
Response via : Initial Calibration

