

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001524.D
 Acq On : 02 Jun 2015 21:51
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215

Quant Time: Jun 03 02:31:20 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 02:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	0.00
2	1,4-Dioxane	1.000	0.927	7.3	100	0.00
3	n-Nitrosodimethylamine	1.000	0.988	1.2	99	0.00
4 S	2-Fluorophenol	1.000	0.959	4.1	101	-0.02
5 S	Phenol-d6	1.000	0.954	4.6	101	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
7 S	Nitrobenzene-d5	1.000	0.948	5.2	101	0.00
8	Nitrobenzene	1.000	0.941	5.9	100	0.00
9	Naphthalene	1.000	0.963	3.7	101	0.00
10	Hexachlorobutadiene	1.000	0.964	3.6	101	0.00
11	2-Methylnaphthalene	1.000	0.975	2.5	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	105	0.00
13 S	2,4,6-Tribromophenol	1.000	1.000	0.0	106	0.00
14 S	2-Fluorobiphenyl	1.000	0.956	4.4	102	0.00
15	Acenaphthylene	1.000	0.950	5.0	102	0.00
16	Acenaphthene	1.000	0.920	8.0	97	0.00
17	Fluorene	1.000	0.939	6.1	101	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	109	0.00
19	4-Bromophenyl-phenylether	1.000	0.908	9.2	101	0.00
20	Hexachlorobenzene	1.000	0.878	12.2	99	0.00
21	Pentachlorophenol	1.000	0.875	12.5	99	0.00
22	Phenanthrene	1.000	1.077	-7.7	125	0.00
23	Anthracene	1.000	0.865	13.5	97	0.00
24	Fluoranthene	1.000	0.932	6.8	106	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	106	0.00
26	Pyrene	1.000	1.004	-0.4	109	0.00
27 S	Terphenyl-d14	1.000	0.999	0.1	106	0.00
28	Benzo(a)anthracene	1.000	0.987	1.3	109	-0.01
29	Chrysene	1.000	1.004	-0.4	109	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.892	10.8	108	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	0.969	3.1	105	0.00
32 I	Perylene-d12	5.000	5.000	0.0	107	0.00
33	Benzo(b)fluoranthene	1.000	0.935	6.5	97	0.00
34	Benzo(k)fluoranthene	1.000	1.048	-4.8	116	0.00
35 C	Benzo(a)pyrene	1.000	0.977	2.3	105	0.00
36	Dibenzo(a,h)anthracene	1.000	0.973	2.7	104	0.00
37	Benzo(g,h,i)perylene	1.000	0.970	3.0	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0