

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001532.D
 Acq On : 03 Jun 2015 02:36
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 03 03:36:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 02:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.518	0.467	9.8	98	0.00
3	n-Nitrosodimethylamine	0.710	0.735	-3.5	96	0.00
4 S	2-Fluorophenol	1.093	1.113	-1.8	99	-0.02
5 S	Phenol-d6	1.351	1.364	-1.0	98	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
7 S	Nitrobenzene-d5	0.311	0.318	-2.3	99	0.00
8	Nitrobenzene	0.331	0.335	-1.2	98	0.00
9	Naphthalene	1.055	1.037	1.7	94	0.00
10	Hexachlorobutadiene	0.184	0.184	0.0	95	0.00
11	2-Methylnaphthalene	0.662	0.655	1.1	94	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
13 S	2,4,6-Tribromophenol	0.093	0.117	-25.8#	119	0.00
14 S	2-Fluorobiphenyl	1.501	1.475	1.7	94	0.00
15	Acenaphthylene	2.136	2.089	2.2	94	0.00
16	Acenaphthene	1.335	1.318	1.3	93	0.00
17	Fluorene	2.227	1.870	16.0	81	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19	4-Bromophenyl-phenylether	0.070	0.066	5.7	97	0.00
20	Hexachlorobenzene	0.187	0.135	27.8#	75	0.00
21	Pentachlorophenol	0.081	0.066	18.5	82	0.00
22	Phenanthrene	0.874	1.012	-15.8	123	0.00
23	Anthracene	1.057	0.802	24.1	78	0.00
24	Fluoranthene	0.877	0.793	9.6	95	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
26	Pyrene	1.509	1.495	0.9	96	0.00
27 S	Terphenyl-d14	0.616	0.606	1.6	93	0.00
28	Benzo(a)anthracene	1.346	1.360	-1.0	100	-0.01
29	Chrysene	1.275	1.252	1.8	96	-0.01
30	Bis(2-ethylhexyl)phthalate	0.951	0.908	4.5	104	-0.01
31	Indeno(1,2,3-cd)pyrene	1.417	1.430	-0.9	98	-0.02
32 I	Perylene-d12	1.000	1.000	0.0	98	0.00
33	Benzo(b)fluoranthene	1.428	1.398	2.1	93	0.00
34	Benzo(k)fluoranthene	1.248	1.260	-1.0	103	0.00
35 C	Benzo(a)pyrene	1.217	1.212	0.4	98	0.00
36	Dibenzo(a,h)anthracene	1.211	1.198	1.1	97	-0.02
37	Benzo(g,h,i)perylene	1.267	1.256	0.9	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0