

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001538.D
 Acq On : 03 Jun 2015 07:20
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 03 08:05:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 05:42:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.518	0.469	9.5	103	0.00
3	n-Nitrosodimethylamine	0.710	0.713	-0.4	97	0.00
4 S	2-Fluorophenol	1.093	1.111	-1.6	103	-0.02
5 S	Phenol-d6	1.351	1.364	-1.0	103	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
7 S	Nitrobenzene-d5	0.311	0.333	-7.1	108	0.00
8	Nitrobenzene	0.331	0.352	-6.3	107	-0.01
9	Naphthalene	1.055	1.034	2.0	98	0.00
10	Hexachlorobutadiene	0.184	0.181	1.6	98	0.00
11	2-Methylnaphthalene	0.662	0.650	1.8	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
13 S	2,4,6-Tribromophenol	0.093	0.131	-40.9#	138	0.00
14 S	2-Fluorobiphenyl	1.501	1.468	2.2	97	0.00
15	Acenaphthylene	2.136	2.113	1.1	99	0.00
16	Acenaphthene	1.335	1.326	0.7	97	0.00
17	Fluorene	2.227	1.698	23.8	76	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
19	4-Bromophenyl-phenylether	0.070	0.083	-18.6	121	0.00
20	Hexachlorobenzene	0.187	0.122	34.8#	67	0.00
21	Pentachlorophenol	0.081	0.069	14.8	85	0.00
22	Phenanthrene	0.874	1.137	-30.1#	137	0.00
23	Anthracene	1.057	0.728	31.1#	70	0.00
24	Fluoranthene	0.877	0.836	4.7	99	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
26	Pyrene	1.509	1.496	0.9	100	0.00
27 S	Terphenyl-d14	0.616	0.604	1.9	96	0.00
28	Benzo(a)anthracene	1.346	1.413	-5.0	108	-0.01
29	Chrysene	1.275	1.267	0.6	101	-0.01
30	Bis(2-ethylhexyl)phthalate	0.951	0.966	-1.6	115	-0.01
31	Indeno(1,2,3-cd)pyrene	1.417	1.485	-4.8	106	-0.02
32 I	Perylene-d12	1.000	1.000	0.0	106	0.00
33	Benzo(b)fluoranthene	1.428	1.373	3.9	99	0.00
34	Benzo(k)fluoranthene	1.248	1.273	-2.0	112	0.00
35 C	Benzo(a)pyrene	1.217	1.227	-0.8	107	0.00
36	Dibenzo(a,h)anthracene	1.211	1.199	1.0	105	-0.02
37	Benzo(g,h,i)perylene	1.267	1.263	0.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0