

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001519.D
 Acq On : 02 Jun 2015 18:53
 Operator : TP/IZ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICCC001

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:14 PM

Quant Time: Jun 03 02:13:11 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 02 15:30:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	18323	5.00	ng	0.00
6) Naphthalene-d8	10.49	136	64698	5.00	ng	0.00
12) Acenaphthene-d10	14.36	164	31529	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	91007	5.00	ng	0.00
25) Chrysene-d12	21.30	240	53711	5.00	ng	0.00
32) Perylene-d12	23.54	264	49104	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.28	112	3930	0.97	ng	-0.02
5) Phenol-d6	6.86	99	4834	0.97	ng	0.00
7) Nitrobenzene-d5	8.87	82	3959	0.99	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	587	0.87	ng	0.00
14) 2-Fluorobiphenyl	12.98	172	9378	1.00	ng	0.00
27) Terphenyl-d14	19.75	244	6657	0.99	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.23	88	1663	0.89	ng	99
3) n-Nitrosodimethylamine	3.52	42	2671	1.02	ng	# 99
8) Nitrobenzene	8.91	77	4203	0.99	ng	96
9) Naphthalene	10.54	128	13543	0.99	ng	98
10) Hexachlorobutadiene	10.83	225	2381	1.00	ng	100
11) 2-Methylnaphthalene	12.17	142	8570	0.99	ng	97
15) Acenaphthylene	14.07	152	13215	0.97	ng	99
16) Acenaphthene	14.42	154	8440	1.01	ng	95
17) Fluorene	15.39	166	13719	1.05	ng	98
19) 4-Bromophenyl-phenylether	16.31	248	1245	0.82	ng	# 45
20) Hexachlorobenzene	16.41	284	3306	1.04	ng	# 61
21) Pentachlorophenol	16.75	266	1472	1.04	ng	91
22) Phenanthrene	17.12	178	15027	0.86	ng	99
23) Anthracene	17.23	178	18772	1.03	ng	99
24) Fluoranthene	19.16	202	15366	0.93	ng	98
26) Pyrene	19.53	202	15929	0.97	ng	100
28) Benzo(a)anthracene	21.27	228	13871	0.95	ng	94
29) Chrysene	21.33	228	13390m	0.98	ng	
30) Bis(2-ethylhexyl)phthalate	21.21	149	8969	0.74	ng	98
31) Indeno(1,2,3-cd)pyrene	25.80	276	14953	1.01	ng	# 94
33) Benzo(b)fluoranthene	22.87	252	14401	1.01	ng	94
34) Benzo(k)fluoranthene	22.91	252	11813	0.93	ng	# 95
35) Benzo(a)pyrene	23.44	252	11852	0.97	ng	# 96
36) Dibenzo(a,h)anthracene	25.83	278	11861	1.01	ng	# 94
37) Benzo(g,h,i)perylene	26.50	276	12344	1.00	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
Data File : BM001519.D
Acq On : 02 Jun 2015 18:53
Operator : TP/IZ
Sample : SSTDICCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICCC001

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM

Quant Time: Jun 03 02:13:11 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 02 15:30:22 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.71 min Scan# 326

Delta R.T. -0.00 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :

SSTDICCC001



Tgt Ion: 152 Resp: 18323

Ion Ratio Lower Upper




152 100

150 155.1 119.0 178.6

115 61.5 41.8 62.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#2

1,4-Dioxane

Concen: 0.89 ng

RT: 3.23 min Scan# 36

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53




Tgt Ion: 88 Resp: 1663


Ion Ratio Lower Upper

88 100

58 87.8 71.3 106.9

43 44.3 35.5 53.3





#3

n-Nitrosodimethylamine

Concen: 1.02 ng

RT: 3.52 min Scan# 55

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M





ClientSampleId :

SSTDICCC001

Tgt Ion:	42	Resp:	2671
Ion	Ratio	Lower	Upper
42	100		
74	90.6	72.2	108.2
44	10.7	6.2	9.2#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#4

2-Fluorophenol

Concen: 0.97 ng




RT: 5.28 min Scan# 169


Delta R.T. -0.02 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	112	Resp:	3930
Ion	Ratio	Lower	Upper
112	100		
64	66.2	51.7	77.5
63	36.7	29.7	44.5





#5

Phenol-d6

Concen: 0.97 ng

RT: 6.86 min Scan# 271

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M





ClientSampleId :

SSTDICCC001

Tgt Ion:	99	Resp:	4834
Ion Ratio	Lower	Upper	
99	100		
42	27.1	21.8	32.6
71	36.1	29.2	43.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.49 min Scan# 553


Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	136	Resp:	64698
Ion Ratio	Lower	Upper	
136	100		
137	10.3	9.1	13.7
54	12.6	4.3	6.5#
68	7.4	2.6	4.0#





#7

Nitrobenzene-d5

Concen: 0.99 ng

RT: 8.87 min Scan# 423

Delta R.T. -0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M





ClientSampleId :

SSTDICCC001

Tgt Ion:	82	Resp:	3959
Ion	Ratio	Lower	Upper
82	100		
128	38.5	30.2	45.4
54	56.4	52.9	79.3

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#8

Nitrobenzene

Concen: 0.99 ng




RT: 8.91 min Scan# 427


Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	77	Resp:	4203
Ion	Ratio	Lower	Upper
77	100		
123	39.0	33.7	50.5
65	14.4	11.3	16.9





#9

Naphthalene

Concen: 0.99 ng

RT: 10.54 min Scan# 556

Delta R.T. 0.00 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :




SSTDICCC001



Tgt Ion:	128	Resp:	13543
Ion Ratio	Lower	Upper	
128	100		
129	11.2	8.6	13.0
127	13.2	11.7	17.5

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#10

Hexachlorobutadiene

Concen: 1.00 ng




RT: 10.83 min Scan# 573


Delta R.T. -0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	225	Resp:	2381
Ion Ratio	Lower	Upper	
225	100		
223	62.9	50.2	75.2
227	63.5	51.1	76.7





#11

2-Methylnaphthalene

Concen: 0.99 ng

RT: 12.17 min Scan# 683

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M





ClientSampleId :

SSTDICCC001

Tgt Ion:	142	Resp:	8570
Ion Ratio	Lower	Upper	
142	100		
141	87.1	72.6	109.0
115	37.8	31.2	46.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#12

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.36 min Scan# 881


Delta R.T. -0.01 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	164	Resp:	31529
Ion Ratio	Lower	Upper	
164	100		
162	106.3	83.4	125.2
160	48.3	36.9	55.3









#13
 2,4,6-Tribromophenol
 Concen: 0.87 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.01 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC001

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	587		
332	0.0	90.3	135.5#	
141	63.5	0.0	0.0#	




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:14 PM



#14
 2-Fluorobiphenyl
 Concen: 1.00 ng
 RT: 12.98 min Scan# 761
 Delta R.T. -0.01 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	9378		
171	34.6	29.4	44.0	
170	23.6	20.2	30.4	









#15
Acenaphthylene
Concen: 0.97 ng
RT: 14.07 min Scan# 857
Delta R.T. -0.01 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53

Instrument :
BNA_M
ClientSampleId :
SSTDICCC001

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.1	22.7
153	13.0	10.2	15.4




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:14 PM



#16
Acenaphthene
Concen: 1.01 ng
RT: 14.42 min Scan# 886
Delta R.T. -0.01 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53

Tgt Ion	Ratio	Lower	Upper
154	100		
153	107.7	91.4	137.2
152	52.2	44.0	66.0





#17

Fluorene

Concen: 1.05 ng

RT: 15.39 min Scan# 930

Delta R.T. -0.01 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :




SSTDICCC001



Tgt Ion:	166	Resp:	13719
Ion Ratio	Lower	Upper	
166	100		
165	92.4	72.6	108.8
167	13.7	10.9	16.3

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.09 min Scan# 980


Delta R.T. -0.01 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	188	Resp:	91007
Ion Ratio	Lower	Upper	
188	100		
94	5.5	3.0	4.4#
80	5.1	2.5	3.7#









#19
4-Bromophenyl-phenylether
Concen: 0.82 ng
RT: 16.31 min Scan# 957
Delta R.T. 0.03 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53

Instrument :
BNA_M
ClientSampleId :
SSTDICCC001

Tgt Ion	Ratio	Lower	Upper
248	100		
250	126.6	76.8	115.2#
141	0.0	57.7	86.5#




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:14 PM



#20
Hexachlorobenzene
Concen: 1.04 ng
RT: 16.41 min Scan# 960
Delta R.T. -0.01 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	13.9	20.9#
249	22.0	39.0	58.4#





#21

Pentachlorophenol

Concen: 1.04 ng

RT: 16.75 min Scan# 970

Delta R.T. -0.01 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :




SSTDICCC001



Tgt Ion:	266	Resp:	1472
Ion Ratio	Lower	Upper	
266	100		
268	69.2	63.0	94.4
264	61.4	52.8	79.2

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#22

Phenanthrene

Concen: 0.86 ng




RT: 17.12 min Scan# 981


Delta R.T. -0.01 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53


Tgt Ion:	178	Resp:	15027
Ion Ratio	Lower	Upper	
178	100		
176	21.6	16.8	25.2
179	14.0	11.0	16.4





#23
 Anthracene
 Concen: 1.03 ng
 RT: 17.23 min Scan# 984
 Delta R.T. -0.01 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC001



Tgt Ion	Ratio	Lower	Upper
178	100		
176	17.3	13.1	19.7
179	16.0	12.8	19.2




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:14 PM



#24
 Fluoranthene
 Concen: 0.93 ng
 RT: 19.16 min Scan# 1090
 Delta R.T. 0.00 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.7	8.6	12.8
203	17.1	13.8	20.6





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.30 min Scan# 1253

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M


ClientSampleId :

SSTDICCC001

Tgt Ion:	240	Resp:	53711
Ion Ratio	Lower	Upper	
240	100		
120	2.9	9.7	14.5#
236	23.7	24.8	37.2#

Manual Integrations
APPROVED




apatel
6/3/2015 5:28:14 PM



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#26

Pyrene

Concen: 0.97 ng


RT: 19.53 min Scan# 1120

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53



Tgt Ion:	202	Resp:	15929
Ion Ratio	Lower	Upper	
202	100		
200	20.6	16.4	24.6
203	17.6	13.9	20.9




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#27

Terphenyl-d14

Concen: 0.99 ng

RT: 19.75 min Scan# 1138

Delta R.T. 0.00 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :




SSTDICCC001



Tgt Ion:	244	Resp:	6657
Ion Ratio	Lower	Upper	
244	100		
212	9.8	7.2	10.8
122	8.0	5.1	7.7

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#28

Benzo(a)anthracene

Concen: 0.95 ng




RT: 21.27 min Scan# 1251


Delta R.T. -0.01 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	228	Resp:	13871
Ion Ratio	Lower	Upper	
228	100		
226	31.1	21.9	32.9
229	17.2	15.1	22.7





#29

Chrysene

Concen: 0.98 ng m

RT: 21.33 min Scan# 1255

Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Instrument :

BNA_M





ClientSampleId :

SSTDICCC001

Tgt Ion:	228	Resp:	13390
Ion Ratio	Lower	Upper	
228	100		
226	26.5	27.6	41.4#
229	21.6	13.6	20.4#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.74 ng




RT: 21.21 min Scan# 1247


Delta R.T. -0.01 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	149	Resp:	8969
Ion Ratio	Lower	Upper	
149	100		
167	25.5	21.0	31.6
279	2.2	2.2	3.2









#31
 Indeno(1,2,3-cd)pyrene
 Concen: 1.01 ng
 RT: 25.80 min Scan# 1663
 Delta R.T. -0.01 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC001

Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	35.5	23.4	35.0#
277	24.4	19.5	29.3




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:14 PM



#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.54 min Scan# 1446
 Delta R.T. 0.00 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Tgt Ion	Ratio	Resp Lower	Resp Upper
264	100		
260	25.6	20.1	30.1
265	23.2	17.1	25.7





#33

Benzo(b)fluoranthene

Concen: 1.01 ng

RT: 22.87 min Scan# 1381

Delta R.T. 0.00 min

Lab File: BM001519.D


Acq: 02 Jun 2015 18:53

Instrument :

BNA_M

ClientSampleId :




SSTDICCC001



Tgt Ion:	252	Resp:	14401
Ion Ratio	Lower	Upper	
252	100		
253	24.7	17.4	26.0
125	14.9	10.2	15.4

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:14 PM



#34

Benzo(k)fluoranthene

Concen: 0.93 ng




RT: 22.91 min Scan# 1385


Delta R.T. 0.00 min

Lab File: BM001519.D

Acq: 02 Jun 2015 18:53

Tgt Ion:	252	Resp:	11813
Ion Ratio	Lower	Upper	
252	100		
253	23.8	17.6	26.4
125	15.2	9.7	14.5#









#35
Benzo(a)pyrene
Concen: 0.97 ng
RT: 23.44 min Scan# 1436
Delta R.T. 0.00 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53

Instrument :
BNA_M
ClientSampleId :
SSTDICCC001

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.4	18.6	28.0
125	18.1	10.7	16.1




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:14 PM



#36
Dibenzo(a,h)anthracene
Concen: 1.01 ng
RT: 25.83 min Scan# 1665
Delta R.T. -0.01 min
Lab File: BM001519.D
Acq: 02 Jun 2015 18:53


Tgt Ion	Ratio	Lower	Upper
278	100		
139	24.4	15.7	23.5
279	25.2	19.4	29.2





#37
 Benzo(g,h,i)perylene
 Concen: 1.00 ng
 RT: 26.50 min Scan# 1730
 Delta R.T. 0.00 min
 Lab File: BM001519.D
 Acq: 02 Jun 2015 18:53

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC001



Tgt Ion:	276	Resp:	12344
Ion Ratio		Lower	Upper
276	100		
277	24.1	19.8	29.6
138	30.3	20.2	30.2#

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:14 PM

