

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001524.D
 Acq On : 02 Jun 2015 21:51
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:35 PM

Quant Time: Jun 03 02:31:20 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 02:19:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	18895	5.00	ng	0.00
6) Naphthalene-d8	10.49	136	67476	5.00	ng	0.00
12) Acenaphthene-d10	14.36	164	33251	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	99601	5.00	ng	0.00
25) Chrysene-d12	21.30	240	57048	5.00	ng	0.00
32) Perylene-d12	23.54	264	52557	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	3960	0.96	ng	-0.02
5) Phenol-d6	6.86	99	4869	0.95	ng	0.00
7) Nitrobenzene-d5	8.87	82	3980	0.95	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	620	1.00	ng	0.00
14) 2-Fluorobiphenyl	12.98	172	9546	0.96	ng	0.00
27) Terphenyl-d14	19.73	244	7024	1.00	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.23	88	1659	0.93	ng	98
3) n-Nitrosodimethylamine	3.52	42	2653	0.99	ng	99
8) Nitrobenzene	8.91	77	4200	0.94	ng	97
9) Naphthalene	10.54	128	13708	0.96	ng	97
10) Hexachlorobutadiene	10.83	225	2394	0.96	ng	100
11) 2-Methylnaphthalene	12.17	142	8713	0.97	ng	93
15) Acenaphthylene	14.07	152	13489	0.95	ng	100
16) Acenaphthene	14.42	154	8173	0.92	ng	99
17) Fluorene	15.39	166	13911	0.94	ng	99
19) 4-Bromophenyl-phenylether	16.27	248	1259	0.91	ng	# 1
20) Hexachlorobenzene	16.41	284	3275	0.88	ng	# 59
21) Pentachlorophenol	16.75	266	1461	0.88	ng	91
22) Phenanthrene	17.12	178	18748	1.08	ng	99
23) Anthracene	17.23	178	18210	0.86	ng	99
24) Fluoranthene	19.16	202	16282	0.93	ng	97
26) Pyrene	19.53	202	17287	1.00	ng	100
28) Benzo(a)anthracene	21.27	228	15159	0.99	ng	97
29) Chrysene	21.33	228	14608m	1.00	ng	
30) Bis(2-ethylhexyl)phthalate	21.21	149	9673	0.89	ng	97
31) Indeno(1,2,3-cd)pyrene	25.80	276	15669	0.97	ng	# 94
33) Benzo(b)fluoranthene	22.85	252	14037m	0.94	ng	
34) Benzo(k)fluoranthene	22.91	252	13748	1.05	ng	95
35) Benzo(a)pyrene	23.44	252	12495	0.98	ng	# 96
36) Dibenzo(a,h)anthracene	25.83	278	12389	0.97	ng	96
37) Benzo(g,h,i)perylene	26.49	276	12915	0.97	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
Data File : BM001524.D
Acq On : 02 Jun 2015 21:51
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_M
Client Sampled :
ICVBM060215

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM

Quant Time: Jun 03 02:31:20 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 03 02:19:00 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.71 min Scan# 326

Delta R.T. 0.00 min

Lab File: BM001524.D


Acq: 02 Jun 2015 21:51

Instrument :

BNA_M

ClientSampleId :




ICVBM060215



Tgt Ion:	152	Resp:	18895
Ion Ratio	Lower	Upper	
152	100		
150	150.4	119.0	178.6
115	58.4	41.8	62.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#2

1,4-Dioxane

Concen: 0.93 ng




RT: 3.23 min Scan# 36


Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	88	Resp:	1659
Ion Ratio	Lower	Upper	
88	100		
58	88.2	71.3	106.9
43	42.7	35.5	53.3





#3

n-Nitrosodimethylamine

Concen: 0.99 ng

RT: 3.52 min Scan# 55

Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Instrument :

BNA_M


ClientSampleId :

ICVBM060215

Tgt Ion:	42	Resp:	2653
Ion	Ratio	Lower	Upper
42	100		
74	90.0	72.2	108.2
44	9.2	6.2	9.2

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM




Abundance Ion 42.00 (41.70 to 42.70): BM001521.D (-51) (-)


Ion 74.00 (73.70 to 74.70): BM001521.D (-51) (-)

Ion 44.00 (43.70 to 44.70): BM001521.D (-51) (-)

Time-->



Time-->



#4

2-Fluorophenol

Concen: 0.96 ng


RT: 5.28 min Scan# 169

Delta R.T. -0.02 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	112	Resp:	3960
Ion	Ratio	Lower	Upper
112	100		
64	66.0	51.7	77.5
63	36.6	29.7	44.5




Abundance Ion 112.00 (111.70 to 112.70): BM001521.D (-165) (-)




Ion 64.00 (63.70 to 64.70): BM001521.D (-165) (-)

Ion 63.00 (62.70 to 63.70): BM001521.D (-165) (-)

Time-->



Time-->



#5

Phenol-d6

Concen: 0.95 ng

RT: 6.86 min Scan# 271

Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Resp	Lower	Upper
99	100	4869		
42	27.1		21.8	32.6
71	35.7		29.2	43.8

Instrument :





BNA_M

ClientSampleId :

ICVBM060215

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#6

Naphthalene-d8

Concen: 5.00 ng


RT: 10.49 min Scan# 553


Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	67476		
137	10.3		9.1	13.7
54	11.3		4.3	6.5#
68	6.8		2.6	4.0#





#7

Nitrobenzene-d5

Concen: 0.95 ng

RT: 8.87 min Scan# 423

Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Instrument :

BNA_M





ClientSampleId :

ICVBM060215

Tgt Ion:	82	Resp:	3980
Ion Ratio	Lower	Upper	
82	100		
128	39.7	30.2	45.4
54	54.9	52.9	79.3

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#8

Nitrobenzene

Concen: 0.94 ng




RT: 8.91 min Scan# 427


Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	77	Resp:	4200
Ion Ratio	Lower	Upper	
77	100		
123	40.0	33.7	50.5
65	14.6	11.3	16.9





#9

Naphthalene

Concen: 0.96 ng

RT: 10.54 min Scan# 556

Delta R.T. 0.00 min

Lab File: BM001524.D


Acq: 02 Jun 2015 21:51

Instrument :

BNA_M

ClientSampleId :




ICVBM060215



Tgt Ion:	128	Resp:	13708
Ion Ratio	Lower	Upper	
128	100		
129	11.5	8.6	13.0
127	13.1	11.7	17.5

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#10

Hexachlorobutadiene


Concen: 0.96 ng

RT: 10.83 min Scan# 573



Delta R.T. 0.00 min


Lab File: BM001524.D

Acq: 02 Jun 2015 21:51



Tgt Ion:	225	Resp:	2394
Ion Ratio	Lower	Upper	
225	100		
223	63.0	50.2	75.2
227	63.7	51.1	76.7









#11
2-Methylnaphthalene
Concen: 0.97 ng
RT: 12.17 min Scan# 683
Delta R.T. 0.00 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Instrument :
BNA_M
ClientSampleId :
ICVBM060215

Tgt Ion	Ratio	Lower	Upper
142	100		
141	83.2	72.6	109.0
115	35.1	31.2	46.8




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:35 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.36 min Scan# 881
Delta R.T. -0.01 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.0	83.4	125.2
160	44.4	36.9	55.3









#13
 2,4,6-Tribromophenol
 Concen: 1.00 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	620		
332	89.7	90.3	135.5	#
141	57.6	0.0	0.0	#




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:35 PM



#14
 2-Fluorobiphenyl
 Concen: 0.96 ng
 RT: 12.98 min Scan# 761
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	9546		
171	34.5	29.4	44.0	
170	22.7	20.2	30.4	





#15

Acenaphthylene

Concen: 0.95 ng

RT: 14.07 min Scan# 857

Delta R.T. -0.01 min

Lab File: BM001524.D


Acq: 02 Jun 2015 21:51

Instrument :

BNA_M

ClientSampleId :




ICVBM060215



Tgt Ion:	152	Resp:	13489
Ion Ratio	Lower	Upper	
152	100		
151	18.9	15.1	22.7
153	12.8	10.2	15.4

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#16

Acenaphthene

Concen: 0.92 ng




RT: 14.42 min Scan# 886


Delta R.T. -0.01 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	154	Resp:	8173
Ion Ratio	Lower	Upper	
154	100		
153	114.9	91.4	137.2
152	55.8	44.0	66.0









#17
Fluorene
 Concen: 0.94 ng
 RT: 15.39 min Scan# 930
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215

Tgt Ion	Ratio	Lower	Upper
166	100		
165	91.1	72.6	108.8
167	14.1	10.9	16.3




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:35 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.09 min Scan# 980
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
188	100		
94	4.7	3.0	4.4#
80	4.2	2.5	3.7#









#19
4-Bromophenyl-phenylether
Concen: 0.91 ng
RT: 16.27 min Scan# 956
Delta R.T. -0.01 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Instrument :
BNA_M
ClientSampleId :
ICVBM060215

Tgt Ion	Ratio	Lower	Upper
248	100		
250	0.0	76.8	115.2#
141	293.3	57.7	86.5#




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:35 PM



#20
Hexachlorobenzene
Concen: 0.88 ng
RT: 16.41 min Scan# 960
Delta R.T. -0.01 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	13.9	20.9#
249	20.5	39.0	58.4#






#21
 Pentachlorophenol
 Concen: 0.88 ng
 RT: 16.75 min Scan# 970
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215


Tgt Ion	Ratio	Lower	Upper
266	100		
268	70.8	63.0	94.4
264	58.8	52.8	79.2

Manual Integrations
 APPROVED



apatel
 6/3/2015 5:28:35 PM



Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E




Time-->





#22
 Phenanthrene
 Concen: 1.08 ng
 RT: 17.12 min Scan# 981
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51


Tgt Ion	Ratio	Lower	Upper
178	100		
176	21.3	16.8	25.2
179	14.1	11.0	16.4

Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E




Time-->





#23
 Anthracene
 Concen: 0.86 ng
 RT: 17.23 min Scan# 984
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51




Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215



Tgt Ion	Ratio	Lower	Upper
178	100		
176	16.9	13.1	19.7
179	16.2	12.8	19.2




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:35 PM



#24
 Fluoranthene
 Concen: 0.93 ng
 RT: 19.16 min Scan# 1090
 Delta R.T. 0.00 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.7	8.6	12.8
203	17.6	13.8	20.6









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.30 min Scan# 1253
Delta R.T. 0.00 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Instrument :
BNA_M
ClientSampleId :
ICVBM060215

Tgt Ion	Ratio	Lower	Upper
240	100		
120	2.4	9.7	14.5#
236	23.7	24.8	37.2#




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:35 PM



#26
Pyrene
Concen: 1.00 ng
RT: 19.53 min Scan# 1120
Delta R.T. 0.00 min
Lab File: BM001524.D
Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.4	24.6
203	17.3	13.9	20.9






#27
 Terphenyl-d14
 Concen: 1.00 ng
 RT: 19.73 min Scan# 1137
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215




Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.6	7.2	10.8
122	25.8	5.1	7.7

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:35 PM






Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E




#28
 Benzo(a)anthracene
 Concen: 0.99 ng
 RT: 21.27 min Scan# 1251
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.1	21.9	32.9
229	17.9	15.1	22.7

Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E





#29

Chrysene

Concen: 1.00 ng m

RT: 21.33 min Scan# 1255

Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Instrument :

BNA_M


ClientSampleId :

ICVBM060215




Tgt Ion:	228	Resp:	14608
Ion Ratio	Lower	Upper	
228	100		
226	25.3	27.6	41.4#
229	22.1	13.6	20.4#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.89 ng

RT: 21.21 min Scan# 1247




Delta R.T. -0.01 min


Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	149	Resp:	9673
Ion Ratio	Lower	Upper	
149	100		
167	24.6	21.0	31.6
279	2.2	2.2	3.2

Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E









#31
 Indeno(1,2,3-cd)pyrene
 Concen: 0.97 ng
 RT: 25.80 min Scan# 1663
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215

Tgt Ion	Ratio	Resp	Lower	Upper
276	100	15669		
138	35.5	23.4	35.0	
277	24.4	19.5	29.3	




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:35 PM



#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.54 min Scan# 1446
 Delta R.T. 0.00 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	52557		
260	24.8	20.1	30.1	
265	23.4	17.1	25.7	





#33

Benzo(b)fluoranthene

Concen: 0.94 ng m

RT: 22.85 min Scan# 1380

Delta R.T. -0.01 min

Lab File: BM001524.D


Acq: 02 Jun 2015 21:51

Instrument :

BNA_M

ClientSampleId :




ICVBM060215



Tgt Ion:	252	Resp:	14037
Ion Ratio	Lower	Upper	
252	100		
253	22.0	17.4	26.0
125	15.7	10.2	15.4#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#34

Benzo(k)fluoranthene

Concen: 1.05 ng




RT: 22.91 min Scan# 1385


Delta R.T. 0.00 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51

Tgt Ion:	252	Resp:	13748
Ion Ratio	Lower	Upper	
252	100		
253	24.5	17.6	26.4
125	14.2	9.7	14.5





#35

Benzo(a)pyrene

Concen: 0.98 ng

RT: 23.44 min Scan# 1436

Delta R.T. 0.00 min

Lab File: BM001524.D


Acq: 02 Jun 2015 21:51

Instrument :

BNA_M

ClientSampleId :

ICVBM060215



Tgt Ion: 252 Resp: 12495

Ion Ratio Lower Upper




252 100

253 24.1 18.6 28.0

125 17.2 10.7 16.1#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:35 PM



#36

Dibenzo(a,h)anthracene

Concen: 0.97 ng

RT: 25.83 min Scan# 1665

Delta R.T. -0.01 min

Lab File: BM001524.D

Acq: 02 Jun 2015 21:51




Tgt Ion: 278 Resp: 12389


Ion Ratio Lower Upper

278 100

139 22.7 15.7 23.5


279 25.4 19.4 29.2





#37
 Benzo(g,h,i)perylene
 Concen: 0.97 ng
 RT: 26.49 min Scan# 1729
 Delta R.T. -0.01 min
 Lab File: BM001524.D
 Acq: 02 Jun 2015 21:51

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060215



Tgt Ion:	276	Resp:	12915
Ion Ratio	Lower	Upper	
276	100		
277	24.1	19.8	29.6
138	32.5	20.2	30.2

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:35 PM

