

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
 Data File : BM001534.D
 Acq On : 03 Jun 2015 04:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:55 PM

Quant Time: Jun 03 05:42:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 02:19:00 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	17305	5.00	ng	0.00
6) Naphthalene-d8	10.49	136	60932	5.00	ng	0.00
12) Acenaphthene-d10	14.36	164	29590	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	90106	5.00	ng	0.00
25) Chrysene-d12	21.28	240	49714	5.00	ng	-0.01
32) Perylene-d12	23.53	264	46672	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	3706	0.98	ng	-0.02
5) Phenol-d6	6.86	99	4525	0.97	ng	0.00
7) Nitrobenzene-d5	8.87	82	3780	1.00	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	685	1.24	ng	0.00
14) 2-Fluorobiphenyl	12.98	172	8829	0.99	ng	0.00
27) Terphenyl-d14	19.73	244	6136	1.00	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.23	88	1593	0.98	ng	99
3) n-Nitrosodimethylamine	3.52	42	2485	1.01	ng	100
8) Nitrobenzene	8.91	77	4019	1.00	ng	97
9) Naphthalene	10.54	128	12752	0.99	ng	96
10) Hexachlorobutadiene	10.83	225	2250	1.00	ng	100
11) 2-Methylnaphthalene	12.16	142	7973	0.99	ng	94
15) Acenaphthylene	14.07	152	12205	0.97	ng	100
16) Acenaphthene	14.42	154	7784	0.98	ng	97
17) Fluorene	15.39	166	10883	0.83	ng	98
19) 4-Bromophenyl-phenylether	16.27	248	1266	1.01	ng	# 1
20) Hexachlorobenzene	16.41	284	2523	0.75	ng	# 37
21) Pentachlorophenol	16.75	266	1215	0.82	ng	90
22) Phenanthrene	17.12	178	19051	1.21	ng	98
23) Anthracene	17.23	178	15014	0.79	ng	99
24) Fluoranthene	19.16	202	14204	0.90	ng	98
26) Pyrene	19.53	202	14929	1.00	ng	100
28) Benzo(a)anthracene	21.27	228	13663	1.02	ng	99
29) Chrysene	21.31	228	12107	0.96	ng	99
30) Bis(2-ethylhexyl)phthalate	21.21	149	8341	0.88	ng	# 97
31) Indeno(1,2,3-cd)pyrene	25.79	276	14085	1.00	ng	94
33) Benzo(b)fluoranthene	22.85	252	13183	0.99	ng	96
34) Benzo(k)fluoranthene	22.90	252	11613m	1.00	ng	
35) Benzo(a)pyrene	23.43	252	11118	0.98	ng	# 96
36) Dibenzo(a,h)anthracene	25.82	278	11139	0.99	ng	96
37) Benzo(g,h,i)perylene	26.48	276	11701	0.99	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060215\
Data File : BM001534.D
Acq On : 03 Jun 2015 04:58
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM

Quant Time: Jun 03 05:42:16 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 03 02:19:00 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.71 min Scan# 326

Delta R.T. -0.00 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :

SSTDCCC001



Tgt Ion: 152 Resp: 17305

Ion Ratio Lower Upper




152 100

150 148.3 119.0 178.6

115 56.4 41.8 62.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#2

1,4-Dioxane

Concen: 0.98 ng

RT: 3.23 min Scan# 36

Delta R.T. 0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58




Tgt Ion: 88 Resp: 1593


Ion Ratio Lower Upper

88 100

58 87.6 71.3 106.9

43 43.9 35.5 53.3





#3





n-Nitrosodimethylamine
Concen: 1.01 ng
RT: 3.52 min Scan# 55
Delta R.T. -0.00 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
42	100		
74	90.7	72.2	108.2
44	7.6	6.2	9.2

Manual Integrations
APPROVED




apatel
6/3/2015 5:28:55 PM




#4

2-Fluorophenol
Concen: 0.98 ng
RT: 5.28 min Scan# 169
Delta R.T. -0.02 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.1	51.7	77.5
63	37.1	29.7	44.5





#5

Phenol-d6

Concen: 0.97 ng

RT: 6.86 min Scan# 271

Delta R.T. -0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001

Tgt Ion:	99	Resp:	4525
Ion Ratio	Lower	Upper	
99	100		
42	27.4	21.8	32.6
71	36.4	29.2	43.8

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.49 min Scan# 553


Delta R.T. -0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	136	Resp:	60932
Ion Ratio	Lower	Upper	
136	100		
137	10.7	9.1	13.7
54	9.9	4.3	6.5#
68	5.9	2.6	4.0#





#7

Nitrobenzene-d5

Concen: 1.00 ng

RT: 8.87 min Scan# 423

Delta R.T. -0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001

Tgt Ion:	82	Resp:	3780
Ion Ratio	Lower	Upper	
82	100		
128	42.5	30.2	45.4
54	50.7	52.9	79.3

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#8

Nitrobenzene

Concen: 1.00 ng




RT: 8.91 min Scan# 427


Delta R.T. -0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	77	Resp:	4019
Ion Ratio	Lower	Upper	
77	100		
123	39.7	33.7	50.5
65	14.4	11.3	16.9





#9

Naphthalene

Concen: 0.99 ng

RT: 10.54 min Scan# 556

Delta R.T. -0.00 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion:	128	Resp:	12752
Ion Ratio	Lower	Upper	
128	100		
129	11.8	8.6	13.0
127	12.6	11.7	17.5

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#10

Hexachlorobutadiene

Concen: 1.00 ng




RT: 10.83 min Scan# 573


Delta R.T. -0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	225	Resp:	2250
Ion Ratio	Lower	Upper	
225	100		
223	62.6	50.2	75.2
227	63.8	51.1	76.7









#11
2-Methylnaphthalene
Concen: 0.99 ng
RT: 12.16 min Scan# 682
Delta R.T. -0.01 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Tgt Ion:142 Resp: 7973
Ion Ratio Lower Upper
142 100
141 94.5 72.6 109.0
115 45.5 31.2 46.8




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:55 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.36 min Scan# 881
Delta R.T. -0.01 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Tgt Ion:164 Resp: 29590
Ion Ratio Lower Upper
164 100
162 100.0 83.4 125.2
160 43.5 36.9 55.3









#13
 2,4,6-Tribromophenol
 Concen: 1.24 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
330	100		
332	84.2	90.3	135.5#
141	42.8	0.0	0.0#




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:55 PM



#14
 2-Fluorobiphenyl
 Concen: 0.99 ng
 RT: 12.98 min Scan# 761
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.2	29.4	44.0
170	21.2	20.2	30.4





#15

Acenaphthylene

Concen: 0.97 ng

RT: 14.07 min Scan# 857

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.0	15.1	22.7
153	13.0	10.2	15.4

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#16

Acenaphthene

Concen: 0.98 ng




RT: 14.42 min Scan# 886


Delta R.T. -0.01 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
154	100		
153	110.0	91.4	137.2
152	53.6	44.0	66.0





#17

Fluorene

Concen: 0.83 ng

RT: 15.39 min Scan# 930

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion:	166	Resp:	10883
Ion Ratio	Lower	Upper	
166	100		
165	88.5	72.6	108.8
167	14.4	10.9	16.3

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.09 min Scan# 980


Delta R.T. -0.01 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	188	Resp:	90106
Ion Ratio	Lower	Upper	
188	100		
94	3.2	3.0	4.4
80	2.8	2.5	3.7









#19
4-Bromophenyl-phenylether
Concen: 1.01 ng
RT: 16.27 min Scan# 956
Delta R.T. -0.01 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
248	100		
250	0.0	76.8	115.2#
141	260.1	57.7	86.5#




Manual Integrations
APPROVED


apatel
6/3/2015 5:28:55 PM



#20
Hexachlorobenzene
Concen: 0.75 ng
RT: 16.41 min Scan# 960
Delta R.T. -0.01 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
284	100		
142	0.0	13.9	20.9#
249	0.0	39.0	58.4#






#21
 Pentachlorophenol
 Concen: 0.82 ng
 RT: 16.75 min Scan# 970
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001




Tgt Ion:	266	Resp:	1215
Ion Ratio	Lower	Upper	
266	100		
268	71.4	63.0	94.4
264	55.7	52.8	79.2

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:55 PM




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#22
 Phenanthrene
 Concen: 1.21 ng
 RT: 17.12 min Scan# 981
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Tgt Ion:	178	Resp:	19051
Ion Ratio	Lower	Upper	
178	100		
176	20.3	16.8	25.2
179	14.8	11.0	16.4




Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#23
 Anthracene
 Concen: 0.79 ng
 RT: 17.23 min Scan# 984
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001



Tgt Ion	Ratio	Lower	Upper
178	100		
176	16.1	13.1	19.7
179	16.5	12.8	19.2




Manual Integrations
 APPROVED


apatel
 6/3/2015 5:28:55 PM



#24
 Fluoranthene
 Concen: 0.90 ng
 RT: 19.16 min Scan# 1090
 Delta R.T. 0.00 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.2	8.6	12.8
203	17.2	13.8	20.6





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.28 min Scan# 1252

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion:	240	Resp:	49714
Ion Ratio	Lower	Upper	
240	100		
120	13.8	9.7	14.5
236	31.3	24.8	37.2

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#26

Pyrene

Concen: 1.00 ng




RT: 19.53 min Scan# 1120


Delta R.T. 0.00 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	202	Resp:	14929
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.4	24.6
203	17.4	13.9	20.9





#27

Terphenyl-d14

Concen: 1.00 ng

RT: 19.73 min Scan# 1137

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion:	244	Resp:	6136
Ion Ratio	Lower	Upper	
244	100		
212	11.9	7.2	10.8#
122	17.9	5.1	7.7#

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#28

Benzo(a)anthracene

Concen: 1.02 ng




RT: 21.27 min Scan# 1251


Delta R.T. -0.01 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	228	Resp:	13663
Ion Ratio	Lower	Upper	
228	100		
226	26.7	21.9	32.9
229	19.4	15.1	22.7





#29

Chrysene

Concen: 0.96 ng

RT: 21.31 min Scan# 1254

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion:	228	Resp:	12107
Ion Ratio	Lower	Upper	
228	100		
226	33.6	27.6	41.4
229	17.6	13.6	20.4

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.88 ng




RT: 21.21 min Scan# 1247


Delta R.T. -0.01 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion:	149	Resp:	8341
Ion Ratio	Lower	Upper	
149	100		
167	24.9	21.0	31.6
279	2.1	2.2	3.2#






#31
 Indeno(1,2,3-cd)pyrene
 Concen: 1.00 ng
 RT: 25.79 min Scan# 1662
 Delta R.T. -0.02 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001




Tgt Ion	Ratio	Resp	Lower	Upper
276	100	14085		
138	34.5	23.4	35.0	
277	24.6	19.5	29.3	

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:55 PM






Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E




#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.53 min Scan# 1445
 Delta R.T. -0.01 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	46672		
260	25.5	20.1	30.1	
265	22.4	17.1	25.7	

Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E





#33

Benzo(b)fluoranthene

Concen: 0.99 ng

RT: 22.85 min Scan# 1380

Delta R.T. -0.01 min

Lab File: BM001534.D


Acq: 03 Jun 2015 04:58

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001



Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.0	17.4	26.0
125	14.1	10.2	15.4

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM



#34

Benzo(k)fluoranthene

Concen: 1.00 ng m




RT: 22.90 min Scan# 1384


Delta R.T. -0.01 min

Lab File: BM001534.D

Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	17.6	26.4
125	14.8	9.7	14.5#






#35
Benzo(a)pyrene
Concen: 0.98 ng
RT: 23.43 min Scan# 1435
Delta R.T. -0.01 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001




Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.5	18.6	28.0
125	16.9	10.7	16.1

Manual Integrations
APPROVED

apatel
6/3/2015 5:28:55 PM






Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#36
Dibenzo(a,h)anthracene
Concen: 0.99 ng
RT: 25.82 min Scan# 1664
Delta R.T. -0.02 min
Lab File: BM001534.D
Acq: 03 Jun 2015 04:58

Tgt Ion	Ratio	Lower	Upper
278	100		
139	23.2	15.7	23.5
279	25.0	19.4	29.2

Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#37
 Benzo(g,h,i)perylene
 Concen: 0.99 ng
 RT: 26.48 min Scan# 1728
 Delta R.T. -0.02 min
 Lab File: BM001534.D
 Acq: 03 Jun 2015 04:58

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.8	29.6
138	30.3	20.2	30.2

Manual Integrations
 APPROVED

apatel
 6/3/2015 5:28:55 PM

