

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060217\
 Data File : BM010408.D
 Acq On : 03 Jun 2017 03:35
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.418

Quant Time: Jun 05 01:52:36 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM053017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 03 05:17:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
3	Naphthalene	1.055	1.035	1.9	99	0.00
4 SURR	2-Methylnaphthalene-d10	0.581	0.587	-1.0	100	0.00
5	2-Methylnaphthalene	0.744	0.758	-1.9	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
7	Acenaphthylene	1.997	2.105	-5.4	108	0.00
8 C	Acenaphthene	1.510	1.512	-0.1	103	0.00
9	Fluorene	1.808	1.811	-0.2	103	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
11	Pentachlorophenol	0.153	0.184	-20.3#	123	0.00
12	Phenanthrene	1.286	1.277	0.7	102	0.00
13	Anthracene	1.204	1.251	-3.9	107	0.00
14 SURR	Fluoranthene-d10	1.061	1.124	-5.9	106	0.00
15 C	Fluoranthene	1.595	1.660	-4.1	106	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
17	Pyrene	1.704	1.759	-3.2	113	0.00
18	Benzo(a)anthracene	1.626	1.726	-6.2	115	0.00
19	Chrysene	1.573	1.557	1.0	106	0.00
20 I	Perylene-d12	1.000	1.000	0.0	102	0.00
21	Benzo(b)fluoranthene	1.751	1.680	4.1	99	0.00
22	Benzo(k)fluoranthene	1.665	1.653	0.7	104	0.00
23 C	Benzo(a)pyrene	1.657	1.657	0.0	104	0.00
24	Indeno(1,2,3-cd)pyrene	2.178	2.065	5.2	98	0.00
25	Dibenzo(a,h)anthracene	1.735	1.647	5.1	98	0.00
26	Benzo(g,h,i)perylene	1.878	1.724	8.2	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0