


Benzaldehyde


Response Ratio



$R = -4.586e-002 A^2 + 5.918e-001 A + 9.622e-002$
Coef of Det (r^2) = 0.991526 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060222.M
Calibration Table Last Updated: Thu Jun 02 16:16:23 2022

Hexachlorocyclopentadiene

Response Ratio



$$\text{Response} = 3.400\text{e-}001 * \text{Amt} - 8.157\text{e-}002$$


Coef of Det (r^2) = 0.995136 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060222.M

Calibration Table Last Updated: Thu Jun 02 16:16:23 2022

2,4-Dinitrophenol

Response Ratio



$$\text{Response} = 2.312\text{e-}001 * \text{Amt} - 3.054\text{e-}002$$

Coef of Det (r^2) = 0.992567 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM060222.M

Calibration Table Last Updated: Thu Jun 02 16:16:23 2022