

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060315\
 Data File : BM001541.D
 Acq On : 03 Jun 2015 19:00
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

apatel
 6/5/2015 8:04:04 AM

Quant Time: Jun 04 04:00:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 04 03:14:13 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	26772	5.00	ng	0.00
6) Naphthalene-d8	10.49	136	96348	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	46108	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	110458	5.00	ng	0.00
25) Chrysene-d12	21.28	240	76793	5.00	ng	0.00
32) Perylene-d12	23.53	264	73442	5.00	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	30111	5.11	ng	0.00
5) Phenol-d6	6.86	99	37721	5.18	ng	0.00
7) Nitrobenzene-d5	8.86	82	34472	5.65	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	9805	9.92	ng	0.00
14) 2-Fluorobiphenyl	12.97	172	66465	4.81	ng	0.00
27) Terphenyl-d14	19.73	244	45944	4.84	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	11937	4.27	ng	96
3) n-Nitrosodimethylamine	3.51	42	19927	5.22	ng	98
8) Nitrobenzene	8.90	77	36724	5.67	ng	99
9) Naphthalene	10.52	128	96728	4.76	ng	99
10) Hexachlorobutadiene	10.82	225	16690	4.71	ng	100
11) 2-Methylnaphthalene	12.16	142	61935	4.85	ng	99
15) Acenaphthylene	14.06	152	103709	5.25	ng	100
16) Acenaphthene	14.41	154	60625	4.95	ng	96
17) Fluorene	15.39	166	51201	2.66	ng	99
19) 4-Bromophenyl-phenylether	16.27	248	19970	10.55	ng	95
20) Hexachlorobenzene	16.41	284	10930	2.79	ng	# 100
21) Pentachlorophenol	16.75	266	7524	4.39	ng	94
22) Phenanthrene	17.12	178	150170	7.14	ng	99
23) Anthracene	17.23	178	67466m	3.03	ng	
24) Fluoranthene	19.15	202	109117	5.49	ng	99
26) Pyrene	19.52	202	114037	4.89	ng	100
28) Benzo(a)anthracene	21.27	228	102410	4.93	ng	99
29) Chrysene	21.31	228	86788	4.43	ng	99
30) Bis(2-ethylhexyl)phthalate	21.21	149	79152	5.33	ng	99
31) Indeno(1,2,3-cd)pyrene	25.78	276	101115	4.67	ng	100
33) Benzo(b)fluoranthene	22.84	252	91771	4.42	ng	97
34) Benzo(k)fluoranthene	22.90	252	90332	4.88	ng	96
35) Benzo(a)pyrene	23.43	252	85778	4.78	ng	96
36) Dibenzo(a,h)anthracene	25.80	278	79369	4.48	ng	97
37) Benzo(g,h,i)perylene	26.48	276	82211	4.44	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060315\
Data File : BM001541.D
Acq On : 03 Jun 2015 19:00
Operator : TP/IZ
Sample : SSTDIC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampled :
SSTDIC005

Manual Integrations
APPROVED

apatel
6/5/2015 8:04:04 AM

Quant Time: Jun 04 04:00:53 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 04 03:14:13 2015
Response via : Initial Calibration

