


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060321\
 Data File : BM030276.D
 Acq On : 03 Jun 2021 09:04
 Operator : CG/JU
 Sample : DFTPP45
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP045

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM060121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 03 10:49:01 2021



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.2	306402	PASS
68	69	0.00	2	1.6	4838	PASS
69	198	0.00	100	38.9	304104	PASS
70	69	0.00	2	0.5	1410	PASS
127	198	10	80	47.8	373632	PASS
197	198	0.00	2	0.5	3721	PASS
198	198	100	100	100.0	781483	PASS
199	198	5	9	6.8	53381	PASS
275	198	10	60	27.4	214229	PASS
365	198	1	100	3.9	30685	PASS
441	198	0.01	100	11.0	85688	PASS
442	198	50	100	71.4	558059	PASS
443	442	15	24	18.9	105405	PASS

m/z	Abundance
36.00	215.0
38.00	188.0
39.00	213.0
40.00	1346.0
41.10	444.0
42.00	162.0
43.10	557.0
44.00	1348.0
47.00	155.0
49.00	525.0
51.00	265.0
55.10	324.0
56.00	157.0
57.00	292.0
67.00	171.0
68.90	237.0
70.10	165.0
73.00	1048.0
77.10	158.0
77.90	252.0
81.00	187.0
83.20	285.0
83.90	442.0
85.90	297.0
95.90	494.0
102.80	151.0
117.10	183.0
133.10	470.0
134.80	156.0
147.00	544.0
167.00	273.0
176.80	210.0
191.00	419.0
192.90	294.0
207.00	3706.0
208.00	808.0
209.10	542.0
221.10	173.0
267.00	218.0
281.10	586.0
282.00	169.0
282.80	164.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

m/z	Abundance
36.00	428.0
37.00	1721.0
38.00	4513.0
39.10	27128.0
40.00	2305.0
41.00	1373.0
42.00	250.0
43.10	491.0
44.00	1667.0
45.00	748.0
47.90	368.0
49.10	2891.0
50.00	101240.0
51.10	383872.0
52.00	18800.0
53.00	965.0
55.00	2025.0
56.00	11827.0
57.00	26384.0
58.00	1332.0
59.00	446.0
60.00	430.0
61.00	4834.0
62.00	5367.0
63.00	14551.0
64.00	2424.0
65.00	6942.0
65.90	665.0
67.00	733.0
68.00	6086.0
69.00	383232.0
70.00	2071.0
71.00	459.0
72.00	265.0
73.10	4588.0
74.00	38272.0
75.00	59256.0
76.10	21312.0
77.10	440000.0
78.10	32416.0
79.00	26552.0
80.00	21880.0
81.00	31504.0
82.00	7767.0
83.00	7101.0
84.00	1454.0
85.00	5446.0
86.00	8993.0
87.00	4482.0
88.00	1252.0
89.00	617.0
91.00	6916.0
92.00	7858.0
93.00	50944.0
94.00	3371.0
95.00	890.0
96.00	2767.0
97.10	971.0
98.00	37464.0
99.00	32168.0
100.00	2729.0
101.00	15909.0
101.90	788.0
103.00	6070.0
104.00	10218.0
105.00	10165.0
106.00	3272.0
107.00	128368.0
108.00	20896.0
109.10	3165.0
110.00	220800.0
111.00	36584.0
112.00	4593.0
112.90	1621.0
114.20	436.0
115.00	580.0
116.00	6639.0
117.00	118176.0
118.00	8541.0
119.00	1195.0
120.00	1961.0
121.00	558.0
122.00	8693.0
123.00	13391.0
124.00	6447.0
125.00	5373.0
126.10	1902.0
127.00	448704.0
128.00	35104.0
129.00	167808.0
130.00	14432.0
131.00	2806.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

131.90	1534.0
133.00	1153.0
134.00	4678.0
135.00	14067.0
136.00	5657.0
137.00	8440.0
138.00	1703.0
139.00	802.0
140.00	2402.0
141.00	21848.0
142.00	7230.0
143.00	5127.0
144.00	1630.0
145.00	1326.0
146.00	3540.0
147.00	12044.0
148.00	31200.0
149.00	6530.0
150.00	2065.0
151.10	2838.0
151.90	1525.0
153.00	7471.0
154.00	5277.0
155.00	12804.0
156.00	19144.0
157.00	3774.0
158.00	4329.0
159.00	3299.0
160.00	6904.0
161.00	10297.0
162.00	3394.0
163.00	1242.0
163.90	1035.0
165.00	9645.0
166.00	6821.0
167.00	50520.0
168.00	31704.0
169.00	4804.0
170.00	1574.0
171.10	1751.0
172.00	4617.0
173.00	5715.0
174.00	9632.0
175.10	18360.0
176.10	5020.0
177.00	8340.0
178.00	2341.0
179.00	35504.0
180.00	26816.0
181.00	11846.0
182.00	1890.0
183.00	942.0
184.00	3421.0
185.00	16003.0
186.00	125008.0
187.00	36032.0
188.00	3936.0
189.00	8583.0
190.00	1310.0
191.00	4071.0
192.00	11939.0
193.00	13111.0
194.00	2821.0
195.00	1265.0
196.00	25384.0
197.00	4362.0
198.00	887360.0
199.00	61416.0
200.00	4897.0
201.40	3559.0
203.00	7168.0
204.00	34784.0
205.00	59168.0
206.00	240128.0
207.00	35712.0
208.00	9551.0
209.00	3965.0
210.00	3629.0
211.10	9372.0
211.70	1628.0
213.00	743.0
214.10	291.0
215.00	2964.0
216.00	5740.0
217.00	68688.0
218.00	9243.0
219.00	783.0
221.00	35568.0
221.80	8546.0
223.00	14390.0
224.00	133952.0
225.00	34184.0
226.00	3928.0
227.00	57920.0
228.00	8223.0
229.00	13175.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

230.00	1432.0
231.00	5196.0
231.90	851.0
233.00	1079.0
234.00	3554.0
235.00	4378.0
236.00	2781.0
237.00	6195.0
238.00	909.0
238.90	1925.0
240.00	1879.0
241.00	3314.0
242.00	6865.0
243.00	6950.0
244.10	102512.0
245.00	13264.0
246.00	20200.0
247.00	4415.0
248.00	1105.0
249.00	4252.0
250.00	650.0
251.00	958.0
252.00	841.0
253.00	2282.0
254.00	3431.0
255.00	522880.0
256.00	77632.0
257.00	6872.0
258.00	31432.0
259.00	4801.0
260.00	1002.0
260.90	757.0
262.10	184.0
263.00	285.0
264.00	735.0
265.00	12149.0
266.00	2015.0
267.00	399.0
268.10	228.0
270.00	622.0
271.00	1277.0
272.00	1377.0
273.00	15033.0
274.00	38680.0
275.00	221504.0
276.00	30528.0
277.00	20272.0
278.00	4119.0
278.90	734.0
281.00	1059.0
281.90	760.0
283.00	2202.0
284.00	1358.0
285.00	3624.0
286.00	672.0
287.90	432.0
288.90	1051.0
290.00	717.0
290.90	701.0
292.00	938.0
293.00	4357.0
294.00	1508.0
295.10	1449.0
296.00	67664.0
297.00	9480.0
298.00	898.0
299.00	200.0
301.00	667.0
302.00	1007.0
303.00	7469.0
304.00	2203.0
305.00	303.0
308.00	823.0
309.00	685.0
310.00	896.0
310.90	222.0
313.00	470.0
314.00	2764.0
315.00	6856.0
316.00	3903.0
316.90	670.0
319.90	391.0
321.00	2014.0
322.00	827.0
323.00	18800.0
324.00	3940.0
325.00	383.0
326.00	267.0
327.00	4230.0
328.00	1971.0
328.90	474.0
332.00	1261.0
333.00	1717.0
334.00	11389.0
335.00	3012.0
336.10	467.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

338.90	354.0
340.00	273.0
341.00	2233.0
342.00	705.0
346.00	4661.0
347.00	1101.0
351.00	332.0
352.00	5590.0
353.00	3699.0
354.00	6018.0
355.10	1428.0
356.00	199.0
359.00	608.0
361.10	188.0
364.10	226.0
365.00	28248.0
366.00	4237.0
366.90	232.0
370.10	534.0
371.00	1131.0
372.00	9204.0
373.00	2265.0
373.80	255.0
376.90	313.0
383.00	3088.0
384.00	931.0
384.90	295.0
390.00	1273.0
391.10	920.0
392.10	584.0
394.90	160.0
397.00	156.0
400.90	767.0
402.00	3503.0
403.00	5211.0
404.00	1810.0
405.00	725.0
415.00	228.0
421.00	4276.0
422.00	4079.0
423.10	28944.0
424.00	9088.0
425.00	1198.0
425.90	187.0
439.20	151.0
441.10	75792.0
442.10	486976.0
443.10	93384.0
444.00	8516.0
444.90	581.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

m/z	Abundance
35.00	518.0
37.00	1701.0
38.10	4115.0
39.10	25792.0
40.00	2305.0
41.00	1100.0
43.00	384.0
44.00	1623.0
45.00	769.0
46.80	248.0
48.00	242.0
49.10	2738.0
50.00	92944.0
51.10	356992.0
52.10	17808.0
53.00	818.0
55.00	1676.0
56.00	10740.0
57.00	23880.0
58.00	1218.0
59.10	438.0
59.90	481.0
61.00	4424.0
62.10	4711.0
63.00	13305.0
64.00	1544.0
65.10	7059.0
66.00	381.0
67.00	620.0
68.00	5678.0
69.00	353344.0
70.00	1739.0
71.00	261.0
73.10	4168.0
74.00	36672.0
75.00	56896.0
76.10	19592.0
77.10	417024.0
78.10	29792.0
79.00	25304.0
80.00	19992.0
81.00	27896.0
82.00	7023.0
83.00	6359.0
84.00	1069.0
85.00	4921.0
86.00	8162.0
87.00	3634.0
88.00	1655.0
88.90	624.0
91.00	6027.0
92.00	7259.0
93.00	48208.0
94.00	3440.0
95.10	886.0
96.00	2594.0
97.00	772.0
98.00	36336.0
99.00	30920.0
100.00	2310.0
101.00	16063.0
102.00	1096.0
103.00	4883.0
104.00	9755.0
105.00	9662.0
106.10	3076.0
107.00	125776.0
108.00	18936.0
109.00	3383.0
110.00	214208.0
111.00	34312.0
112.00	4534.0
113.00	1253.0
114.10	219.0
115.00	613.0
116.00	6654.0
117.00	116320.0
118.00	8648.0
119.00	1224.0
120.00	1662.0
121.00	576.0
122.00	7641.0
123.00	12389.0
124.00	5948.0
125.00	4883.0
126.10	1881.0
127.00	441728.0
128.00	32456.0
129.00	167104.0
130.00	13882.0
131.00	2655.0
132.10	1277.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

133.00	809.0
134.00	4815.0
135.00	13931.0
136.00	5559.0
137.00	8165.0
138.00	1603.0
138.90	826.0
140.00	2025.0
141.00	22688.0
142.00	6524.0
143.00	4891.0
144.00	1358.0
145.00	1201.0
146.00	3578.0
147.00	12202.0
148.00	30288.0
149.00	6138.0
150.00	1734.0
151.10	2948.0
151.80	1677.0
153.00	7185.0
154.00	5414.0
155.00	13410.0
156.00	19408.0
157.00	3516.0
158.00	3950.0
159.00	3030.0
160.00	8025.0
161.00	10584.0
162.00	2969.0
163.00	1046.0
163.90	1149.0
165.00	8590.0
166.00	6738.0
167.00	49872.0
168.00	30296.0
169.00	4722.0
170.00	1414.0
171.00	2024.0
172.00	4513.0
173.00	5765.0
174.00	10399.0
175.00	19312.0
176.10	4508.0
177.00	8747.0
178.00	2856.0
179.00	37600.0
180.00	26992.0
181.00	11528.0
182.00	1852.0
183.10	846.0
184.00	2757.0
185.00	17744.0
186.00	135680.0
187.00	37752.0
188.10	4610.0
189.00	8972.0
190.00	1009.0
191.00	4262.0
192.00	12888.0
193.00	13312.0
194.00	2583.0
195.10	1253.0
196.00	26856.0
197.00	4076.0
198.00	937920.0
199.00	64088.0
200.00	5681.0
201.40	3592.0
203.00	6869.0
204.00	37232.0
205.00	62248.0
206.10	260416.0
207.10	39200.0
208.00	9693.0
209.10	3859.0
210.00	4781.0
211.00	11131.0
212.00	1689.0
213.00	784.0
214.00	242.0
215.10	2708.0
216.00	6021.0
217.00	72016.0
218.00	8909.0
219.00	840.0
221.00	39024.0
222.00	8599.0
223.00	16288.0
224.00	144832.0
225.00	35056.0
226.00	3977.0
227.00	63768.0
228.00	9345.0
229.00	13061.0
230.00	1496.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

231.00	5246.0
232.00	1065.0
233.00	1194.0
234.00	4426.0
235.00	4051.0
236.00	3131.0
237.00	6970.0
238.00	943.0
239.00	2619.0
240.00	1846.0
241.00	3645.0
242.00	8074.0
243.10	8898.0
244.00	115040.0
245.00	14779.0
246.00	23032.0
247.00	4735.0
248.00	1021.0
249.00	4085.0
250.00	812.0
251.00	1092.0
252.00	932.0
253.00	2287.0
254.10	3886.0
255.00	598784.0
256.00	88728.0
257.00	7976.0
258.00	35560.0
259.00	5528.0
260.00	937.0
261.00	1151.0
262.00	208.0
262.90	365.0
264.00	716.0
265.00	14491.0
266.00	2542.0
267.00	531.0
268.00	297.0
270.00	709.0
271.00	1396.0
272.00	1443.0
273.00	17696.0
274.00	45480.0
275.00	266176.0
276.00	35568.0
277.00	23544.0
278.00	4592.0
279.00	834.0
281.00	1122.0
282.00	886.0
283.00	2778.0
284.00	1846.0
285.00	3881.0
286.00	849.0
288.10	261.0
288.90	1154.0
290.00	690.0
291.00	502.0
291.90	1110.0
293.00	5493.0
294.00	1536.0
295.00	1727.0
296.00	81440.0
297.00	11681.0
298.00	870.0
300.00	153.0
301.00	1049.0
302.00	1382.0
303.00	8542.0
304.00	2755.0
304.90	315.0
308.00	1482.0
309.00	681.0
310.00	1240.0
310.90	287.0
313.00	570.0
314.00	3591.0
315.00	8836.0
316.00	4836.0
317.10	928.0
319.00	330.0
320.00	421.0
321.00	2214.0
322.00	1456.0
323.00	24256.0
324.00	4371.0
325.00	390.0
326.00	575.0
327.00	4281.0
328.00	2291.0
329.00	416.0
332.00	1901.0
333.00	2442.0
334.00	14634.0
335.00	3530.0
336.00	542.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

339.00	568.0
340.00	374.0
341.00	2818.0
342.00	1054.0
343.00	291.0
346.00	5832.0
347.00	1093.0
350.00	205.0
351.00	418.0
352.00	7400.0
353.00	5641.0
354.10	6896.0
355.00	1363.0
356.00	202.0
357.00	217.0
359.10	525.0
363.80	243.0
365.00	37944.0
366.00	5166.0
367.00	268.0
370.00	746.0
371.00	2051.0
372.00	12169.0
373.00	2684.0
374.00	401.0
377.00	369.0
383.00	3056.0
384.00	903.0
385.10	400.0
390.00	1870.0
391.00	1220.0
392.00	775.0
392.80	177.0
400.90	720.0
402.00	5208.0
403.00	7152.0
404.00	2642.0
405.00	435.0
415.00	306.0
421.00	5526.0
422.00	6051.0
423.10	39384.0
424.00	11223.0
425.10	1207.0
438.50	193.0
439.10	232.0
439.90	276.0
441.10	103048.0
442.10	676416.0
443.10	128744.0
444.10	11651.0
444.90	612.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

m/z	Abundance
36.00	381.0
37.00	905.0
38.10	2384.0
39.10	13263.0
40.00	1878.0
41.10	909.0
42.10	174.0
43.00	432.0
44.00	1639.0
45.00	486.0
47.00	192.0
48.00	190.0
49.00	1842.0
50.00	46016.0
51.10	179136.0
52.00	8897.0
53.00	438.0
55.00	1329.0
56.00	5711.0
57.00	11600.0
58.00	565.0
58.90	181.0
59.90	184.0
61.00	2342.0
62.00	2095.0
63.00	6811.0
64.00	1169.0
65.00	2817.0
66.00	358.0
67.00	405.0
68.00	2751.0
69.00	176448.0
70.00	915.0
70.90	291.0
72.00	163.0
73.10	2828.0
74.10	17792.0
75.00	29024.0
76.00	10411.0
77.10	216704.0
78.10	15058.0
79.00	12811.0
80.00	10143.0
81.00	15000.0
82.00	3832.0
83.00	3499.0
84.00	718.0
85.00	2516.0
86.00	4702.0
87.00	1701.0
88.00	733.0
88.90	450.0
91.00	2923.0
92.00	3973.0
93.00	24080.0
94.00	1370.0
94.90	528.0
96.10	1398.0
96.80	405.0
97.10	399.0
98.00	19424.0
99.00	15939.0
100.00	1133.0
101.00	8076.0
102.10	478.0
103.00	2656.0
104.00	5394.0
105.00	4921.0
106.00	1520.0
107.00	62768.0
108.00	10004.0
109.10	1540.0
110.00	109624.0
111.00	17696.0
112.00	2344.0
113.00	693.0
113.90	162.0
115.00	444.0
116.00	3677.0
117.00	60376.0
118.00	4255.0
118.90	622.0
119.90	816.0
121.10	367.0
122.00	3836.0
123.00	6396.0
124.00	3208.0
125.00	3016.0
126.00	863.0
127.00	230464.0
128.00	16728.0
129.00	86416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

130.00	6917.0
131.00	1571.0
132.10	884.0
132.90	571.0
134.00	2614.0
135.00	7390.0
136.00	2810.0
137.00	4283.0
137.80	668.0
138.00	683.0
138.90	692.0
140.10	1093.0
141.00	12785.0
142.00	3755.0
143.00	2494.0
144.00	962.0
145.00	726.0
146.00	1973.0
147.00	6365.0
148.00	16235.0
149.00	3372.0
150.00	882.0
151.10	1365.0
152.10	805.0
153.00	4324.0
154.00	2808.0
155.00	6473.0
156.00	10727.0
157.00	2154.0
158.00	1973.0
159.00	2144.0
160.00	3933.0
161.00	6022.0
162.10	1838.0
163.00	623.0
164.00	561.0
165.00	4820.0
166.00	3793.0
167.00	29272.0
168.00	16432.0
169.00	2406.0
170.00	920.0
171.00	1200.0
171.90	2548.0
173.00	3005.0
174.00	5528.0
175.00	10852.0
176.00	2927.0
177.00	4989.0
178.00	1652.0
179.00	19984.0
180.00	14279.0
181.00	6128.0
182.00	1208.0
183.10	505.0
184.00	1346.0
185.00	9745.0
186.00	74408.0
187.00	20080.0
188.00	2314.0
189.00	5017.0
190.00	978.0
191.00	3038.0
192.00	6756.0
193.00	7765.0
194.00	1402.0
195.00	828.0
196.00	15467.0
197.10	2726.0
198.00	519168.0
199.00	34640.0
200.00	2895.0
201.60	2001.0
203.00	3902.0
204.00	21144.0
205.00	35600.0
206.10	145152.0
207.10	22432.0
208.00	5674.0
209.00	2093.0
210.20	2724.0
211.00	6573.0
212.90	324.0
214.10	166.0
215.00	1526.0
216.00	2744.0
217.00	40904.0
218.00	5343.0
219.00	485.0
221.00	20880.0
221.70	4904.0
223.00	9052.0
224.00	82960.0
225.00	20120.0
226.00	2296.0
227.00	36472.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

228.00	4725.0
229.00	8016.0
229.90	951.0
231.00	3471.0
231.90	439.0
233.00	922.0
234.00	2115.0
235.10	2411.0
236.00	1647.0
237.10	3318.0
238.10	527.0
239.00	1663.0
240.00	1136.0
241.10	1990.0
242.00	5006.0
243.00	5112.0
244.10	67896.0
245.00	8068.0
246.00	12583.0
247.00	2502.0
248.10	636.0
249.00	2354.0
250.00	532.0
251.00	636.0
251.90	667.0
253.10	1319.0
254.10	2266.0
255.00	353216.0
256.00	50680.0
257.00	3877.0
258.00	21992.0
259.00	3388.0
259.90	414.0
261.00	506.0
262.90	157.0
263.90	589.0
265.00	9184.0
266.00	1234.0
266.90	497.0
268.10	153.0
270.00	332.0
271.00	811.0
272.00	811.0
273.00	10695.0
274.00	27352.0
275.00	155008.0
276.00	19216.0
277.00	14554.0
278.00	2754.0
279.00	447.0
281.00	933.0
282.00	613.0
283.10	1650.0
284.00	954.0
285.00	2558.0
286.10	233.0
288.90	565.0
289.90	480.0
291.00	366.0
292.00	806.0
293.00	3028.0
294.00	974.0
295.00	1091.0
296.00	49552.0
297.00	6097.0
298.00	463.0
301.00	521.0
302.00	879.0
303.00	5650.0
304.10	1307.0
305.00	279.0
307.80	740.0
309.00	419.0
309.90	788.0
311.00	169.0
313.00	670.0
314.00	2550.0
314.90	5721.0
316.00	3256.0
317.00	521.0
321.00	1916.0
322.00	500.0
323.00	14188.0
324.00	2629.0
324.90	228.0
326.00	271.0
327.00	2904.0
328.00	1432.0
329.10	377.0
330.90	166.0
332.00	1193.0
332.90	1652.0
334.00	9859.0
335.00	2494.0
336.00	377.0
339.00	342.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045

340.00	222.0
341.00	1742.0
342.10	452.0
345.90	3726.0
347.00	872.0
350.00	258.0
351.00	337.0
352.10	5532.0
353.00	2932.0
354.00	4734.0
355.00	989.0
357.10	221.0
359.10	353.0
365.00	25864.0
366.00	3164.0
367.00	256.0
370.00	372.0
371.00	1471.0
372.00	8773.0
373.00	1871.0
374.10	199.0
376.80	248.0
383.00	2600.0
384.00	729.0
384.90	213.0
390.10	959.0
391.00	611.0
392.00	720.0
401.00	616.0
402.00	3462.0
403.00	5296.0
404.00	1849.0
405.10	284.0
415.10	196.0
421.00	4217.0
422.00	4158.0
423.10	27704.0
424.00	7424.0
425.00	1234.0
440.10	165.0
441.10	78224.0
442.10	510784.0
443.10	94088.0
444.00	7836.0
445.00	335.0

Instrument :
BNA_M
ClientSampleId :
DFTPP045