

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060424\
 Data File : BM045846.D
 Acq On : 05 Jun 2024 12:18
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020114

Quant Time: Jun 06 01:17:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM053124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 31 13:01:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.463	152	287684	20.000	ng/u1	0.00
20) Naphthalene-d8	10.228	136	1173100	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.104	164	668839	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.851	188	1164139	20.000	ng/u1	0.00
79) Chrysene-d12	21.068	240	876962	20.000	ng/u1	0.00
88) Perylene-d12	23.786	264	1002731	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.052	96	54926	7.402	ng/uL	0.00
4) Pyridine-d5	3.458	84	385510	20.178	ng/u1	-0.01
7) Phenol-d5	6.716	99	511156	20.915	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.828	67	361577	20.603	ng/u1	0.00
11) 2-Chlorophenol-d4	7.028	132	394025	21.584	ng/u1	0.00
15) 4-Methylphenol-d8	8.234	113	391819	21.221	ng/u1	0.00
21) Nitrobenzene-d5	8.622	128	197447	21.527	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.339	143	212882	22.933	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.886	165	384661	22.097	ng/u1	0.00
31) 4-Chloroaniline-d4	10.410	131	490898	20.376	ng/u1	-0.01
46) Dimethylphthalate-d6	13.551	166	993348	21.453	ng/u1	0.00
49) Acenaphthylene-d8	13.798	160	1161536	21.602	ng/u1	0.00
54) 4-Nitrophenol-d4	14.404	143	200531	20.067	ng/u1	0.00
60) Fluorene-d10	15.104	176	809552	20.494	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.245	200	104334	17.551	ng/u1	0.00
73) Anthracene-d10	16.951	188	1090018	21.538	ng/u1	0.00
81) Pyrene-d10	19.251	212	1127295	20.670	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.603	264	998079	21.297	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.081	88	61001	7.428	ng/uL	95
5) Pyridine	3.475	79	395552	20.159	ng/u1	99
6) Benzaldehyde	6.634	77	262990	22.879	ng/u1	99
8) Phenol	6.746	94	528073	20.835	ng/u1	96
10) Bis(2-Chloroethyl)ether	6.916	93	424086	20.046	ng/u1	98
12) 2-Chlorophenol	7.057	128	404294	21.022	ng/u1	100
13) 2-Methylphenol	7.963	108	389999	21.013	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	7.998	45	882450	21.178	ng/u1	99
16) Acetophenone	8.298	105	638558	20.751	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.287	70	356851	20.986	ng/u1	97
18) 4-Methylphenol	8.292	108	408243	21.258	ng/u1	100
19) Hexachloroethane	8.528	117	186791	20.221	ng/u1	98
22) Nitrobenzene	8.663	77	553658	21.569	ng/u1	99
23) Isophorone	9.192	82	973242	21.457	ng/u1	99
25) 2-Nitrophenol	9.369	139	225685	22.241	ng/u1	98
26) 2,4-Dimethylphenol	9.469	107	446387	21.684	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.675	93	570369	20.713	ng/u1	99
29) 2,4-Dichlorophenol	9.910	162	374231	21.578	ng/u1	97
30) Naphthalene	10.275	128	1304513	20.854	ng/u1	99
32) 4-Chloroaniline	10.434	127	481609	20.451	ng/u1	97
33) Hexachlorobutadiene	10.569	225	249334	20.986	ng/u1	99
34) Caprolactam	11.251	113	56425	11.551	ng/u1	93
35) 4-Chloro-3-methylphenol	11.586	107	399454	21.230	ng/u1	100
36) 2-Methylnaphthalene	11.898	142	857793	20.705	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060424\
 Data File : BM045846.D
 Acq On : 05 Jun 2024 12:18
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 37 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020114


Quant Time: Jun 06 01:17:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM053124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 31 13:01:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.122	142	848901	20.495	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.275	216	431490	21.541	ng/ul	97
40) Hexachlorocyclopentadiene	12.263	237	164512	13.532	ng/ul	100
41) 2,4,6-Trichlorophenol	12.545	196	276805	23.751	ng/ul	94
42) 2,4,5-Trichlorophenol	12.633	196	282908	22.370	ng/ul	99
43) 1,1'-Biphenyl	12.939	154	1062803	21.418	ng/ul	99
44) 2-Chloronaphthalene	12.969	162	836698	21.517	ng/ul	98
45) 2-Nitroaniline	13.216	65	295935	22.287	ng/ul	98
47) Dimethylphthalate	13.604	163	1007003	21.418	ng/ul	100
48) 2,6-Dinitrotoluene	13.710	165	196846	21.763	ng/ul	95
50) Acenaphthylene	13.827	152	1331064	21.389	ng/ul	99
51) 3-Nitroaniline	14.057	138	186502	22.455	ng/ul	97
52) Acenaphthene	14.169	153	879407	21.177	ng/ul	99
53) 2,4-Dinitrophenol	14.245	184	87207	21.175	ng/ul	98
55) 4-Nitrophenol	14.416	109	167378	20.302	ng/ul	99
56) Dibenzofuran	14.510	168	1176420	21.007	ng/ul	100
57) 2,4-Dinitrotoluene	14.498	165	267759	21.397	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.757	232	216602	22.667	ng/ul	93
59) Diethylphthalate	14.969	149	1003263	21.310	ng/ul	99
61) Fluorene	15.157	166	938450	20.798	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.169	204	462391	20.646	ng/ul	100
63) 4-Nitroaniline	15.233	138	165667	24.021	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.257	198	114879	17.607	ng/ul#	91
67) N-Nitrosodiphenylamine	15.386	169	753377	22.837	ng/ul	99
68) 4-Bromophenyl-phenylether	16.057	248	268779	22.145	ng/ul	99
69) Hexachlorobenzene	16.163	284	287189	21.396	ng/ul	93
70) Atrazine	16.357	200	207945	20.794	ng/ul	99
71) Pentachlorophenol	16.527	266	155596	22.846	ng/ul	99
72) Phenanthrene	16.892	178	1302548	21.235	ng/ul	99
74) Anthracene	16.986	178	1314859	21.460	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.892	216	422864	23.422	ng/uL	100
76) Pentachlorobenzene	14.427	250	383560	22.714	ng/uL	98
77) Carbazole	17.274	167	1136048	21.731	ng/ul	99
78) Di-n-butylphthalate	17.845	149	1562414	22.726	ng/ul	100
80) Fluoranthene	18.915	202	1382135	20.905	ng/ul	99
82) Pyrene	19.280	202	1406532	20.706	ng/ul	99
83) Butylbenzylphthalate	20.209	149	634774	22.929	ng/ul	100
84) 3,3'-Dichlorobenzidine	20.998	252	366705	21.170	ng/ul	99
85) Benzo(a)anthracene	21.050	228	1296932	21.418	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.003	149	948842	23.068	ng/ul	99
87) Chrysene	21.103	228	1186025	21.173	ng/ul	99
89) Di-n-octyl phthalate	22.033	149	1616665	21.081	ng/ul	100
90) Benzo(b)fluoranthene	22.939	252	1286355	20.875	ng/ul	99
91) Benzo(k)fluoranthene	22.992	252	1252860	20.816	ng/ul	98
93) Benzo(a)pyrene	23.662	252	1171055	21.626	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.791	276	1324081	22.391	ng/ul	99
95) Dibenzo(a,h)anthracene	26.838	278	1049765	22.445	ng/ul	100
96) Benzo(g,h,i)perylene	27.727	276	1131091	21.717	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 06 01:17:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM053124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 31 13:01:11 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.428 ng/uL
RT: 3.081 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion: 88 Resp: 61001
Ion Ratio Lower Upper
88 100
43 45.1 35.1 52.7
58 90.5 67.4 101.2



#5
Pyridine
Concen: 20.159 ng/uL
RT: 3.475 min Scan# 83
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 79 Resp: 395552
Ion Ratio Lower Upper
79 100
52 113.5 90.5 135.7
51 54.1 41.5 62.3









#6
Benzaldehyde
Concen: 22.879 ng/ul
RT: 6.634 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion: 77 Resp: 262990
Ion Ratio Lower Upper
77 100
105 90.9 72.7 109.1
106 86.1 70.8 106.2



#8
Phenol
Concen: 20.835 ng/ul
RT: 6.746 min Scan# 639
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 94 Resp: 528073
Ion Ratio Lower Upper
94 100
65 35.2 26.8 40.2
66 46.6 35.1 52.7









#10
Bis(2-Chloroethyl)ether
Concen: 20.046 ng/ul
RT: 6.916 min Scan# 668
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion: 93 Resp: 424086
Ion Ratio Lower Upper
93 100
63 92.1 71.8 107.8
95 32.2 25.4 38.2



#12
2-Chlorophenol
Concen: 21.022 ng/ul
RT: 7.057 min Scan# 692
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 128 Resp: 404294
Ion Ratio Lower Upper
128 100
64 58.0 46.1 69.1
130 32.2 25.8 38.8









#13
2-Methylphenol
Concen: 21.013 ng/ul
RT: 7.963 min Scan# 846
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:108 Resp: 389999
Ion Ratio Lower Upper
108 100
107 88.9 74.8 112.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.178 ng/ul
RT: 7.998 min Scan# 852
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 45 Resp: 882450
Ion Ratio Lower Upper
45 100
77 11.3 8.7 13.1
79 7.9 6.6 10.0









#16
Acetophenone
Concen: 20.751 ng/ul
RT: 8.298 min Scan# 903
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:105 Resp: 638558
Ion Ratio Lower Upper
105 100
77 105.9 84.7 127.1
51 51.8 39.4 59.0



#17
N-Nitroso-di-n-propylamine
Concen: 20.986 ng/ul
RT: 8.287 min Scan# 901
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 70 Resp: 356851
Ion Ratio Lower Upper
70 100
42 83.7 64.6 96.8
101 9.6 7.0 10.4
130 19.0 14.5 21.7





#18

4-Methylphenol

Concen: 21.258 ng/ul

RT: 8.292 min Scan# 902

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :





SSTD020114

Tgt Ion:108 Resp: 408243

Ion Ratio Lower Upper

108 100

107 113.3 90.6 136.0



#19

Hexachloroethane

Concen: 20.221 ng/ul

RT: 8.528 min Scan# 942

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:117 Resp: 186791


Ion Ratio Lower Upper

117 100

201 94.2 75.0 112.6

199 60.3 45.0 67.4









#22
Nitrobenzene
Concen: 21.569 ng/ul
RT: 8.663 min Scan# 90
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion: 77 Resp: 553658
Ion Ratio Lower Upper
77 100
123 39.2 31.2 46.8
65 15.3 11.8 17.8



#23
Isophorone
Concen: 21.457 ng/ul
RT: 9.192 min Scan# 1055
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion: 82 Resp: 973242
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 15.2 12.6 18.8









#25
2-Nitrophenol
Concen: 22.241 ng/ul
RT: 9.369 min Scan# 1102
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:139 Resp: 225685
Ion Ratio Lower Upper
139 100
65 68.8 55.8 83.8
109 35.7 29.6 44.4



#26
2,4-Dimethylphenol
Concen: 21.684 ng/ul
RT: 9.469 min Scan# 1102
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:107 Resp: 446387
Ion Ratio Lower Upper
107 100
121 47.2 37.4 56.0
122 81.2 63.1 94.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.713 ng/ul
 RT: 9.675 min Scan# 1137
 Delta R.T. -0.012 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020114


Tgt Ion: 93 Resp: 570369
 Ion Ratio Lower Upper
 93 100
 95 31.4 25.6 38.4
 123 11.2 8.6 13.0



#29
 2,4-Dichlorophenol
 Concen: 21.578 ng/ul
 RT: 9.910 min Scan# 1177
 Delta R.T. -0.006 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18

Tgt Ion: 162 Resp: 374231
 Ion Ratio Lower Upper
 162 100
 164 65.7 51.5 77.3
 98 39.0 29.0 43.6









#30
Naphthalene
Concen: 20.854 ng/ul
RT: 10.275 min Scan# 1239
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:128 Resp: 1304513
Ion Ratio Lower Upper
128 100
129 11.0 8.3 12.5
127 12.6 10.3 15.5



#32
4-Chloroaniline
Concen: 20.451 ng/ul
RT: 10.434 min Scan# 1266
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:127 Resp: 481609
Ion Ratio Lower Upper
127 100
129 32.6 24.6 37.0





#33

Hexachlorobutadiene

Concen: 20.986 ng/ul

RT: 10.569 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:225 Resp: 249334

Ion Ratio Lower Upper

225 100

223 63.0 49.4 74.0

227 63.1 51.1 76.7



#34

Caprolactam

Concen: 11.551 ng/ul

RT: 11.251 min Scan# 1405

Delta R.T. -0.029 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:113 Resp: 56425


Ion Ratio Lower Upper

113 100

55 235.4 179.5 269.3

56 171.4 130.1 195.1









#35
4-Chloro-3-methylphenol
Concen: 21.230 ng/ul
RT: 11.586 min Scan# 1462
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:107 Resp: 399454
Ion Ratio Lower Upper
107 100
144 24.8 20.1 30.1
142 76.5 61.4 92.0



#36
2-Methylnaphthalene
Concen: 20.705 ng/ul
RT: 11.898 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:142 Resp: 857793
Ion Ratio Lower Upper
142 100
141 89.0 71.1 106.7









#37
1-Methylnaphthalene
Concen: 20.495 ng/ul
RT: 12.122 min Scan# 1553
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:142 Resp: 848901
Ion Ratio Lower Upper
142 100
141 91.7 73.5 110.3
116 4.2 3.3 4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.541 ng/ul
RT: 12.275 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:216 Resp: 431490
Ion Ratio Lower Upper
216 100
214 79.6 61.3 91.9
179 21.2 17.1 25.7
108 20.4 18.8 28.2





#40

Hexachlorocyclopentadiene

Concen: 13.532 ng/ul

RT: 12.263 min Scan# 1577

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:237 Resp: 164512

Ion Ratio Lower Upper

237 100

235 61.5 49.2 73.8

272 12.5 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 23.751 ng/ul

RT: 12.545 min Scan# 1625

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:196 Resp: 276805


Ion Ratio Lower Upper

196 100

198 92.8 78.2 117.4

200 29.1 26.5 39.7





#42

2,4,5-Trichlorophenol

Concen: 22.370 ng/ul

RT: 12.633 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:196 Resp: 282908

Ion Ratio Lower Upper

196 100

198 95.5 77.6 116.4

200 31.9 25.0 37.6



#43

1,1'-Biphenyl

Concen: 21.418 ng/ul

RT: 12.939 min Scan# 1692

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:154 Resp: 1062803


Ion Ratio Lower Upper

154 100

153 40.1 31.6 47.4

76 14.3 11.4 17.2





#44

2-Chloronaphthalene

Concen: 21.517 ng/ul

RT: 12.969 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:162 Resp: 836698

Ion Ratio Lower Upper

162 100

164 33.2 26.2 39.4

127 41.3 31.9 47.9



#45

2-Nitroaniline

Concen: 22.287 ng/ul

RT: 13.216 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion: 65 Resp: 295935


Ion Ratio Lower Upper

65 100

92 60.0 48.4 72.6

138 85.7 66.3 99.5









#47
Dimethylphthalate
Concen: 21.418 ng/ul
RT: 13.604 min Scan# 1805
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:163 Resp: 1007003
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 9.8 7.8 11.6



#48
2,6-Dinitrotoluene
Concen: 21.763 ng/ul
RT: 13.710 min Scan# 1823
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:165 Resp: 196846
Ion Ratio Lower Upper
165 100
89 74.0 55.7 83.5
121 24.0 17.8 26.8









#50
Acenaphthylene
Concen: 21.389 ng/ul
RT: 13.827 min Scan# 1843
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:152 Resp: 1331064
Ion Ratio Lower Upper
152 100
151 19.1 15.6 23.4
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 22.455 ng/ul
RT: 14.057 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:138 Resp: 186502
Ion Ratio Lower Upper
138 100
108 10.0 7.9 11.9
92 125.3 103.1 154.7





#52

Acenaphthene

Concen: 21.177 ng/ul

RT: 14.169 min Scan# 1901

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:153 Resp: 879407

Ion Ratio Lower Upper

153 100

152 47.3 38.6 58.0

154 87.8 70.7 106.1



#53

2,4-Dinitrophenol

Concen: 21.175 ng/ul

RT: 14.245 min Scan# 1914

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:184 Resp: 87207


Ion Ratio Lower Upper

184 100

63 98.0 76.4 114.6

154 67.8 52.9 79.3









#55
4-Nitrophenol
Concen: 20.302 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:109 Resp: 167378
Ion Ratio Lower Upper
109 100
139 98.8 79.0 118.6
65 132.6 104.3 156.5



#56
Dibenzofuran
Concen: 21.007 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:168 Resp: 1176420
Ion Ratio Lower Upper
168 100
139 40.2 32.2 48.2









#57
 2,4-Dinitrotoluene
 Concen: 21.397 ng/ul
 RT: 14.498 min Scan# 1957
 Delta R.T. -0.006 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020114


Tgt Ion:165 Resp: 267759
 Ion Ratio Lower Upper
 165 100
 63 77.7 61.9 92.9
 182 2.7 2.2 3.2



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 22.667 ng/ul
 RT: 14.757 min Scan# 2001
 Delta R.T. -0.006 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18

Tgt Ion:232 Resp: 216602
 Ion Ratio Lower Upper
 232 100
 131 55.9 39.0 58.6
 130 3.2 2.2 3.4
 166 32.8 25.7 38.5









#59
Diethylphthalate
Concen: 21.310 ng/ul
RT: 14.969 min Scan# 2037
Delta R.T. -0.012 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:149 Resp: 1003263
Ion Ratio Lower Upper
149 100
177 21.1 17.1 25.7
150 12.4 9.6 14.4



#61
Fluorene
Concen: 20.798 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:166 Resp: 938450
Ion Ratio Lower Upper
166 100
165 98.3 78.1 117.1
167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.646 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:204 Resp: 462391
Ion Ratio Lower Upper
204 100
206 32.3 26.2 39.2
141 62.5 50.2 75.2



#63
4-Nitroaniline
Concen: 24.021 ng/ul
RT: 15.233 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:138 Resp: 165667
Ion Ratio Lower Upper
138 100
92 58.9 47.2 70.8
108 107.6 97.4 146.0









#66
4,6-Dinitro-2-methylphenol
Concen: 17.607 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:198 Resp: 114879
Ion Ratio Lower Upper
198 100
182 5.6 3.4 5.0#
77 36.4 24.7 37.1



#67
N-Nitrosodiphenylamine
Concen: 22.837 ng/ul
RT: 15.386 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:169 Resp: 753377
Ion Ratio Lower Upper
169 100
168 67.2 53.6 80.4
167 37.3 30.4 45.6









#68
4-Bromophenyl-phenylether
Concen: 22.145 ng/ul
RT: 16.057 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18




Instrument :
BNA_M
ClientSampleId :
SSTD020114


Tgt Ion:248 Resp: 268779
Ion Ratio Lower Upper
248 100
250 98.3 79.3 118.9
141 82.8 65.8 98.6



#69
Hexachlorobenzene
Concen: 21.396 ng/ul
RT: 16.163 min Scan# 2240
Delta R.T. 0.000 min
Lab File: BM045846.D
Acq: 05 Jun 2024 12:18

Tgt Ion:284 Resp: 287189
Ion Ratio Lower Upper
284 100
142 38.5 35.4 53.0
249 29.2 25.8 38.8





#70

Atrazine

Concen: 20.794 ng/ul

RT: 16.357 min Scan# 21

Delta R.T. -0.012 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:200 Resp: 207945

Ion Ratio Lower Upper

200 100

173 26.2 20.6 31.0

215 46.7 36.7 55.1



#71

Pentachlorophenol

Concen: 22.846 ng/ul

RT: 16.527 min Scan# 2302

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:266 Resp: 155596


Ion Ratio Lower Upper

266 100

264 59.8 48.2 72.2

268 64.0 50.8 76.2





#72

Phenanthrene

Concen: 21.235 ng/ul

RT: 16.892 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:178 Resp: 1302548

Ion Ratio Lower Upper

178 100

179 15.6 12.4 18.6

176 19.9 15.3 22.9



#74

Anthracene

Concen: 21.460 ng/ul

RT: 16.986 min Scan# 2380

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:178 Resp: 1314859


Ion Ratio Lower Upper

178 100

179 15.3 12.0 18.0

176 19.1 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 23.422 ng/uL

RT: 12.892 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:216 Resp: 422864

Ion Ratio Lower Upper

216 100

214 79.2 63.2 94.8

218 47.7 38.2 57.2



#76

Pentachlorobenzene

Concen: 22.714 ng/uL

RT: 14.427 min Scan# 1945

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:250 Resp: 383560


Ion Ratio Lower Upper

250 100

248 63.3 50.2 75.2

252 64.3 49.5 74.3









#77
 Carbazole
 Concen: 21.731 ng/ul
 RT: 17.274 min Scan# 2429
 Delta R.T. -0.006 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020114


Tgt Ion:167 Resp: 1136048
 Ion Ratio Lower Upper
 167 100
 166 21.1 16.6 24.8
 139 13.7 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 22.726 ng/ul
 RT: 17.845 min Scan# 2526
 Delta R.T. -0.006 min
 Lab File: BM045846.D
 Acq: 05 Jun 2024 12:18

Tgt Ion:149 Resp: 1562414
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.1 10.7
 104 6.2 4.9 7.3





#80

Fluoranthene

Concen: 20.905 ng/ul

RT: 18.915 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:202 Resp: 1382135

Ion Ratio Lower Upper

202 100

101 11.3 8.9 13.3

100 8.5 6.6 10.0



#82

Pyrene

Concen: 20.706 ng/ul

RT: 19.280 min Scan# 2770

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:202 Resp: 1406532


Ion Ratio Lower Upper

202 100

101 13.2 10.5 15.7

100 11.2 8.6 12.8





#83

Butylbenzylphthalate

Concen: 22.929 ng/ul

RT: 20.209 min Scan# 2928

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:149 Resp: 634774

Ion Ratio Lower Upper

149 100

91 79.8 63.7 95.5

206 19.4 15.9 23.9



#84

3,3'-Dichlorobenzidine

Concen: 21.170 ng/ul

RT: 20.998 min Scan# 3062

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:252 Resp: 366705


Ion Ratio Lower Upper

252 100

254 63.9 51.4 77.2

126 12.7 9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.418 ng/ul

RT: 21.050 min Scan# 3070

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:228 Resp: 1296932

Ion Ratio Lower Upper

228 100

229 19.3 15.8 23.8

226 26.6 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.068 ng/ul

RT: 21.003 min Scan# 3063

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:149 Resp: 948842


Ion Ratio Lower Upper

149 100

167 26.1 20.4 30.6

279 3.8 3.1 4.7





#87

Chrysene

Concen: 21.173 ng/ul

RT: 21.103 min Scan# 3080

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:228 Resp: 1186025

Ion Ratio Lower Upper

228 100

226 30.0 24.2 36.2

229 19.4 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 21.081 ng/ul




RT: 22.033 min Scan# 3238


Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Tgt Ion:149 Resp: 1616665





#90

Benzo(b)fluoranthene

Concen: 20.875 ng/ul

RT: 22.939 min Scan# 31

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:252 Resp: 1286355

Ion Ratio Lower Upper

252 100

253 21.4 17.8 26.6

125 10.4 8.4 12.6



#91

Benzo(k)fluoranthene

Concen: 20.816 ng/ul

RT: 22.992 min Scan# 3401

Delta R.T. -0.006 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:252 Resp: 1252860


Ion Ratio Lower Upper

252 100

253 22.1 16.8 25.2

125 10.5 8.2 12.4





#93

Benzo(a)pyrene

Concen: 21.626 ng/ul

RT: 23.662 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:252 Resp: 1171055

Ion Ratio Lower Upper

252 100

253 21.0 17.5 26.3

125 11.3 9.3 13.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.391 ng/ul

RT: 26.791 min Scan# 4047

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18




Tgt Ion:276 Resp: 1324081


Ion Ratio Lower Upper

276 100

138 23.2 17.9 26.9

277 24.1 19.0 28.6





#95

Dibenzo(a,h)anthracene

Concen: 22.445 ng/ul

RT: 26.838 min Scan# 4054

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Instrument :

BNA_M

ClientSampleId :

SSTD020114





Tgt Ion:278 Resp: 1049765

Ion Ratio Lower Upper

278 100

139 16.7 13.2 19.8

279 23.8 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 21.717 ng/ul

RT: 27.727 min Scan# 4206

Delta R.T. 0.000 min

Lab File: BM045846.D

Acq: 05 Jun 2024 12:18

Tgt Ion:276 Resp: 1131091

Ion Ratio Lower Upper

276 100

138 22.6 18.6 27.8

277 23.4 18.7 28.1

