

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060424\
 Data File : BM045862.D
 Acq On : 05 Jun 2024 23:34
 Operator : MA/JU
 Sample : SP6526
 Misc : SFAM SURROGATES
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP6526

Quant Time: Jun 06 02:20:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM053124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 31 13:01:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.457	152	264401	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.228	136	1063289	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.104	164	597513	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.851	188	1033965	20.000	ng/u1	0.00
79) Chrysene-d12	21.062	240	816318	20.000	ng/u1	0.00
88) Perylene-d12	23.785	264	915947	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.052	96	106928	15.680	ng/uL	0.00
4) Pyridine-d5	3.452	84	1546711	88.085	ng/u1	-0.02
7) Phenol-d5	6.716	99	1918769	85.424	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.822	67	1360121	84.327	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.022	132	1543395	91.989	ng/u1	-0.01
15) 4-Methylphenol-d8	8.228	113	1510762	89.027	ng/u1	0.00
21) Nitrobenzene-d5	8.628	128	772844	92.963	ng/u1	0.00
24) 2-Nitrophenol-d4	9.339	143	851724	101.228	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.886	165	1380499	87.495	ng/u1	0.00
31) 4-Chloroaniline-d4	10.404	131	1965221	89.996	ng/u1	-0.02
46) Dimethylphthalate-d6	13.551	166	3642604	88.058	ng/u1	0.00
49) Acenaphthylene-d8	13.798	160	4242805	88.327	ng/u1	0.00
54) 4-Nitrophenol-d4	14.404	143	557349	62.432	ng/u1	0.00
60) Fluorene-d10	15.104	176	2955365	83.747	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.245	200	379987	71.968	ng/u1	0.00
73) Anthracene-d10	16.951	188	4083202	90.841	ng/u1	0.00
81) Pyrene-d10	19.250	212	4267820	84.069	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.603	264	3825075	89.352	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060424\
Data File : BM045862.D
Acq On : 05 Jun 2024 23:34
Operator : MA/JU
Sample : SP6526
Misc : SFAM SURROGATES
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SP6526

Quant Time: Jun 06 02:20:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM053124.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 31 13:01:11 2024
Response via : Initial Calibration

