

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
 Data File : BM001580.D
 Acq On : 06 Jun 2015 06:18
 Operator : TP/IZ
 Sample : PB83718BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB83718BL

Quant Time: Jun 08 05:46:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 05:27:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	32402	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	114812	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	54286	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	79321	5.00	ng	0.00
25) Chrysene-d12	21.28	240	83521	5.00	ng	0.00
32) Perylene-d12	23.51	264	71229	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.27	112	90823	12.69	ng	0.00
5) Phenol-d6	6.84	99	116664	13.26	ng	0.00
7) Nitrobenzene-d5	8.85	82	70585	9.03	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	35115	13.51	ng	0.00
14) 2-Fluorobiphenyl	12.97	172	150520	9.00	ng	0.01
27) Terphenyl-d14	19.72	244	101385	9.40	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
Data File : BM001580.D
Acq On : 06 Jun 2015 06:18
Operator : TP/IZ
Sample : PB83718BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB83718BL

Quant Time: Jun 08 05:46:49 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 05:27:25 2015
Response via : Initial Calibration

