

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
 Data File : BM001586.D
 Acq On : 06 Jun 2015 09:52
 Operator : TP/IZ
 Sample : G2491-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SS043MW007-150601

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:48:21 PM

Quant Time: Jun 08 13:44:08 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 05:27:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	25001	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	87670	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	40609	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	51062	5.00	ng	0.00
25) Chrysene-d12	21.28	240	68942	5.00	ng	0.00
32) Perylene-d12	23.52	264	63607	5.00	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	14563	2.64	ng	0.02
5) Phenol-d6	6.84	99	16734	2.47	ng	0.00
7) Nitrobenzene-d5	8.85	82	32120	5.38	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	25416	13.07	ng	0.00
14) 2-Fluorobiphenyl	12.96	172	74424	5.95	ng	0.00
27) Terphenyl-d14	19.72	244	69006	7.75	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	26923	12.50	ng	97
9) Naphthalene	10.52	128	1373	0.07	ng	# 79
11) 2-Methylnaphthalene	12.15	142	3209	0.27	ng	97
15) Acenaphthylene	14.06	152	384	0.02	ng	# 55
16) Acenaphthene	14.41	154	1903	0.17	ng	89
21) Pentachlorophenol	16.75	266	501	0.64	ng	93
22) Phenanthrene	17.12	178	4084	0.23	ng	# 90
23) Anthracene	17.19	178	1280m	0.17	ng	
24) Fluoranthene	19.15	202	1954	0.14	ng	# 86
26) Pyrene	19.52	202	1609	0.07	ng	# 48
30) Bis(2-ethylhexyl)phthalate	21.19	149	8281	0.73	ng	98
31) Indeno(1,2,3-cd)pyrene	25.77	276	1379	0.07	ng	98
36) Dibenzo(a,h)anthracene	25.79	278	1229	0.08	ng	# 58
37) Benzo(g,h,i)perylene	26.46	276	1080	0.06	ng	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
Data File : BM001586.D
Acq On : 06 Jun 2015 09:52
Operator : TP/IZ
Sample : G2491-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SS043MW007-150601

Manual Integrations
APPROVED

apatel
6/8/2015 2:48:21 PM

Quant Time: Jun 08 13:44:08 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 05:27:25 2015
Response via : Initial Calibration

