

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
 Data File : BM001591.D
 Acq On : 06 Jun 2015 12:52
 Operator : TP/IZ
 Sample : G2491-06
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 LF016MW027-150601

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:48:43 PM

Quant Time: Jun 08 06:30:55 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 05:27:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	26318	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	92425	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	41341	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	56870	5.00	ng	0.00
25) Chrysene-d12	21.28	240	67392	5.00	ng	0.00
32) Perylene-d12	23.52	264	62589	5.00	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	15050	2.59	ng	0.02
5) Phenol-d6	6.86	99	11705	1.64	ng	0.02
7) Nitrobenzene-d5	8.85	82	35860	5.70	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	23474	11.86	ng	0.00
14) 2-Fluorobiphenyl	12.96	172	78499	6.16	ng	0.00
27) Terphenyl-d14	19.72	244	69081	7.94	ng	0.00
Target Compounds						
17) Fluorene	15.42	166	999m	0.14	ng	Qvalue
30) Bis(2-ethylhexyl)phthalate	21.19	149	128475	10.05	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
Data File : BM001591.D
Acq On : 06 Jun 2015 12:52
Operator : TP/IZ
Sample : G2491-06
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
LF016MW027-150601

Manual Integrations
APPROVED

apatel
6/8/2015 2:48:43 PM

Quant Time: Jun 08 06:30:55 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 05:27:25 2015
Response via : Initial Calibration

