

Data Path : Z:\HPCHEM1\BNA_M\Data\BM060615\
 Data File : BM001579.D
 Acq On : 06 Jun 2015 05:42
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 08 05:43:25 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 05:27:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.523	0.471	9.9	102	0.00
3	n-Nitrosodimethylamine	0.763	0.742	2.8	103	0.00
4 S	2-Fluorophenol	1.104	1.052	4.7	102	0.00
5 S	Phenol-d6	1.357	1.302	4.1	103	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
7 S	Nitrobenzene-d5	0.341	0.321	5.9	102	0.00
8	Nitrobenzene	0.365	0.346	5.2	103	0.00
9	Naphthalene	1.080	1.063	1.6	105	0.00
10	Hexachlorobutadiene	0.188	0.184	2.1	105	0.00
11	2-Methylnaphthalene	0.676	0.662	2.1	105	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	0.207	0.207	0.0	112	0.00
14 S	2-Fluorobiphenyl	1.541	1.457	5.5	105	0.00
15	Acenaphthylene	2.158	2.050	5.0	105	0.00
16	Acenaphthene	1.359	1.294	4.8	106	0.00
17	Fluorene	0.875	0.845	3.4	111	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
19	4-Bromophenyl-phenylether	0.340	0.420	-23.5	115	0.00
20	Hexachlorobenzene	0.159	0.219	-37.7#	127	0.00
21	Pentachlorophenol	0.068	0.072	-5.9	105	0.00
22	Phenanthrene	1.744	1.717	1.5	87	0.00
23	Anthracene	0.735	0.879	-19.6	116	0.00
24	Fluoranthene	1.357	1.667	-22.8	113	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
26	Pyrene	1.608	1.571	2.3	113	0.00
27 S	Terphenyl-d14	0.646	0.638	1.2	114	0.00
28	Benzo(a)anthracene	1.408	1.331	5.5	114	0.00
29	Chrysene	1.345	1.291	4.0	114	0.00
30	Bis(2-ethylhexyl)phthalate	0.829	0.804	3.0	121	0.00
31	Indeno(1,2,3-cd)pyrene	1.444	1.123	22.2	96	0.00
32 I	Perylene-d12	1.000	1.000	0.0	104	0.00
33	Benzo(b)fluoranthene	1.494	1.389	7.0	106	0.00
34	Benzo(k)fluoranthene	1.360	1.362	-0.1	103	0.00
35 C	Benzo(a)pyrene	1.305	1.210	7.3	102	0.00
36	Dibenzo(a,h)anthracene	1.280	1.083	15.4	96	0.00
37	Benzo(g,h,i)perylene	1.345	1.139	15.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0