

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
 Data File : BM001572.D
 Acq On : 05 Jun 2015 19:01
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:29 PM

Quant Time: Jun 06 06:14:51 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 06 01:16:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	26605	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	94440	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	45170	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	78769	5.00	ng	0.00
25) Chrysene-d12	21.28	240	70388	5.00	ng	0.00
32) Perylene-d12	23.52	264	62672	5.00	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	4596	0.77	ng	0.00
5) Phenol-d6	6.86	99	5603	0.76	ng	0.00
7) Nitrobenzene-d5	8.86	82	4966	0.75	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	1427	0.76	ng	0.00
14) 2-Fluorobiphenyl	12.97	172	11028	0.82	ng	0.00
27) Terphenyl-d14	19.72	244	6995	0.80	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	2086	0.75	ng	99
3) n-Nitrosodimethylamine	3.52	42	3216	0.78	ng	99
8) Nitrobenzene	8.90	77	5304	0.76	ng	100
9) Naphthalene	10.52	128	16091	0.81	ng	99
10) Hexachlorobutadiene	10.82	225	2838	0.83	ng	100
11) 2-Methylnaphthalene	12.15	142	10018	0.80	ng	99
15) Acenaphthylene	14.06	152	15172	0.78	ng	100
16) Acenaphthene	14.41	154	9646	0.81	ng	100
17) Fluorene	15.39	166	6359	0.68	ng	# 100
19) 4-Bromophenyl-phenylether	16.27	248	4148	1.25	ng	98
20) Hexachlorobenzene	16.38	284	1916	1.11	ng	# 73
21) Pentachlorophenol	16.75	266	744	0.81	ng	98
22) Phenanthrene	17.12	178	21803	0.93	ng	100
23) Anthracene	17.23	178	9451m	0.90	ng	
24) Fluoranthene	19.15	202	16521	0.99	ng	100
26) Pyrene	19.52	202	17295	0.80	ng	100
28) Benzo(a)anthracene	21.27	228	14879	0.76	ng	# 86
29) Chrysene	21.31	228	14820	0.84	ng	97
30) Bis(2-ethylhexyl)phthalate	21.19	149	7903	0.57	ng	100
31) Indeno(1,2,3-cd)pyrene	25.77	276	15337	0.79	ng	99
33) Benzo(b)fluoranthene	22.84	252	14544	0.84	ng	97
34) Benzo(k)fluoranthene	22.89	252	12916	0.79	ng	98
35) Benzo(a)pyrene	23.42	252	12472	0.79	ng	98
36) Dibenzo(a,h)anthracene	25.79	278	12139	0.83	ng	99
37) Benzo(g,h,i)perylene	26.46	276	12758	0.83	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
Data File : BM001572.D
Acq On : 05 Jun 2015 19:01
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM

Quant Time: Jun 06 06:14:51 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 06 01:16:45 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.70 min Scan# 325
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tgt Ion	Ratio	Lower	Upper
152	100		
150	157.8	126.6	190.0
115	64.4	51.6	77.4

Manual Integrations
APPROVED




apatel
6/8/2015 2:47:29 PM




#2

1,4-Dioxane
Concen: 0.75 ng
RT: 3.21 min Scan# 35
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
88	100		
58	82.2	66.1	99.1
43	39.4	32.3	48.5





#3





n-Nitrosodimethylamine
 Concen: 0.78 ng
 RT: 3.52 min Scan# 55
 Delta R.T. 0.02 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	42	Resp:	3216
Ion	Ratio	Lower	Upper
42	100		
74	97.0	77.2	115.8
44	8.2	7.4	11.2

Manual Integrations
 APPROVED




apatel
 6/8/2015 2:47:29 PM




#4

2-Fluorophenol
 Concen: 0.77 ng
 RT: 5.28 min Scan# 169
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Tgt Ion:	112	Resp:	4596
Ion	Ratio	Lower	Upper
112	100		
64	66.6	53.3	79.9
63	36.9	29.7	44.5





#5

Phenol-d6

Concen: 0.76 ng

RT: 6.86 min Scan# 271

Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Instrument :

BNA_M





ClientSampleId :

SSTDICC0.8

Tgt Ion:	99	Resp:	5603
Ion	Ratio	Lower	Upper
99	100		
42	27.2	21.4	32.2
71	35.8	29.0	43.4

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.47 min Scan# 552


Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Tgt Ion:	136	Resp:	94440
Ion	Ratio	Lower	Upper
136	100		
137	10.0	8.1	12.1
54	14.5	11.7	17.5
68	8.1	6.6	9.8





#7

Nitrobenzene-d5

Concen: 0.75 ng

RT: 8.86 min Scan# 422

Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Instrument :

BNA_M


ClientSampleId :

SSTDIC0.8

Tgt Ion:	82	Resp:	4966
Ion	Ratio	Lower	Upper
82	100		
128	42.0	34.1	51.1
54	48.6	37.8	56.6

Manual Integrations
APPROVED




apatel
6/8/2015 2:47:29 PM



Abundance Ion 82.00 (81.70 to 82.70): BM001571.D (-417) (-)

Ion 128.00 (127.70 to 128.70): BM001571.D (-417) (-)

Ion 54.00 (53.70 to 54.70): BM001571.D (-417) (-)



#8

Nitrobenzene

Concen: 0.76 ng


RT: 8.90 min Scan# 426

Delta R.T. 0.01 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01



Tgt Ion:	77	Resp:	5304
Ion	Ratio	Lower	Upper
77	100		
123	39.5	31.7	47.5
65	14.0	11.4	17.0




Abundance Ion 77.00 (76.70 to 77.70): BM001571.D (-415) (-)

Ion 123.00 (122.70 to 123.70): BM001571.D (-415) (-)

Ion 65.00 (64.70 to 65.70): BM001571.D (-415) (-)





#9

Naphthalene

Concen: 0.81 ng

RT: 10.52 min Scan# 555

Delta R.T. -0.00 min

Lab File: BM001572.D


Acq: 05 Jun 2015 19:01

Instrument :

BNA_M

ClientSampleId :




SSTDIC0.8



Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.7	8.7	13.1
127	13.8	10.7	16.1

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM



#10

Hexachlorobutadiene

Concen: 0.83 ng




RT: 10.82 min Scan# 572


Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.9	50.4	75.6
227	63.6	50.8	76.2









#11
2-Methylnaphthalene
Concen: 0.80 ng
RT: 12.15 min Scan# 681
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
142	100		
141	93.3	75.6	113.4
115	44.6	35.5	53.3




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:29 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.35 min Scan# 880
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.8	83.8	125.6
160	46.7	37.1	55.7









#13
 2,4,6-Tribromophenol
 Concen: 0.76 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	1427		
332	89.4	72.0	108.0	
141	0.0	0.0	0.0	




Manual Integrations
 APPROVED


apatel
 6/8/2015 2:47:29 PM



#14
 2-Fluorobiphenyl
 Concen: 0.82 ng
 RT: 12.97 min Scan# 760
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	11028		
171	32.3	26.2	39.2	
170	20.1	16.1	24.1	





#15

Acenaphthylene

Concen: 0.78 ng

RT: 14.06 min Scan# 856

Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.2	22.8
153	12.9	10.3	15.5

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM



#16

Acenaphthene

Concen: 0.81 ng




RT: 14.41 min Scan# 885


Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
154	100		
153	110.7	89.1	133.7
152	53.9	43.2	64.8





#17

Fluorene

Concen: 0.68 ng

RT: 15.39 min Scan# 930

Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Instrument :

BNA_M





ClientSampleId :

SSTDIC0.8

Tgt Ion:	166	Resp:	6359
Ion Ratio	Lower	Upper	
166	100		
165	0.0	0.0	0.0
167	14.1	11.3	16.9

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.09 min Scan# 980


Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Tgt Ion:	188	Resp:	78769
Ion Ratio	Lower	Upper	
188	100		
94	0.9	0.7	1.1
80	0.8	0.6	1.0









#19
4-Bromophenyl-phenylether
Concen: 1.25 ng
RT: 16.27 min Scan# 956
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8

Tgt Ion	Ratio	Lower	Upper
248	100		
250	76.3	63.1	94.7
141	44.3	35.0	52.6




Manual Integrations
APPROVED




apatel
6/8/2015 2:47:29 PM



#20
Hexachlorobenzene
Concen: 1.11 ng
RT: 16.38 min Scan# 959
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
284	100		
142	94.3	64.2	96.2
249	101.3	55.4	83.0









#21
Pentachlorophenol
Concen: 0.81 ng
RT: 16.75 min Scan# 970
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion:	266	Resp:	744
Ion Ratio	Lower	Upper	
266	100		
268	84.1	66.3	99.5
264	59.4	45.5	68.3

Instrument :
BNA_M
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED


apatel
6/8/2015 2:47:29 PM



#22
Phenanthrene
Concen: 0.93 ng
RT: 17.12 min Scan# 981
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion:	178	Resp:	21803
Ion Ratio	Lower	Upper	
178	100		
176	17.0	13.5	20.3
179	16.6	13.3	19.9









#23
 Anthracene
 Concen: 0.90 ng m
 RT: 17.23 min Scan# 984
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	178	Resp:	9451
Ion Ratio	Lower	Upper	
178	100		
176	54.9	45.9	68.9
179	39.2	33.3	49.9




Manual Integrations
 APPROVED


apatel
 6/8/2015 2:47:29 PM



#24
 Fluoranthene
 Concen: 0.99 ng
 RT: 19.15 min Scan# 1089
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Tgt Ion:	202	Resp:	16521
Ion Ratio	Lower	Upper	
202	100		
101	12.6	10.0	15.0
203	17.1	13.9	20.9









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.28 min Scan# 1252
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
240	100		
120	4.5	2.2	3.2#
236	26.8	19.5	29.3




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:29 PM



#26
Pyrene
Concen: 0.80 ng
RT: 19.52 min Scan# 1119
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.6	24.8
203	17.5	13.9	20.9






#27
 Terphenyl-d14
 Concen: 0.80 ng
 RT: 19.72 min Scan# 1136
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8




Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.6	9.6	14.4
122	28.8	19.0	28.4

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:29 PM






Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E




#28
 Benzo(a)anthracene
 Concen: 0.76 ng
 RT: 21.27 min Scan# 1251
 Delta R.T. 0.02 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.1	24.8	37.2
229	22.1	13.4	20.0

Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E





#29

Chrysene

Concen: 0.84 ng

RT: 21.31 min Scan# 1254

Delta R.T. -0.00 min

Lab File: BM001572.D


Acq: 05 Jun 2015 19:01

Instrument :

BNA_M

ClientSampleId :




SSTDIC0.8



Tgt Ion:	228	Resp:	14820
Ion Ratio		Lower	Upper
228	100		
226	28.0	21.0	31.6
229	20.3	17.2	25.8

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:29 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.57 ng




RT: 21.19 min Scan# 1246


Delta R.T. -0.00 min

Lab File: BM001572.D

Acq: 05 Jun 2015 19:01

Tgt Ion:	149	Resp:	7903
Ion Ratio		Lower	Upper
149	100		
167	25.3	20.2	30.2
279	2.1	1.8	2.6









#31
 Indeno(1,2,3-cd)pyrene
 Concen: 0.79 ng
 RT: 25.77 min Scan# 1660
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tgt Ion	Ratio	Resp	Lower	Upper
276	100	15337		
138	34.6	28.0	42.0	
277	24.7	19.7	29.5	




Manual Integrations
 APPROVED


apatel
 6/8/2015 2:47:29 PM



#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.52 min Scan# 1444
 Delta R.T. 0.01 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	62672		
260	25.0	21.6	32.4	
265	21.8	16.5	24.7	









#33
Benzo(b)fluoranthene
Concen: 0.84 ng
RT: 22.84 min Scan# 1379
Delta R.T. 0.01 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.4	17.4	26.2
125	14.0	11.8	17.8




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:29 PM



#34
Benzo(k)fluoranthene
Concen: 0.79 ng
RT: 22.89 min Scan# 1383
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.8	28.2
125	14.3	10.7	16.1









#35
Benzo(a)pyrene
Concen: 0.79 ng
RT: 23.42 min Scan# 1434
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.8	28.2
125	16.6	12.1	18.1




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:29 PM



#36
Dibenzo(a,h)anthracene
Concen: 0.83 ng
RT: 25.79 min Scan# 1662
Delta R.T. -0.00 min
Lab File: BM001572.D
Acq: 05 Jun 2015 19:01

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.6	17.4	26.0
279	24.4	20.0	30.0






#37
 Benzo(g,h,i)perylene
 Concen: 0.83 ng
 RT: 26.46 min Scan# 1726
 Delta R.T. -0.00 min
 Lab File: BM001572.D
 Acq: 05 Jun 2015 19:01

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.2	28.8
138	31.1	23.7	35.5

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:29 PM



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

