

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
 Data File : BM001578.D
 Acq On : 06 Jun 2015 05:06
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:54 PM

Quant Time: Jun 08 05:41:17 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 05:27:25 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	27107	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	96358	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	44829	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	61652	5.00	ng	0.00
25) Chrysene-d12	21.28	240	71894	5.00	ng	0.00
32) Perylene-d12	23.51	264	63077	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.27	112	5803	0.97	ng	0.00
5) Phenol-d6	6.84	99	7123	0.97	ng	0.00
7) Nitrobenzene-d5	8.85	82	6248	0.95	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	1901	0.94	ng	0.00
14) 2-Fluorobiphenyl	12.96	172	13456	0.97	ng	0.00
27) Terphenyl-d14	19.72	244	8505	0.92	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	2582	0.93	ng	100
3) n-Nitrosodimethylamine	3.51	42	4080	0.99	ng	# 98
8) Nitrobenzene	8.89	77	6669	0.95	ng	100
9) Naphthalene	10.52	128	20094	0.97	ng	99
10) Hexachlorobutadiene	10.82	225	3504	0.97	ng	99
11) 2-Methylnaphthalene	12.15	142	12407	0.95	ng	96
15) Acenaphthylene	14.06	152	18625	0.96	ng	100
16) Acenaphthene	14.41	154	11415	0.94	ng	98
17) Fluorene	15.36	166	7704	0.98	ng	# 65
19) 4-Bromophenyl-phenylether	16.27	248	5468	1.18	ng	# 89
20) Hexachlorobenzene	16.38	284	2601	1.33	ng	96
21) Pentachlorophenol	16.72	266	1133	1.20	ng	# 64
22) Phenanthrene	17.12	178	21959	1.02	ng	99
23) Anthracene	17.19	178	11361m	1.25	ng	
24) Fluoranthene	19.15	202	20246	1.21	ng	99
26) Pyrene	19.52	202	21621	0.93	ng	100
28) Benzo(a)anthracene	21.25	228	18544	0.92	ng	99
29) Chrysene	21.31	228	18073	0.93	ng	100
30) Bis(2-ethylhexyl)phthalate	21.19	149	10547	0.87	ng	100
31) Indeno(1,2,3-cd)pyrene	25.77	276	18336	0.88	ng	100
33) Benzo(b)fluoranthene	22.83	252	16948	0.90	ng	99
34) Benzo(k)fluoranthene	22.89	252	17024	0.99	ng	100
35) Benzo(a)pyrene	23.42	252	15404	0.94	ng	99
36) Dibenzo(a,h)anthracene	25.79	278	14369	0.89	ng	99
37) Benzo(g,h,i)perylene	26.46	276	15217	0.90	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060615\
Data File : BM001578.D
Acq On : 06 Jun 2015 05:06
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
ICVBM060615

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM

Quant Time: Jun 08 05:41:17 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 05:27:25 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.70 min Scan# 325

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Instrument :

BNA_M





ClientSampleId :

ICVBM060615

Tgt Ion:	152	Resp:	27107
Ion Ratio	Lower	Upper	
152	100		
150	153.3	126.6	190.0
115	61.0	51.6	77.4

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#2

1,4-Dioxane

Concen: 0.93 ng




RT: 3.21 min Scan# 35


Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion:	88	Resp:	2582
Ion Ratio	Lower	Upper	
88	100		
58	83.0	66.1	99.1
43	40.2	32.3	48.5





#3

n-Nitrosodimethylamine

Concen: 0.99 ng

RT: 3.51 min Scan# 54

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Instrument :

BNA_M


ClientSampleId :

ICVBM060615

Tgt Ion:	42	Resp:	4080
Ion	Ratio	Lower	Upper
42	100		
74	94.8	77.2	115.8
44	6.0	7.4	11.2#

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM






Abundance Ion 42.00 (41.70 to 42.70): BM001571.D (-51) (-)

Ion 74.00 (73.70 to 74.70): BM001571.D (-51) (-)

Ion 44.00 (43.70 to 44.70): BM001571.D (-51) (-)

Time-->



#4

2-Fluorophenol

Concen: 0.97 ng


RT: 5.27 min Scan# 168

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion:	112	Resp:	5803
Ion	Ratio	Lower	Upper
112	100		
64	67.8	53.3	79.9
63	37.9	29.7	44.5






Abundance Ion 112.00 (111.70 to 112.70): BM001571.D (-165) (-)

Ion 64.00 (63.70 to 64.70): BM001571.D (-165) (-)

Ion 63.00 (62.70 to 63.70): BM001571.D (-165) (-)

Time-->





#5

Phenol-d6

Concen: 0.97 ng

RT: 6.84 min Scan# 270

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Instrument :

BNA_M





ClientSampleId :

ICVBM060615

Tgt Ion:	99	Resp:	7123
Ion Ratio	Lower	Upper	
99	100		
42	27.2	21.4	32.2
71	36.5	29.0	43.4

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.47 min Scan# 552


Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion:	136	Resp:	96358
Ion Ratio	Lower	Upper	
136	100		
137	10.1	8.1	12.1
54	12.2	11.7	17.5
68	7.1	6.6	9.8





#7

Nitrobenzene-d5

Concen: 0.95 ng

RT: 8.85 min Scan# 421

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Instrument :

BNA_M





ClientSampled :

ICVBM060615

Tgt Ion:	82	Resp:	6248
Ion Ratio	Lower	Upper	
82	100		
128	33.1	34.1	51.1#
54	66.7	37.8	56.6#

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#8

Nitrobenzene

Concen: 0.95 ng




RT: 8.89 min Scan# 425


Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion:	77	Resp:	6669
Ion Ratio	Lower	Upper	
77	100		
123	39.4	31.7	47.5
65	14.2	11.4	17.0





#9

Naphthalene

Concen: 0.97 ng

RT: 10.52 min Scan# 555

Delta R.T. -0.00 min

Lab File: BM001578.D


Acq: 06 Jun 2015 05:06

Instrument :

BNA_M

ClientSampleId :




ICVBM060615



Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.7	13.1
127	13.2	10.7	16.1

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#10

Hexachlorobutadiene


Concen: 0.97 ng

RT: 10.82 min Scan# 572



Delta R.T. -0.00 min


Lab File: BM001578.D

Acq: 06 Jun 2015 05:06



Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.8	50.4	75.6
227	63.6	50.8	76.2









#11
2-Methylnaphthalene
Concen: 0.95 ng
RT: 12.15 min Scan# 681
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Instrument :
BNA_M
ClientSampleId :
ICVBM060615

Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.6	75.6	113.4
115	41.8	35.5	53.3




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:54 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.35 min Scan# 880
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.3	83.8	125.6
160	43.6	37.1	55.7









#13
 2,4,6-Tribromophenol
 Concen: 0.94 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	1901		
332	93.0	72.0	108.0	
141	0.0	0.0	0.0	




Manual Integrations
 APPROVED


apatel
 6/8/2015 2:47:54 PM



#14
 2-Fluorobiphenyl
 Concen: 0.97 ng
 RT: 12.96 min Scan# 759
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	13456		
171	37.4	26.2	39.2	
170	26.7	16.1	24.1	









#15
Acenaphthylene
Concen: 0.96 ng
RT: 14.06 min Scan# 856
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Instrument :
BNA_M
ClientSampleId :
ICVBM060615

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.0	15.2	22.8
153	13.0	10.3	15.5




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:54 PM



#16
Acenaphthene
Concen: 0.94 ng
RT: 14.41 min Scan# 885
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.7	89.1	133.7
152	54.9	43.2	64.8






#17
Fluorene
 Concen: 0.98 ng
 RT: 15.36 min Scan# 929
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615




Tgt Ion	Ratio	Lower	Upper
166	100		
165	105.3	0.0	0.0#
167	0.0	11.3	16.9#

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:54 PM






Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E




#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.09 min Scan# 980
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.8	0.7	1.1
80	0.7	0.6	1.0

Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BM0
 Ion 80.00 (79.70 to 80.70): BM0






#19
4-Bromophenyl-phenylether
Concen: 1.18 ng
RT: 16.27 min Scan# 956
Delta R.T. 0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Instrument :
BNA_M
ClientSampleId :
ICVBM060615




Tgt Ion	Ratio	Lower	Upper
248	100		
250	84.7	63.1	94.7
141	31.8	35.0	52.6

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM






Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E




#20
Hexachlorobenzene
Concen: 1.33 ng
RT: 16.38 min Scan# 959
Delta R.T. 0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
284	100		
142	73.2	64.2	96.2
249	69.5	55.4	83.0

Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E





#21

Pentachlorophenol

Concen: 1.20 ng

RT: 16.72 min Scan# 969

Delta R.T. -0.00 min

Lab File: BM001578.D


Acq: 06 Jun 2015 05:06

Instrument :

BNA_M

ClientSampleId :




ICVBM060615



Tgt Ion	Ratio	Lower	Upper
266	100		
268	53.2	66.3	99.5#
264	86.7	45.5	68.3#

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#22

Phenanthrene

Concen: 1.02 ng




RT: 17.12 min Scan# 981


Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
178	100		
176	16.7	13.5	20.3
179	16.9	13.3	19.9









#23
 Anthracene
 Concen: 1.25 ng m
 RT: 17.19 min Scan# 983
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615

Tgt Ion	Ratio	Lower	Upper
178	100		
176	52.0	45.9	68.9
179	46.2	33.3	49.9




Manual Integrations
 APPROVED


apatel
 6/8/2015 2:47:54 PM



#24
 Fluoranthene
 Concen: 1.21 ng
 RT: 19.15 min Scan# 1089
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.3	10.0	15.0
203	17.1	13.9	20.9









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.28 min Scan# 1252
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Instrument :
BNA_M
ClientSampleId :
ICVBM060615

Tgt Ion	Ratio	Lower	Upper
240	100		
120	2.4	2.2	3.2
236	24.2	19.5	29.3




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:54 PM



#26
Pyrene
Concen: 0.93 ng
RT: 19.52 min Scan# 1119
Delta R.T. -0.00 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.6	24.8
203	17.3	13.9	20.9





#27

Terphenyl-d14

Concen: 0.92 ng

RT: 19.72 min Scan# 1136

Delta R.T. -0.00 min

Lab File: BM001578.D


Acq: 06 Jun 2015 05:06

Instrument :

BNA_M

ClientSampleId :

ICVBM060615



Tgt Ion:244 Resp: 8505

Ion Ratio Lower Upper




244 100

212 11.6 9.6 14.4

122 18.5 19.0 28.4#

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#28

Benzo(a)anthracene

Concen: 0.92 ng

RT: 21.25 min Scan# 1250

Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06




Tgt Ion:228 Resp: 18544


Ion Ratio Lower Upper

228 100

226 30.4 24.8 37.2

229 17.2 13.4 20.0





#29

Chrysene

Concen: 0.93 ng

RT: 21.31 min Scan# 1254

Delta R.T. -0.00 min

Lab File: BM001578.D


Acq: 06 Jun 2015 05:06

Instrument :

BNA_M

ClientSampleId :




ICVBM060615



Tgt Ion:	228	Resp:	18073
Ion Ratio		Lower	Upper
228	100		
226	26.3	21.0	31.6
229	21.5	17.2	25.8

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.87 ng




RT: 21.19 min Scan# 1246


Delta R.T. -0.00 min

Lab File: BM001578.D

Acq: 06 Jun 2015 05:06

Tgt Ion:	149	Resp:	10547
Ion Ratio		Lower	Upper
149	100		
167	25.3	20.2	30.2
279	2.1	1.8	2.6






#31
 Indeno(1,2,3-cd)pyrene
 Concen: 0.88 ng
 RT: 25.77 min Scan# 1660
 Delta R.T. 0.01 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615




Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	34.8	28.0	42.0
277	24.5	19.7	29.5

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:54 PM




Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E






#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.51 min Scan# 1443
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Tgt Ion	Ratio	Resp Lower	Resp Upper
264	100		
260	27.4	21.6	32.4
265	20.7	16.5	24.7



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E





#33

Benzo(b)fluoranthene

Concen: 0.90 ng

RT: 22.83 min Scan# 1378

Delta R.T. -0.00 min

Lab File: BM001578.D


Acq: 06 Jun 2015 05:06

Instrument :

BNA_M

ClientSampleId :



ICVBM060615



Tgt Ion:	252	Resp:	16948
Ion Ratio	Lower	Upper	
252	100		
253	21.4	17.4	26.2
125	15.1	11.8	17.8

Manual Integrations
APPROVED

apatel
6/8/2015 2:47:54 PM



#34

Benzo(k)fluoranthene


Concen: 0.99 ng

RT: 22.89 min Scan# 1383



Delta R.T. 0.01 min


Lab File: BM001578.D

Acq: 06 Jun 2015 05:06




Tgt Ion:	252	Resp:	17024
Ion Ratio	Lower	Upper	
252	100		
253	23.6	18.8	28.2
125	13.2	10.7	16.1





#35
Benzo(a)pyrene
Concen: 0.94 ng
RT: 23.42 min Scan# 1434
Delta R.T. 0.01 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06




Instrument :
BNA_M
ClientSampleId :
ICVBM060615



Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.9	18.8	28.2
125	15.2	12.1	18.1




Manual Integrations
APPROVED


apatel
6/8/2015 2:47:54 PM



#36
Dibenzo(a,h)anthracene
Concen: 0.89 ng
RT: 25.79 min Scan# 1662
Delta R.T. 0.01 min
Lab File: BM001578.D
Acq: 06 Jun 2015 05:06


Tgt Ion	Ratio	Lower	Upper
278	100		
139	21.4	17.4	26.0
279	24.7	20.0	30.0





#37
 Benzo(g,h,i)perylene
 Concen: 0.90 ng
 RT: 26.46 min Scan# 1726
 Delta R.T. -0.00 min
 Lab File: BM001578.D
 Acq: 06 Jun 2015 05:06

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060615



Tgt	Ion	276	Resp:	15217
	Ion	Ratio	Lower	Upper
	276	100		
	277	24.1	19.2	28.8
	138	29.3	23.7	35.5

Manual Integrations
 APPROVED

apatel
 6/8/2015 2:47:54 PM

