

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060617\
 Data File : BM010422.D
 Acq On : 06 Jun 2017 09:52
 Operator : SJ/MA
 Sample : I3445-05
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SVOC-GPC-BLANK

Quant Time: Jun 07 01:56:46 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jun 07 01:44:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	238718	20.00	ng/ul	0.00
18) Naphthalene-d8	10.72	136	852431	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.55	164	522056	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.30	188	1155533	20.00	ng/ul	0.00
75) Chrysene-d12	21.46	240	1480869	20.00	ng/ul	0.00
83) Perylene-d12	23.81	264	1662823	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0d	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
46) Acenaphthylene-d8	0.00	160	0d	0.00	ng/ul
51) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
57) Fluorene-d10	0.00	176	0d	0.00	ng/ul
62) 4,6-Dinitro-2-methylphenol	0.00	200	0d	0.00	ng/ul
70) Anthracene-d10	0.00	188	0d	0.00	ng/ul
76) Pyrene-d10	0.00	212	0d	0.00	ng/ul
87) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060617\
Data File : BM010422.D
Acq On : 06 Jun 2017 09:52
Operator : SJ/MA
Sample : I3445-05
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Jun 07 01:56:46 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM052717.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jun 07 01:44:59 2017
Response via : Initial Calibration

