

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060624\
 Data File : BM045870.D
 Acq On : 06 Jun 2024 17:13
 Operator : MA/JU
 Sample : SSTD3.250
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2013

Quant Time: Jun 06 17:56:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM060624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 17:32:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.480	152	4875	0.400	ng/ul	0.00
4) Naphthalene-d8	10.244	136	16780	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.114	164	9884	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.858	188	19906	0.400	ng/ul	0.00
17) Chrysene-d12	21.052	240	15113	0.400	ng/ul	0.00
23) Perylene-d12	23.165	264	17391	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.067	96	18677	3.407	ng/ul	-0.02
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.101	88	19836	3.103	ng/ul#	55
14) Pentachlorophenol	16.538	266	13830	3.872	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060624\
Data File : BM045870.D
Acq On : 06 Jun 2024 17:13
Operator : MA/JU
Sample : SST3.250
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2013

Quant Time: Jun 06 17:56:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM060624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 06 17:32:58 2024
Response via : Initial Calibration

