


Data Path : Z:\HPCHEM1\BNA M\DATA\BM060716\
 Data File : BM005689.D
 Acq On : 07 Jun 2016 13:04
 Operator : UM/SJ
 Sample : DFTPP50
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP50

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM060216.M
 Title : SVOA CALIBRATION
 Last Update : Wed Jun 08 01:24:32 2016



AutoFind: Scans 2548, 2549, 2550; Background Corrected with Scan 2542

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	39594	PASS
68	69	0.00	2	1.2	448	PASS
69	198	0.00	100	34.9	38311	PASS
70	69	0.00	2	0.1	22	PASS
127	198	10	80	48.2	52898	PASS
197	198	0.00	2	0.5	524	PASS
198	198	100	100	100.0	109781	PASS
199	198	5	9	6.6	7284	PASS
275	198	10	60	26.1	28666	PASS
365	198	1	100	3.3	3606	PASS
441	443	0.01	100	79.0	14349	PASS
442	198	50	100	85.9	94256	PASS
443	442	15	24	19.3	18168	PASS

m/z	Abundance
39.00	621.0
39.90	708.0
41.00	1338.0
42.00	667.0
43.00	2309.0
44.00	1527.0
45.00	968.0
55.00	1362.0
56.00	557.0
57.00	1013.0
60.00	764.0
67.00	401.0
68.00	348.0
69.00	801.0
70.00	419.0
71.00	546.0
73.00	775.0
77.00	363.0
81.00	385.0
82.10	377.0
83.00	766.0
83.90	581.0
85.00	545.0
90.90	346.0
95.10	340.0
97.00	564.0
104.90	359.0
166.90	343.0
207.00	579.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.00	589.0
39.00	3826.0
39.90	692.0
41.00	1023.0
42.00	531.0
43.00	2013.0
44.00	1473.0
45.00	865.0
49.00	581.0
50.00	11647.0
51.00	47008.0
52.00	2742.0
53.00	314.0
55.00	1236.0
56.00	1578.0
57.00	3857.0
60.00	867.0
61.00	524.0
62.00	584.0
63.00	1827.0
65.00	1170.0
67.00	349.0
68.00	858.0
69.00	45872.0
70.00	544.0
71.00	430.0
73.00	1017.0
74.00	4955.0
75.00	7289.0
76.00	2985.0
77.00	58440.0
78.00	4268.0
79.00	3440.0
80.00	2620.0
81.00	3980.0
82.00	1187.0
83.00	1283.0
84.00	489.0
85.00	1132.0
85.90	1143.0
86.90	568.0
91.00	1174.0
91.90	868.0
93.00	5890.0
93.90	554.0
95.00	468.0
96.00	428.0
96.90	392.0
98.00	4659.0
99.00	3826.0
100.00	385.0
100.90	2279.0
103.00	674.0
103.90	1208.0
105.00	1425.0
105.90	464.0
107.00	17896.0
108.00	3024.0
109.00	631.0
110.00	30368.0
111.00	4860.0
112.00	696.0
115.90	838.0
117.00	13566.0
118.00	1104.0
122.00	1078.0
122.90	1681.0
124.00	743.0
125.00	788.0
126.10	362.0
127.00	59736.0
128.00	4389.0
129.00	22232.0
130.00	1865.0
130.90	510.0
134.00	624.0
135.00	2079.0
136.00	876.0
136.90	1035.0
140.00	319.0
140.90	2558.0
141.90	986.0
143.00	702.0
145.90	529.0
147.00	1571.0
148.00	3535.0
149.00	984.0
150.90	404.0
152.90	916.0
153.90	724.0
155.00	1569.0
156.00	2584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

156.90	508.0
157.90	503.0
159.00	409.0
159.90	1075.0
161.00	1362.0
161.90	472.0
164.90	1156.0
166.00	1070.0
167.00	6342.0
168.00	3466.0
168.90	606.0
170.90	343.0
171.90	433.0
172.90	707.0
174.00	1084.0
175.00	2572.0
175.90	759.0
176.90	1027.0
177.90	395.0
178.90	4336.0
180.00	2948.0
181.00	1633.0
182.00	368.0
183.90	334.0
185.00	2305.0
186.00	16728.0
187.00	4962.0
188.00	512.0
189.00	998.0
191.00	689.0
191.90	1471.0
193.00	1632.0
194.00	388.0
196.00	3186.0
197.00	568.0
197.90	117816.0
198.90	8143.0
200.00	716.0
201.30	398.0
203.00	875.0
204.00	4386.0
205.00	7333.0
206.00	30992.0
207.00	4640.0
208.00	1091.0
209.00	395.0
210.00	413.0
211.00	1336.0
214.90	431.0
215.90	779.0
217.00	7933.0
218.00	1095.0
221.00	4605.0
222.00	954.0
223.00	1957.0
224.00	18024.0
225.00	4596.0
226.00	451.0
227.00	6936.0
228.00	1080.0
229.00	1600.0
230.90	665.0
233.90	431.0
235.00	525.0
235.90	379.0
236.90	568.0
238.90	325.0
241.00	443.0
241.90	944.0
243.00	871.0
244.00	13164.0
245.00	1789.0
246.00	2676.0
246.90	523.0
249.00	479.0
254.00	542.0
255.00	68184.0
256.00	10029.0
257.00	776.0
258.00	3701.0
258.90	611.0
264.90	1671.0
273.00	2000.0
274.00	5390.0
275.00	28528.0
276.00	3728.0
277.00	2538.0
277.90	437.0
285.00	499.0
292.90	763.0
296.00	8357.0
297.00	1176.0
303.00	931.0
314.00	337.0
314.90	1062.0
316.00	628.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

323.00	2610.0
324.00	507.0
326.90	559.0
334.00	1756.0
335.00	436.0
341.00	319.0
345.90	520.0
352.00	797.0
352.90	591.0
354.00	886.0
365.00	3474.0
365.90	613.0
372.00	1312.0
373.00	344.0
383.00	359.0
402.00	571.0
403.00	710.0
421.00	685.0
422.00	672.0
423.00	4284.0
424.00	785.0
441.00	12677.0
442.00	82472.0
443.00	16664.0
444.00	1482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.00	573.0
39.00	3578.0
39.90	753.0
41.00	1013.0
42.00	484.0
43.00	2041.0
44.00	1597.0
45.00	859.0
49.00	497.0
50.00	10896.0
51.00	46120.0
52.00	2686.0
55.00	1191.0
56.00	1616.0
57.00	4045.0
60.00	637.0
61.00	666.0
61.90	623.0
63.00	1970.0
64.00	309.0
65.00	1053.0
67.00	441.0
68.00	997.0
69.00	45752.0
70.00	465.0
71.00	358.0
73.00	1033.0
74.00	4816.0
75.00	7271.0
76.00	2691.0
77.00	59360.0
78.00	4289.0
79.00	3113.0
80.00	2690.0
81.00	4070.0
82.00	1295.0
83.00	1285.0
84.00	558.0
85.00	1008.0
86.00	1096.0
86.90	632.0
91.00	1055.0
92.00	964.0
93.00	6052.0
93.90	525.0
95.00	448.0
96.00	525.0
97.00	436.0
98.00	4496.0
99.00	3902.0
99.90	374.0
101.00	2498.0
102.90	738.0
104.00	1508.0
105.00	1498.0
106.00	494.0
107.00	17768.0
108.00	2829.0
109.00	575.0
110.00	30048.0
111.00	4951.0
112.00	620.0
115.90	880.0
117.00	13762.0
117.90	1064.0
122.00	1189.0
123.00	1845.0
123.90	787.0
124.90	629.0
126.10	342.0
127.00	61600.0
128.00	4672.0
129.00	22680.0
129.90	2067.0
130.90	435.0
133.90	712.0
134.90	2129.0
135.90	856.0
137.00	1033.0
140.90	2909.0
141.90	948.0
142.90	805.0
146.00	588.0
147.00	1704.0
147.90	3707.0
149.00	1022.0
150.00	320.0
151.10	407.0
152.90	1031.0
153.90	781.0
155.00	1676.0
156.00	2598.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

157.00	526.0
157.90	592.0
158.90	506.0
160.00	864.0
160.90	1483.0
162.00	477.0
164.90	1023.0
166.00	852.0
167.00	6543.0
168.00	3320.0
169.00	686.0
171.90	569.0
172.90	739.0
174.00	1379.0
175.00	2596.0
176.00	636.0
177.00	1144.0
177.90	478.0
178.90	4395.0
180.00	3339.0
181.00	1601.0
182.00	346.0
183.90	417.0
185.00	2342.0
186.00	17248.0
187.00	5247.0
187.90	644.0
189.00	1102.0
191.00	613.0
191.90	1516.0
193.00	1784.0
194.00	448.0
196.00	3341.0
197.00	613.0
197.90	128160.0
198.90	8493.0
199.90	620.0
201.60	531.0
202.90	931.0
204.00	4820.0
205.00	8712.0
206.00	34416.0
207.00	4900.0
208.00	1172.0
208.90	380.0
210.20	544.0
210.90	1353.0
215.00	408.0
216.00	703.0
216.90	8631.0
217.90	1010.0
221.00	4746.0
221.90	1127.0
223.00	1981.0
224.00	18808.0
225.00	4903.0
226.00	586.0
227.00	7791.0
228.00	1194.0
229.00	1573.0
231.00	638.0
233.90	437.0
235.00	665.0
235.90	390.0
236.90	581.0
238.90	338.0
240.90	474.0
242.00	1006.0
243.00	1173.0
244.00	15215.0
245.00	1954.0
246.00	3050.0
246.90	486.0
248.90	532.0
253.00	336.0
254.00	516.0
255.00	77600.0
256.00	11643.0
256.90	839.0
258.00	4569.0
259.00	633.0
265.00	1795.0
273.00	2160.0
274.00	6470.0
275.00	33848.0
276.00	4300.0
277.00	2681.0
277.90	502.0
283.00	328.0
285.00	553.0
292.90	605.0
296.00	9937.0
297.00	1354.0
303.00	1124.0
304.00	342.0
314.00	490.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

315.00	1170.0
316.00	508.0
320.90	327.0
323.00	3058.0
324.00	574.0
327.00	649.0
328.00	308.0
334.00	2084.0
335.00	457.0
341.00	401.0
345.90	834.0
352.00	894.0
353.00	739.0
354.00	953.0
365.00	4312.0
366.00	540.0
372.00	1634.0
372.90	358.0
383.00	404.0
402.00	714.0
403.00	777.0
403.90	346.0
421.00	840.0
422.00	725.0
423.00	5549.0
424.00	1338.0
441.00	16369.0
442.00	110144.0
443.00	20992.0
444.00	2019.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.00	445.0
39.00	2337.0
40.00	631.0
41.00	1004.0
42.00	496.0
43.00	2083.0
44.00	1646.0
45.00	807.0
49.00	454.0
50.00	6916.0
51.00	25656.0
52.00	1395.0
55.00	1359.0
55.90	1113.0
57.00	2350.0
60.00	772.0
60.90	473.0
62.00	306.0
63.00	1059.0
65.00	698.0
67.00	313.0
67.90	533.0
69.00	25712.0
70.00	315.0
71.00	373.0
73.00	865.0
74.00	2756.0
75.00	4548.0
76.00	1701.0
77.00	34464.0
78.00	2331.0
79.00	2098.0
79.90	1519.0
81.00	2377.0
81.90	770.0
83.00	915.0
83.90	513.0
84.90	862.0
85.90	748.0
87.00	430.0
91.00	821.0
92.00	610.0
93.00	3507.0
94.00	341.0
94.90	389.0
96.00	306.0
97.10	395.0
98.00	2697.0
99.00	2299.0
100.90	1493.0
103.00	560.0
103.90	852.0
105.00	1034.0
106.00	449.0
107.00	10485.0
108.00	1711.0
109.00	493.0
110.00	18224.0
111.00	2821.0
112.00	379.0
115.90	600.0
117.00	8024.0
117.90	676.0
122.00	717.0
122.90	953.0
123.90	595.0
125.00	553.0
127.00	37360.0
128.00	2826.0
129.00	13845.0
129.90	1105.0
130.90	374.0
133.90	503.0
134.90	1237.0
136.00	558.0
137.00	801.0
140.90	1822.0
141.90	627.0
142.90	474.0
145.90	369.0
146.90	1015.0
148.00	2166.0
148.90	742.0
151.00	321.0
152.90	581.0
153.90	486.0
155.00	1262.0
156.00	1543.0
156.90	354.0
157.90	393.0
160.00	548.0
160.90	830.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

162.00	321.0
164.90	691.0
166.00	681.0
167.00	4381.0
167.90	2204.0
168.90	363.0
172.00	399.0
172.90	428.0
173.90	787.0
175.00	1707.0
175.90	446.0
176.90	812.0
177.90	337.0
178.90	3035.0
179.90	2089.0
181.00	1169.0
184.90	1728.0
186.00	11458.0
187.00	3519.0
188.00	363.0
189.00	694.0
190.90	404.0
191.90	1036.0
193.00	1254.0
193.90	337.0
195.90	2373.0
197.00	392.0
197.90	83368.0
198.90	5217.0
199.90	446.0
201.50	402.0
202.90	614.0
204.00	3414.0
205.00	5372.0
206.00	22392.0
207.00	3176.0
208.00	808.0
209.00	345.0
210.00	315.0
210.90	862.0
214.90	300.0
216.00	405.0
216.90	5730.0
217.90	728.0
221.00	3582.0
221.90	691.0
223.00	1297.0
224.00	13249.0
225.00	3011.0
226.00	386.0
227.00	5091.0
228.00	715.0
228.90	1037.0
231.00	496.0
235.00	365.0
236.00	310.0
237.00	451.0
240.90	311.0
242.00	636.0
242.90	591.0
244.00	10405.0
245.00	1389.0
246.00	1969.0
246.90	401.0
249.00	307.0
253.90	331.0
255.00	53152.0
256.00	7366.0
257.00	575.0
258.00	2908.0
258.90	482.0
264.90	1322.0
273.00	1531.0
274.00	4108.0
275.00	23624.0
276.00	3231.0
276.90	1884.0
277.90	348.0
283.00	338.0
285.00	399.0
292.90	468.0
296.00	7170.0
296.90	919.0
303.00	748.0
313.90	418.0
314.90	881.0
316.00	405.0
323.00	2413.0
324.00	425.0
327.00	421.0
334.00	1441.0
334.90	388.0
346.00	625.0
352.00	717.0
352.90	441.0
354.00	781.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

365.00	3034.0
366.00	492.0
372.00	1268.0
383.00	322.0
402.00	575.0
403.00	624.0
404.00	304.0
421.00	661.0
421.90	603.0
423.00	4721.0
424.00	785.0
441.00	14002.0
442.00	90152.0
443.00	16848.0
444.00	1567.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50