


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060721\
 Data File : BM030304.D
 Acq On : 07 Jun 2021 08:37
 Operator : CG/JU
 Sample : DFTPP49
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP049

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM060621.M
 Title : SVOA CALIBRATION
 Last Update : Tue Jun 08 09:51:19 2021



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.5	133572	PASS
68	69	0.00	2	1.6	2066	PASS
69	198	0.00	100	36.9	131335	PASS
70	69	0.00	2	0.6	823	PASS
127	198	10	80	47.0	167253	PASS
197	198	0.00	2	0.4	1560	PASS
198	198	100	100	100.0	356075	PASS
199	198	5	9	6.9	24587	PASS
275	198	10	60	28.4	101019	PASS
365	198	1	100	4.0	14302	PASS
441	198	0.01	100	11.1	39616	PASS
442	198	50	100	73.3	260864	PASS
443	442	15	24	18.8	49077	PASS

m/z	Abundance
38.80	233.0
39.90	927.0
41.10	419.0
43.00	644.0
44.00	870.0
49.00	443.0
51.00	417.0
55.10	198.0
57.00	512.0
58.90	154.0
68.00	200.0
68.90	286.0
71.10	237.0
73.10	1642.0
74.10	228.0
77.00	464.0
78.10	265.0
81.10	169.0
83.10	250.0
83.90	470.0
85.10	206.0
85.90	195.0
91.00	248.0
95.00	271.0
95.80	187.0
96.90	194.0
105.00	205.0
109.10	151.0
115.00	265.0
119.00	215.0
133.00	328.0
135.00	247.0
147.10	649.0
149.00	216.0
152.00	159.0
163.00	168.0
177.00	158.0
185.10	342.0
186.00	537.0
186.80	185.0
191.00	252.0
193.00	673.0
207.10	2597.0
208.00	591.0
208.90	380.0
221.00	466.0
267.00	342.0
281.10	699.0
281.90	242.0
341.00	175.0
355.10	361.0
356.00	183.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
35.90	333.0
37.00	715.0
38.00	2207.0
39.00	12275.0
40.00	1456.0
41.10	811.0
43.00	783.0
44.00	1258.0
45.00	306.0
47.00	180.0
48.00	193.0
49.00	1768.0
50.00	43368.0
51.00	168384.0
52.10	8600.0
53.00	500.0
55.10	1226.0
56.00	4811.0
57.00	12052.0
58.00	774.0
59.00	519.0
60.00	179.0
61.00	1942.0
62.00	2653.0
63.00	6719.0
64.00	995.0
65.00	3333.0
66.00	280.0
66.90	389.0
68.00	2645.0
69.00	163456.0
70.00	872.0
70.90	213.0
73.10	2712.0
74.00	16133.0
75.00	26360.0
76.10	8700.0
77.10	194112.0
78.00	14155.0
79.00	11458.0
80.00	9499.0
81.00	13119.0
82.00	3432.0
83.00	3291.0
84.00	666.0
85.00	2146.0
86.00	4019.0
87.00	1887.0
88.00	780.0
89.00	358.0
91.00	2594.0
92.00	3538.0
93.00	21488.0
94.00	1405.0
94.90	436.0
96.00	1192.0
96.90	407.0
98.00	16400.0
99.00	15477.0
100.00	1400.0
101.00	7443.0
101.90	455.0
103.00	2568.0
104.00	4428.0
105.00	4633.0
106.00	1388.0
107.00	55576.0
108.00	8443.0
109.00	1497.0
110.00	95984.0
111.00	16840.0
112.00	1781.0
113.00	585.0
114.20	279.0
115.00	392.0
116.00	3318.0
117.00	50464.0
118.00	3861.0
119.00	669.0
120.00	896.0
120.90	393.0
122.00	3922.0
123.00	6160.0
124.00	2702.0
125.00	2372.0
126.10	785.0
127.00	198720.0
128.00	15438.0
129.00	74280.0
130.00	6236.0
131.00	1208.0
131.90	1107.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

133.00	492.0
134.00	2087.0
135.00	6022.0
136.00	2781.0
137.00	4056.0
137.90	559.0
138.90	355.0
140.00	866.0
141.00	9739.0
142.00	2498.0
142.90	2326.0
143.90	538.0
145.00	482.0
146.00	1640.0
147.00	5684.0
148.00	12767.0
149.00	3093.0
150.00	816.0
151.00	1191.0
152.00	1008.0
153.00	2729.0
154.00	2192.0
155.00	5460.0
156.00	8179.0
157.10	1378.0
158.00	1976.0
159.00	1544.0
160.00	3300.0
161.00	4787.0
161.90	1360.0
162.90	368.0
164.10	850.0
165.00	3804.0
166.00	3258.0
167.00	20368.0
168.00	14705.0
169.00	2131.0
170.00	673.0
170.90	1021.0
172.00	1773.0
173.00	2717.0
174.00	4598.0
175.00	7944.0
176.00	2333.0
177.00	3397.0
178.00	1484.0
179.00	15306.0
180.00	11889.0
181.00	5399.0
181.90	753.0
183.00	626.0
184.00	1282.0
185.00	7836.0
186.00	56048.0
187.00	15526.0
188.00	2096.0
189.00	3620.0
190.00	500.0
191.00	1831.0
192.00	5121.0
193.00	6511.0
194.00	1124.0
195.10	536.0
196.00	11096.0
197.00	1496.0
198.00	390912.0
199.00	27880.0
200.00	2181.0
201.50	2095.0
203.00	2870.0
204.00	15861.0
205.00	25152.0
206.00	108008.0
207.10	15940.0
208.00	4172.0
209.00	1610.0
210.00	1565.0
211.00	4425.0
212.90	308.0
215.00	1431.0
216.00	2341.0
217.00	30912.0
218.00	3879.0
219.00	459.0
219.90	422.0
221.00	15374.0
222.00	3950.0
223.00	6688.0
224.00	62240.0
225.00	16193.0
226.10	1947.0
227.00	24920.0
228.00	3472.0
229.00	4845.0
230.00	972.0
231.00	2126.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

231.90	435.0
232.90	325.0
234.00	1769.0
235.00	1884.0
236.00	1438.0
237.00	3053.0
238.00	585.0
239.00	1170.0
239.90	1033.0
241.00	1139.0
242.00	3039.0
243.00	3633.0
244.00	45656.0
245.00	6449.0
246.00	9400.0
247.00	2082.0
248.00	501.0
249.00	1647.0
250.00	168.0
251.00	409.0
251.80	329.0
253.00	1093.0
254.10	1399.0
255.00	236928.0
256.00	34208.0
257.00	3019.0
258.00	14472.0
259.00	2073.0
259.90	422.0
260.90	275.0
263.90	404.0
265.00	5732.0
266.00	898.0
266.90	172.0
270.00	272.0
271.00	577.0
272.00	573.0
273.00	7459.0
274.00	16928.0
275.00	100720.0
276.00	14876.0
277.00	9087.0
278.00	2093.0
279.10	359.0
281.00	820.0
281.90	414.0
283.00	1208.0
284.00	902.0
285.00	1886.0
285.90	393.0
289.00	360.0
289.90	360.0
290.90	188.0
291.90	470.0
293.00	2097.0
293.90	726.0
295.00	787.0
296.00	29760.0
297.00	4652.0
298.00	264.0
300.90	409.0
302.00	597.0
303.00	3291.0
304.00	845.0
305.10	190.0
308.00	423.0
308.90	268.0
309.90	313.0
311.00	151.0
313.00	253.0
314.00	1415.0
315.00	3332.0
316.00	1843.0
317.00	282.0
321.00	1028.0
322.00	492.0
323.00	8866.0
324.00	1786.0
324.90	199.0
325.90	194.0
327.00	2192.0
328.00	997.0
332.00	746.0
333.00	869.0
334.00	5882.0
335.00	1470.0
336.00	325.0
340.00	175.0
341.10	1146.0
342.00	278.0
346.00	2250.0
346.90	558.0
350.90	222.0
352.10	2758.0
353.00	1751.0
354.10	2916.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

355.10	574.0
359.00	160.0
365.00	12958.0
366.00	1822.0
369.90	290.0
370.90	761.0
372.10	3946.0
373.00	1000.0
383.00	1298.0
383.90	312.0
390.00	717.0
391.00	292.0
392.00	286.0
400.90	368.0
402.00	1806.0
403.00	2413.0
404.00	895.0
405.00	355.0
415.10	206.0
421.00	2014.0
422.00	2186.0
423.00	14087.0
424.10	5031.0
425.00	878.0
441.00	32936.0
442.10	225600.0
443.10	43600.0
444.10	4377.0
445.10	187.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
36.00	291.0
37.00	892.0
38.00	2082.0
39.10	11652.0
40.00	1550.0
41.00	752.0
42.00	209.0
43.00	874.0
44.00	1150.0
44.90	488.0
47.90	159.0
49.00	1455.0
50.10	40152.0
51.00	153408.0
52.00	7911.0
53.00	509.0
55.00	1050.0
56.00	4547.0
57.00	10413.0
58.00	603.0
58.90	243.0
61.10	2178.0
62.00	2310.0
63.00	6137.0
64.00	799.0
65.00	2968.0
66.10	226.0
66.90	334.0
68.00	2586.0
69.00	151872.0
70.00	1019.0
71.10	311.0
72.00	169.0
73.00	2610.0
74.00	16792.0
75.00	25544.0
76.10	8923.0
77.00	176512.0
78.00	12452.0
79.00	11518.0
80.00	8346.0
81.00	12035.0
82.00	3366.0
83.00	3155.0
84.00	720.0
85.00	2323.0
86.00	3564.0
87.00	1748.0
87.90	568.0
89.00	402.0
91.00	2852.0
92.00	3221.0
93.00	19608.0
94.00	1589.0
95.00	411.0
96.10	1024.0
97.10	482.0
98.00	14949.0
99.00	13898.0
100.00	1141.0
101.00	6878.0
101.90	375.0
103.00	2000.0
104.00	4122.0
105.00	4355.0
106.00	1355.0
107.00	57112.0
108.00	8777.0
109.00	1553.0
110.00	94168.0
111.00	15096.0
112.00	1974.0
113.00	615.0
115.00	287.0
116.00	2936.0
117.00	49624.0
118.00	3674.0
118.90	620.0
120.00	800.0
121.00	493.0
122.00	4004.0
123.00	5690.0
124.00	2360.0
125.00	2342.0
126.10	887.0
127.00	193664.0
128.00	14975.0
129.00	73608.0
130.00	6814.0
131.00	1497.0
132.00	564.0
132.90	538.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

134.00	2383.0
135.00	6520.0
136.00	2444.0
137.00	3808.0
137.90	876.0
139.00	527.0
139.90	916.0
141.00	10484.0
142.00	3190.0
143.00	2569.0
143.90	641.0
145.00	494.0
146.00	1909.0
147.00	5789.0
148.00	13234.0
149.00	3018.0
149.90	811.0
151.00	1473.0
151.70	1114.0
153.00	3685.0
153.90	2600.0
155.10	5711.0
156.00	9163.0
157.00	1533.0
158.00	1663.0
159.00	1646.0
160.00	3557.0
161.00	4806.0
162.00	1344.0
163.00	556.0
163.90	635.0
165.00	3840.0
166.00	3055.0
167.00	22912.0
168.00	14624.0
169.00	2483.0
170.10	498.0
171.00	875.0
171.90	1913.0
173.00	2287.0
174.00	4802.0
175.00	8748.0
176.00	2179.0
177.00	3811.0
177.90	1379.0
179.00	16568.0
180.00	11803.0
181.00	5685.0
181.90	958.0
182.70	398.0
184.00	1462.0
185.00	7877.0
186.00	61440.0
187.00	17096.0
188.00	1867.0
189.00	3582.0
190.00	904.0
191.00	2240.0
192.00	5456.0
193.00	6419.0
194.00	1257.0
195.10	708.0
196.00	12836.0
196.80	1985.0
198.00	422464.0
199.00	29408.0
200.00	2066.0
201.40	1856.0
203.00	3013.0
204.00	16204.0
205.00	27344.0
206.10	113648.0
207.10	17768.0
208.00	4435.0
209.00	1548.0
210.00	2187.0
211.00	4584.0
211.80	869.0
213.00	372.0
214.00	192.0
214.90	1299.0
216.00	2201.0
217.00	33576.0
218.00	3955.0
219.10	497.0
220.00	620.0
221.00	17448.0
221.90	4241.0
223.00	7144.0
224.00	65256.0
225.00	16944.0
226.00	1944.0
227.00	27376.0
228.00	3792.0
229.00	5589.0
229.90	677.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

231.10	2050.0
232.10	397.0
232.90	500.0
234.00	1562.0
235.00	2002.0
236.00	1129.0
237.00	2949.0
238.00	472.0
239.00	995.0
240.00	961.0
241.00	1525.0
242.00	2966.0
243.00	3634.0
244.10	51176.0
245.00	6517.0
246.00	10381.0
247.00	1936.0
247.80	514.0
249.00	2000.0
249.90	392.0
251.00	595.0
252.00	583.0
252.90	918.0
254.00	1500.0
255.00	269312.0
256.00	38904.0
257.00	3403.0
258.00	16374.0
259.00	2830.0
260.00	530.0
260.90	383.0
264.00	311.0
265.00	6792.0
266.00	758.0
267.00	248.0
269.90	347.0
271.00	672.0
272.00	591.0
273.00	7563.0
274.00	20448.0
275.00	121016.0
276.00	15826.0
277.00	10794.0
278.00	2040.0
279.00	341.0
281.00	726.0
282.00	333.0
283.00	1138.0
284.00	1054.0
285.00	1548.0
285.90	356.0
288.90	390.0
290.00	604.0
290.80	302.0
291.90	559.0
293.00	2576.0
294.00	686.0
294.90	819.0
296.00	37248.0
297.00	4513.0
298.10	272.0
301.10	397.0
301.90	692.0
303.00	3790.0
304.00	947.0
305.00	215.0
308.00	540.0
309.00	269.0
310.00	562.0
312.90	291.0
314.10	1636.0
315.00	4059.0
316.00	1679.0
317.00	442.0
321.00	1073.0
322.00	425.0
323.10	10666.0
324.00	2212.0
325.00	251.0
326.00	311.0
327.00	2279.0
327.90	991.0
329.00	157.0
332.10	544.0
333.00	1091.0
334.00	7384.0
335.00	1791.0
336.00	307.0
339.90	256.0
341.00	1281.0
341.90	332.0
346.00	2640.0
346.90	502.0
350.10	223.0
351.00	243.0
352.00	3101.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

353.00	2412.0
354.00	3268.0
355.00	730.0
358.90	270.0
365.00	18016.0
366.00	2525.0
366.90	181.0
369.90	356.0
370.90	737.0
372.00	6098.0
373.00	1155.0
383.00	1534.0
384.10	564.0
390.00	715.0
391.00	602.0
392.00	329.0
401.00	347.0
402.00	2511.0
403.00	3408.0
404.00	1336.0
404.90	277.0
414.90	214.0
421.00	2772.0
422.10	2611.0
423.00	18304.0
424.00	6295.0
425.00	868.0
440.00	221.0
441.10	49400.0
442.10	312192.0
443.10	60008.0
444.00	5667.0
445.10	364.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
36.00	326.0
37.00	359.0
38.00	1254.0
39.10	5989.0
40.00	1355.0
41.00	622.0
42.00	158.0
43.00	711.0
44.00	1064.0
45.00	269.0
46.90	162.0
49.00	1232.0
50.00	20672.0
51.00	80176.0
52.00	4309.0
53.10	353.0
54.90	608.0
56.00	2925.0
57.00	5608.0
58.10	474.0
59.00	288.0
61.00	1147.0
62.00	1199.0
63.00	3091.0
64.00	366.0
65.00	1728.0
66.10	229.0
67.00	293.0
68.00	1566.0
69.00	79536.0
70.10	577.0
71.00	246.0
73.00	1942.0
74.00	8728.0
75.00	13537.0
76.10	4445.0
77.10	93344.0
78.00	6270.0
79.00	5589.0
80.00	4880.0
81.00	6598.0
82.00	1627.0
83.00	1614.0
83.90	465.0
85.00	1512.0
86.00	2282.0
87.00	863.0
88.00	305.0
89.00	173.0
91.00	1794.0
92.00	1977.0
93.00	11355.0
94.10	707.0
95.00	284.0
96.00	808.0
97.00	520.0
98.00	8256.0
99.00	7480.0
100.00	849.0
101.00	4113.0
102.00	307.0
103.00	1152.0
104.00	2556.0
105.00	2246.0
106.00	719.0
107.00	28888.0
108.00	4355.0
109.00	912.0
110.00	51424.0
111.00	7919.0
112.00	933.0
112.90	430.0
115.00	313.0
116.00	1694.0
117.00	26736.0
117.90	2011.0
118.90	462.0
119.90	394.0
121.00	243.0
121.90	2244.0
123.00	3269.0
124.00	1566.0
125.00	1372.0
126.10	646.0
127.00	109376.0
128.00	7539.0
129.00	41312.0
130.00	3412.0
131.00	588.0
132.10	459.0
133.00	461.0
133.90	1262.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

135.00	3353.0
136.00	1515.0
137.00	2196.0
138.10	395.0
139.00	255.0
139.90	661.0
141.00	6071.0
142.00	2020.0
143.10	1324.0
144.00	517.0
145.10	400.0
146.00	1197.0
147.00	3229.0
148.00	7254.0
149.00	1707.0
149.90	407.0
150.90	611.0
152.00	556.0
153.00	2028.0
154.00	1509.0
155.00	3560.0
156.00	4758.0
157.10	837.0
158.00	1119.0
158.90	787.0
160.00	1610.0
161.00	3117.0
162.00	568.0
163.00	532.0
164.00	365.0
165.00	2250.0
166.00	1734.0
167.00	12130.0
168.00	8526.0
169.00	1607.0
170.00	416.0
171.00	565.0
172.00	1234.0
173.00	1526.0
174.00	2607.0
175.00	4858.0
176.00	1498.0
177.00	2730.0
177.90	933.0
179.00	9741.0
180.00	7360.0
181.00	3293.0
182.00	444.0
182.90	187.0
184.00	763.0
185.00	5253.0
186.00	36824.0
187.00	9401.0
188.00	1331.0
189.00	2145.0
190.00	465.0
191.00	1400.0
192.00	3404.0
193.00	3791.0
194.00	929.0
195.00	532.0
196.00	6905.0
197.00	1200.0
198.00	254848.0
199.00	16472.0
199.90	1350.0
201.40	1035.0
203.00	2017.0
204.00	10621.0
205.00	16912.0
206.00	70400.0
207.00	11988.0
208.00	2472.0
209.00	1335.0
210.20	1622.0
211.00	2528.0
212.80	266.0
215.00	868.0
216.00	1732.0
217.00	20608.0
218.00	2612.0
218.90	267.0
220.10	426.0
221.00	11230.0
222.00	2145.0
223.00	4289.0
224.00	42472.0
225.00	10503.0
226.00	1264.0
227.00	16432.0
228.00	2784.0
229.00	4060.0
230.00	576.0
231.00	1616.0
231.90	345.0
232.80	227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

233.90	1093.0
235.00	1108.0
236.00	1011.0
237.00	1824.0
238.00	374.0
238.90	799.0
239.90	702.0
241.00	1240.0
242.00	2282.0
243.00	2587.0
244.00	34320.0
245.00	4521.0
246.00	6628.0
247.00	1518.0
248.00	268.0
248.90	1252.0
249.90	154.0
251.00	304.0
252.00	276.0
253.00	585.0
254.10	920.0
255.00	174848.0
256.00	25840.0
257.00	2216.0
258.00	10233.0
259.00	1743.0
259.90	355.0
260.80	226.0
263.10	163.0
264.00	267.0
265.00	4128.0
266.00	776.0
266.90	180.0
268.00	175.0
270.00	230.0
270.90	252.0
272.00	434.0
273.00	5176.0
274.00	13036.0
275.00	81320.0
276.00	10667.0
277.00	6722.0
278.00	1177.0
278.90	299.0
281.00	863.0
282.00	359.0
283.00	719.0
283.90	557.0
285.00	983.0
286.00	231.0
288.90	214.0
289.90	284.0
291.00	172.0
292.00	388.0
293.00	2064.0
294.00	315.0
294.90	408.0
296.00	24704.0
297.00	3785.0
297.90	262.0
300.90	224.0
301.90	392.0
303.00	2669.0
303.90	576.0
305.10	193.0
308.00	276.0
308.80	299.0
309.90	283.0
312.00	158.0
313.00	263.0
314.00	1330.0
315.00	3050.0
316.00	1504.0
317.00	259.0
321.00	807.0
322.00	360.0
323.00	7339.0
323.90	1483.0
326.00	157.0
327.00	1737.0
328.10	823.0
329.10	357.0
332.00	601.0
333.00	763.0
334.00	5092.0
335.00	1048.0
341.00	806.0
342.00	302.0
346.00	1758.0
346.80	413.0
350.90	219.0
352.00	2376.0
353.00	1221.0
354.10	2438.0
355.10	445.0
359.10	338.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

365.00	11932.0
365.90	1500.0
369.90	311.0
370.90	682.0
372.00	4321.0
373.00	1083.0
376.90	150.0
383.00	1184.0
383.90	275.0
389.90	391.0
390.90	320.0
391.90	227.0
400.90	509.0
402.00	1838.0
403.00	2494.0
404.00	1088.0
405.00	239.0
415.00	151.0
421.00	2125.0
422.00	1874.0
423.00	14267.0
424.00	4475.0
425.00	436.0
441.00	36512.0
442.00	244800.0
443.10	43624.0
444.00	4271.0
445.00	170.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049