

Data Path : Z:\HPCHEM1\BNA_M\Data\BM060815\
 Data File : BM001604.D
 Acq On : 09 Jun 2015 03:36
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM060815

Quant Time: Jun 09 18:10:11 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 09 17:36:38 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.550	0.539	2.0	95	0.00
3	n-Nitrosodimethylamine	0.787	0.826	-5.0	98	0.00
4 S	2-Fluorophenol	1.137	1.163	-2.3	96	-0.02
5 S	Phenol-d6	1.397	1.408	-0.8	96	-0.02
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.352	0.347	1.4	95	0.00
8	Nitrobenzene	0.378	0.373	1.3	95	0.00
9	Naphthalene	1.120	1.151	-2.8	98	0.00
10	Hexachlorobutadiene	0.193	0.195	-1.0	96	0.00
11	2-Methylnaphthalene	0.686	0.704	-2.6	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
13 S	2,4,6-Tribromophenol	0.202	0.215	-6.4	104	0.00
14 S	2-Fluorobiphenyl	1.628	1.672	-2.7	99	0.00
15	Acenaphthylene	2.245	2.273	-1.2	99	0.00
16	Acenaphthene	1.412	1.462	-3.5	101	0.00
17	Fluorene	0.956	1.024	-7.1	105	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.03
19	4-Bromophenyl-phenylether	0.474	0.521	-9.9	100	0.00
20	Hexachlorobenzene	0.259	0.294	-13.5	106	0.00
21	Pentachlorophenol	0.087	0.098	-12.6	104	0.00
22	Phenanthrene	1.785	1.999	-12.0	93	0.00
23	Anthracene	1.059	1.257	-18.7	112	0.00
24	Fluoranthene	1.933	2.080	-7.6	103	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
26	Pyrene	1.620	1.662	-2.6	104	0.00
27 S	Terphenyl-d14	0.652	0.680	-4.3	104	0.00
28	Benzo(a)anthracene	1.460	1.457	0.2	101	0.00
29	Chrysene	1.380	1.450	-5.1	102	0.00
30	Bis(2-ethylhexyl)phthalate	0.844	0.752	10.9	96	0.00
31	Indeno(1,2,3-cd)pyrene	1.402	1.356	3.3	98	0.00
32 I	Perylene-d12	1.000	1.000	0.0	101	0.00
33	Benzo(b)fluoranthene	1.503	1.458	3.0	102	0.00
34	Benzo(k)fluoranthene	1.450	1.407	3.0	99	0.00
35 C	Benzo(a)pyrene	1.327	1.250	5.8	98	0.00
36	Dibenzo(a,h)anthracene	1.213	1.119	7.7	97	0.00
37	Benzo(g,h,i)perylene	1.308	1.230	6.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0