

Data Path : Z:\HPCHEM1\BNA_M\Data\BM060815\
 Data File : BM001623.D
 Acq On : 09 Jun 2015 20:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 10 04:04:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 03:26:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.550	0.514	6.5	105	0.00
3	n-Nitrosodimethylamine	0.787	0.771	2.0	106	0.00
4 S	2-Fluorophenol	1.137	1.139	-0.2	109	-0.02
5 S	Phenol-d6	1.397	1.415	-1.3	112	-0.02
6 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
7 S	Nitrobenzene-d5	0.352	0.350	0.6	113	0.00
8	Nitrobenzene	0.378	0.375	0.8	114	0.00
9	Naphthalene	1.120	1.117	0.3	113	0.00
10	Hexachlorobutadiene	0.193	0.187	3.1	109	0.00
11	2-Methylnaphthalene	0.686	0.690	-0.6	115	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
13 S	2,4,6-Tribromophenol	0.202	0.203	-0.5	122	0.00
14 S	2-Fluorobiphenyl	1.628	1.574	3.3	116	0.00
15	Acenaphthylene	2.245	2.204	1.8	120	0.00
16	Acenaphthene	1.412	1.376	2.5	118	0.00
17	Fluorene	0.956	1.077	-12.7	137	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.03
19	4-Bromophenyl-phenylether	0.474	0.500	-5.5	116	0.00
20	Hexachlorobenzene	0.259	0.335	-29.3#	147	0.00
21	Pentachlorophenol	0.087	0.108	-24.1	138	0.00
22	Phenanthrene	1.785	1.696	5.0	95	0.00
23	Anthracene	1.059	1.442	-36.2#	156#	0.00
24	Fluoranthene	1.933	2.115	-9.4	126	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	128	-0.01
26	Pyrene	1.620	1.549	4.4	128	0.00
27 S	Terphenyl-d14	0.652	0.636	2.5	129	0.00
28	Benzo(a)anthracene	1.460	1.402	4.0	129	0.00
29	Chrysene	1.380	1.346	2.5	126	0.00
30	Bis(2-ethylhexyl)phthalate	0.844	0.750	11.1	127	0.00
31	Indeno(1,2,3-cd)pyrene	1.402	1.385	1.2	132	0.00
32 I	Perylene-d12	1.000	1.000	0.0	130	0.00
33	Benzo(b)fluoranthene	1.503	1.503	0.0	135	0.00
34	Benzo(k)fluoranthene	1.450	1.391	4.1	125	0.00
35 C	Benzo(a)pyrene	1.327	1.301	2.0	132	0.00
36	Dibenzo(a,h)anthracene	1.213	1.196	1.4	133	0.00
37	Benzo(g,h,i)perylene	1.308	1.270	2.9	131	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0