

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060815\
 Data File : BM001610.D
 Acq On : 09 Jun 2015 07:12
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Manual Integrations
 APPROVED

apatel
 6/10/2015 1:04:57 PM

Quant Time: Jun 09 18:18:26 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 09 17:36:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	20835	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	74025	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	34089	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	41172	5.00	ng	0.03
25) Chrysene-d12	21.28	240	62429	5.00	ng	0.00
32) Perylene-d12	23.51	264	56978	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.27	112	4670	0.99	ng	-0.02
5) Phenol-d6	6.84	99	5791	0.99	ng	-0.02
7) Nitrobenzene-d5	8.85	82	5091	0.98	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	1357	0.94	ng	0.00
14) 2-Fluorobiphenyl	12.96	172	10896	0.98	ng	0.00
27) Terphenyl-d14	19.72	244	7674	0.94	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	2189	0.95	ng	100
3) n-Nitrosodimethylamine	3.51	42	3242	0.99	ng	98
8) Nitrobenzene	8.89	77	5445	0.97	ng	99
9) Naphthalene	10.52	128	16627	1.00	ng	99
10) Hexachlorobutadiene	10.82	225	2781	0.97	ng	100
11) 2-Methylnaphthalene	12.15	142	10016	0.99	ng	97
15) Acenaphthylene	14.06	152	14916	0.97	ng	100
16) Acenaphthene	14.40	154	9293	0.97	ng	98
17) Fluorene	15.36	166	7069	1.08	ng	97
19) 4-Bromophenyl-phenylether	16.27	248	4144	1.06	ng	88
20) Hexachlorobenzene	16.38	284	2827	1.32	ng	# 65
21) Pentachlorophenol	16.72	266	936	1.31	ng	96
22) Phenanthrene	17.12	178	14998m	1.02	ng	
23) Anthracene	17.19	178	11531m	1.32	ng	
24) Fluoranthene	19.14	202	17767	1.12	ng	100
26) Pyrene	19.52	202	18942	0.94	ng	99
28) Benzo(a)anthracene	21.25	228	17508	0.96	ng	98
29) Chrysene	21.31	228	17324	1.01	ng	99
30) Bis(2-ethylhexyl)phthalate	21.19	149	8960	0.85	ng	99
31) Indeno(1,2,3-cd)pyrene	25.76	276	17673	1.01	ng	99
33) Benzo(b)fluoranthene	22.83	252	16904	0.99	ng	99
34) Benzo(k)fluoranthene	22.88	252	16272	0.98	ng	99
35) Benzo(a)pyrene	23.41	252	14828	0.98	ng	96
36) Dibenzo(a,h)anthracene	25.78	278	13884	1.00	ng	98
37) Benzo(g,h,i)perylene	26.45	276	14702	0.99	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060815\
Data File : BM001610.D
Acq On : 09 Jun 2015 07:12
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM

Quant Time: Jun 09 18:18:26 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 09 17:36:38 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.70 min Scan# 325

Delta R.T. 0.00 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :

SSTDCCC001EC



Tgt Ion:152 Resp: 20835

Ion Ratio Lower Upper




152 100

150 145.9 120.1 180.1

115 55.6 47.0 70.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#2

1,4-Dioxane

Concen: 0.95 ng

RT: 3.21 min Scan# 35

Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12




Tgt Ion: 88 Resp: 2189


Ion Ratio Lower Upper

88 100

58 82.4 65.5 98.3

43 40.1 32.2 48.2





#3

n-Nitrosodimethylamine

Concen: 0.99 ng

RT: 3.51 min Scan# 54

Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :





SSTDCCC001EC

Manual Integrations
APPROVED

apatel

6/10/2015 1:04:57 PM

Tgt Ion:	42	Resp:	3242
Ion	Ratio	Lower	Upper
42	100		
74	97.8	79.5	119.3
44	8.2	5.7	8.5



#4

2-Fluorophenol

Concen: 0.99 ng




RT: 5.27 min Scan# 168


Delta R.T. -0.02 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Tgt Ion:	112	Resp:	4670
Ion	Ratio	Lower	Upper
112	100		
64	67.9	54.8	82.2
63	36.9	30.5	45.7





#5

Phenol-d6

Concen: 0.99 ng

RT: 6.84 min Scan# 270

Delta R.T. -0.02 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tgt Ion:	99	Resp:	5791
Ion Ratio	Lower	Upper	
99	100		
42	26.1	21.6	32.4
71	35.2	28.7	43.1

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.47 min Scan# 552


Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Tgt Ion:	136	Resp:	74025
Ion Ratio	Lower	Upper	
136	100		
137	10.5	8.3	12.5
54	10.0	9.2	13.8
68	6.3	5.5	8.3





#7

Nitrobenzene-d5

Concen: 0.98 ng

RT: 8.85 min Scan# 421

Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tgt Ion:	82	Resp:	5091
Ion Ratio	Lower	Upper	
82	100		
128	35.1	26.6	40.0
54	61.3	53.0	79.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#8

Nitrobenzene

Concen: 0.97 ng




RT: 8.89 min Scan# 425


Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Tgt Ion:	77	Resp:	5445
Ion Ratio	Lower	Upper	
77	100		
123	38.5	31.0	46.6
65	13.8	11.4	17.0





#9

Naphthalene

Concen: 1.00 ng

RT: 10.52 min Scan# 555

Delta R.T. 0.00 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001EC



Tgt Ion:	128	Resp:	16627
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.1	13.7
127	12.4	10.4	15.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#10

Hexachlorobutadiene


Concen: 0.97 ng

RT: 10.82 min Scan# 572



Delta R.T. 0.00 min


Lab File: BM001610.D

Acq: 09 Jun 2015 07:12



Tgt Ion:	225	Resp:	2781
Ion Ratio	Lower	Upper	
225	100		
223	62.8	49.8	74.8
227	63.6	50.9	76.3









#11
2-Methylnaphthalene
Concen: 0.99 ng
RT: 12.15 min Scan# 681
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
142	100		
141	86.9	71.3	106.9
115	37.1	32.1	48.1




Manual Integrations
APPROVED


apatel
6/10/2015 1:04:57 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.35 min Scan# 880
Delta R.T. 0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
164	100		
162	95.0	79.8	119.6
160	39.4	34.6	51.8






#13
 2,4,6-Tribromophenol
 Concen: 0.94 ng
 RT: 15.83 min Scan# 943
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC


Tgt Ion	Ratio	Resp	Lower	Upper
330	100			
332	100.4	76.7	115.1	
141	0.0	0.0	0.0	

Manual Integrations
 APPROVED



apatel
 6/10/2015 1:04:57 PM



Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E




Time-->




#14
 2-Fluorobiphenyl
 Concen: 0.98 ng
 RT: 12.96 min Scan# 759
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12


Tgt Ion	Ratio	Resp	Lower	Upper
172	100			
171	35.9	29.6	44.4	
170	24.4	20.6	31.0	




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E



Time-->





#15

Acenaphthylene

Concen: 0.97 ng

RT: 14.06 min Scan# 856

Delta R.T. 0.01 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :



SSTDCCC001EC



Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.5	23.3
153	12.9	10.3	15.5

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#16

Acenaphthene

Concen: 0.97 ng




RT: 14.40 min Scan# 884


Delta R.T. -0.01 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.8	89.4	134.0
152	55.6	43.6	65.4





#17

Fluorene

Concen: 1.08 ng

RT: 15.36 min Scan# 929

Delta R.T. -0.01 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :

SSTDCCC001EC



Tgt Ion:166 Resp: 7069

Ion Ratio Lower Upper




166 100

165 113.4 87.8 131.8

167 12.1 10.0 15.0

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 17.09 min Scan# 980

Delta R.T. 0.03 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12




Tgt Ion:188 Resp: 41172


Ion Ratio Lower Upper

188 100

94 1.5 1.0 1.6

80 1.5 1.0 1.4#









#19
4-Bromophenyl-phenylether
Concen: 1.06 ng
RT: 16.27 min Scan# 956
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
248	100		
250	100.1	68.6	103.0
141	33.0	25.5	38.3




Manual Integrations
APPROVED


apatel
6/10/2015 1:04:57 PM



#20
Hexachlorobenzene
Concen: 1.32 ng
RT: 16.38 min Scan# 959
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
284	100		
142	35.8	47.3	70.9#
249	57.7	75.8	113.6#






#21
 Pentachlorophenol
 Concen: 1.31 ng
 RT: 16.72 min Scan# 969
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Instrument :
 BNA_M
 ClientSampleId :
 SSTCCCC001EC




Tgt Ion	Ratio	Lower	Upper
266	100		
268	55.0	45.8	68.8
264	84.7	70.6	106.0

Manual Integrations
 APPROVED

apatel
 6/10/2015 1:04:57 PM






Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E




#22
 Phenanthrene
 Concen: 1.02 ng m
 RT: 17.12 min Scan# 981
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
178	100		
176	31.8	13.2	19.8#
179	27.2	13.8	20.8#

Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#23
 Anthracene
 Concen: 1.32 ng m
 RT: 17.19 min Scan# 983
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
178	100		
176	38.1	21.9	32.9#
179	33.6	23.0	34.4




Manual Integrations
 APPROVED


apatel
 6/10/2015 1:04:57 PM



#24
 Fluoranthene
 Concen: 1.12 ng
 RT: 19.14 min Scan# 1088
 Delta R.T. -0.01 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.2	10.4	15.6
203	17.3	13.8	20.6





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.28 min Scan# 1252

Delta R.T. 0.00 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :



SSTDCCC001EC



Tgt Ion:	240	Resp:	62429
Ion Ratio	Lower	Upper	
240	100		
120	2.1	2.2	3.4#
236	23.4	19.6	29.4

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#26

Pyrene


Concen: 0.94 ng

RT: 19.52 min Scan# 1119



Delta R.T. 0.00 min


Lab File: BM001610.D

Acq: 09 Jun 2015 07:12




Tgt Ion:	202	Resp:	18942
Ion Ratio	Lower	Upper	
202	100		
200	20.5	16.7	25.1
203	17.4	14.0	21.0





#27
 Terphenyl-d14
 Concen: 0.94 ng
 RT: 19.72 min Scan# 1136
 Delta R.T. 0.00 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC



Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.6	9.2	13.8
122	12.8	14.5	21.7#




Manual Integrations
 APPROVED


apatel
 6/10/2015 1:04:57 PM



#28
 Benzo(a)anthracene
 Concen: 0.96 ng
 RT: 21.25 min Scan# 1250
 Delta R.T. 0.00 min
 Lab File: BM001610.D
 Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.2	24.6	37.0
229	18.0	13.8	20.8





#29

Chrysene

Concen: 1.01 ng

RT: 21.31 min Scan# 1254

Delta R.T. 0.00 min

Lab File: BM001610.D


Acq: 09 Jun 2015 07:12

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001EC



Tgt Ion:	228	Resp:	17324
Ion Ratio	Lower	Upper	
228	100		
226	25.3	20.6	31.0
229	22.2	17.4	26.2

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.85 ng




RT: 21.19 min Scan# 1246


Delta R.T. 0.00 min

Lab File: BM001610.D

Acq: 09 Jun 2015 07:12

Tgt Ion:	149	Resp:	8960
Ion Ratio	Lower	Upper	
149	100		
167	25.9	20.2	30.4
279	1.9	1.7	2.5






#31
Indeno(1,2,3-cd)pyrene
Concen: 1.01 ng
RT: 25.76 min Scan# 1659
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
276	100		
138	36.3	28.2	42.2
277	24.8	19.4	29.2

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM






Abundance

Ion 276.00 (275.70 to 276.70): E


Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.51 min Scan# 1443
Delta R.T. 0.00 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.2	21.8	32.6
265	21.3	16.8	25.2






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#33
Benzo(b)fluoranthene
Concen: 0.99 ng
RT: 22.83 min Scan# 1378
Delta R.T. 0.00 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.6	26.4
125	14.9	12.4	18.6




Manual Integrations
APPROVED


apatel
6/10/2015 1:04:57 PM



#34
Benzo(k)fluoranthene
Concen: 0.98 ng
RT: 22.88 min Scan# 1382
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.1	25.7
125	15.5	12.8	19.2









#35
Benzo(a)pyrene
Concen: 0.98 ng
RT: 23.41 min Scan# 1433
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	19.4	29.0
125	17.0	13.0	19.6




Manual Integrations
APPROVED




apatel
6/10/2015 1:04:57 PM



#36
Dibenzo(a,h)anthracene
Concen: 1.00 ng
RT: 25.78 min Scan# 1661
Delta R.T. 0.00 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Tgt Ion	Ratio	Lower	Upper
278	100		
139	23.5	20.6	31.0
279	24.7	19.7	29.5





#37
Benzo(g,h,i)perylene
Concen: 0.99 ng
RT: 26.45 min Scan# 1725
Delta R.T. -0.01 min
Lab File: BM001610.D
Acq: 09 Jun 2015 07:12

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.4	29.0
138	32.7	24.5	36.7

Manual Integrations
APPROVED

apatel
6/10/2015 1:04:57 PM

