

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060815\
 Data File : BM001623.D
 Acq On : 09 Jun 2015 20:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Manual Integrations
 APPROVED

apatel
 6/10/2015 1:05:59 PM

Quant Time: Jun 10 04:04:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 10 03:26:36 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	20881	5.00	ng	0.00
6) Naphthalene-d8	10.47	136	75196	5.00	ng	0.00
12) Acenaphthene-d10	14.35	164	35917	5.00	ng	0.00
18) Phenanthrene-d10	17.09	188	44372	5.00	ng	0.00
25) Chrysene-d12	21.27	240	64096	5.00	ng	-0.02
32) Perylene-d12	23.51	264	58053	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.27	112	4758	1.00	ng	0.00
5) Phenol-d6	6.84	99	5909	1.01	ng	0.00
7) Nitrobenzene-d5	8.85	82	5257	0.99	ng	0.00
13) 2,4,6-Tribromophenol	15.83	330	1457	0.96	ng	0.00
14) 2-Fluorobiphenyl	12.96	172	11306	0.97	ng	0.00
27) Terphenyl-d14	19.72	244	8155	0.98	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.21	88	2147	0.93	ng	98
3) n-Nitrosodimethylamine	3.51	42	3219	0.98	ng	99
8) Nitrobenzene	8.89	77	5636	0.99	ng	100
9) Naphthalene	10.52	128	16792	1.00	ng	99
10) Hexachlorobutadiene	10.82	225	2807	0.97	ng	99
11) 2-Methylnaphthalene	12.15	142	10375	1.01	ng	96
15) Acenaphthylene	14.06	152	15834	0.98	ng	100
16) Acenaphthene	14.40	154	9886	0.97	ng	99
17) Fluorene	15.36	166	7739	1.13	ng	97
19) 4-Bromophenyl-phenylether	16.27	248	4441	1.06	ng	# 74
20) Hexachlorobenzene	16.38	284	2975	1.29	ng	# 80
21) Pentachlorophenol	16.72	266	954	1.24	ng	97
22) Phenanthrene	17.12	178	15048	0.95	ng	99
23) Anthracene	17.19	178	12793m	1.36	ng	
24) Fluoranthene	19.14	202	18766	1.09	ng	100
26) Pyrene	19.50	202	19855	0.96	ng	99
28) Benzo(a)anthracene	21.25	228	17971	0.96	ng	97
29) Chrysene	21.31	228	17260	0.98	ng	98
30) Bis(2-ethylhexyl)phthalate	21.19	149	9610	0.89	ng	99
31) Indeno(1,2,3-cd)pyrene	25.76	276	17751	0.99	ng	99
33) Benzo(b)fluoranthene	22.83	252	17452	1.00	ng	98
34) Benzo(k)fluoranthene	22.88	252	16155	0.96	ng	98
35) Benzo(a)pyrene	23.41	252	15109	0.98	ng	97
36) Dibenzo(a,h)anthracene	25.78	278	13891	0.99	ng	96
37) Benzo(g,h,i)perylene	26.45	276	14746	0.97	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM060815\
Data File : BM001623.D
Acq On : 09 Jun 2015 20:11
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM

Quant Time: Jun 10 04:04:22 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SIM-BM060815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 10 03:26:36 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.69 min Scan# 325

Delta R.T. 0.00 min

Lab File: BM001623.D


Acq: 09 Jun 2015 20:11

Instrument :

BNA_M

ClientSampleId :

SSTDCCC001EC



Tgt Ion:152 Resp: 20881

Ion Ratio Lower Upper




152 100

150 145.4 120.1 180.1

115 54.8 47.0 70.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#2

1,4-Dioxane

Concen: 0.93 ng

RT: 3.21 min Scan# 35

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11




Tgt Ion: 88 Resp: 2147


Ion Ratio Lower Upper

88 100

58 83.8 65.5 98.3

43 39.4 32.2 48.2





#3

n-Nitrosodimethylamine

Concen: 0.98 ng

RT: 3.51 min Scan# 54

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Instrument :

BNA_M


ClientSampleId :

SSTDCCC001EC

Tgt Ion:	42	Resp:	3219
Ion	Ratio	Lower	Upper
42	100		
74	100.6	79.5	119.3
44	6.7	5.7	8.5

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM




Abundance Ion 42.00 (41.70 to 42.70): BM001599.D (-52) (-)


Ion 74.00 (73.70 to 74.70): BM001599.D (-52) (-)

Ion 44.00 (43.70 to 44.70): BM001599.D (-52) (-)

Time-->



Time-->



#4

2-Fluorophenol

Concen: 1.00 ng


RT: 5.27 min Scan# 168

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	112	Resp:	4758
Ion	Ratio	Lower	Upper
112	100		
64	67.9	54.8	82.2
63	36.9	30.5	45.7




Abundance Ion 112.00 (111.70 to 112.70): BM001599.D (-165) (-)


Ion 64.00 (63.70 to 64.70): BM001599.D (-165) (-)

Ion 63.00 (62.70 to 63.70): BM001599.D (-165) (-)

Time-->



Time-->



#5

Phenol-d6

Concen: 1.01 ng

RT: 6.84 min Scan# 270

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tgt Ion:	99	Resp:	5909
Ion	Ratio	Lower	Upper
99	100		
42	26.3	21.6	32.4
71	35.4	28.7	43.1

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.47 min Scan# 552


Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	136	Resp:	75196
Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.3	12.5
54	9.8	9.2	13.8
68	6.1	5.5	8.3





#7

Nitrobenzene-d5

Concen: 0.99 ng

RT: 8.85 min Scan# 421

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Instrument :

BNA_M





ClientSampleId :

SSTDCCC001EC

Tgt Ion:	82	Resp:	5257
Ion	Ratio	Lower	Upper
82	100		
128	36.1	26.6	40.0
54	59.6	53.0	79.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#8

Nitrobenzene

Concen: 0.99 ng




RT: 8.89 min Scan# 425


Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	77	Resp:	5636
Ion	Ratio	Lower	Upper
77	100		
123	38.5	31.0	46.6
65	14.2	11.4	17.0





#9

Naphthalene

Concen: 1.00 ng

RT: 10.52 min Scan# 555

Delta R.T. 0.00 min

Lab File: BM001623.D


Acq: 09 Jun 2015 20:11

Instrument :

BNA_M

ClientSampleId :




SSTDCCC001EC



Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.6	9.1	13.7
127	12.7	10.4	15.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#10

Hexachlorobutadiene

Concen: 0.97 ng




RT: 10.82 min Scan# 572


Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.2	49.8	74.8
227	63.4	50.9	76.3









#11
2-Methylnaphthalene
Concen: 1.01 ng
RT: 12.15 min Scan# 681
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
142	100		
141	85.8	71.3	106.9
115	36.2	32.1	48.1




Manual Integrations
APPROVED


apatel
6/10/2015 1:05:59 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.35 min Scan# 880
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
164	100		
162	94.6	79.8	119.6
160	39.1	34.6	51.8









#13
 2,4,6-Tribromophenol
 Concen: 0.96 ng
 RT: 15.83 min Scan# 943
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	1457		
332	100.1	76.7	115.1	
141	0.0	0.0	0.0	




Manual Integrations
 APPROVED


apatel
 6/10/2015 1:05:59 PM



#14
 2-Fluorobiphenyl
 Concen: 0.97 ng
 RT: 12.96 min Scan# 759
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	11306		
171	34.5	29.6	44.4	
170	23.2	20.6	31.0	









#15
Acenaphthylene
Concen: 0.98 ng
RT: 14.06 min Scan# 856
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.5	23.3
153	12.9	10.3	15.5




Manual Integrations
APPROVED


apatel
6/10/2015 1:05:59 PM



#16
Acenaphthene
Concen: 0.97 ng
RT: 14.40 min Scan# 884
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.3	89.4	134.0
152	54.0	43.6	65.4









#17
Fluorene
 Concen: 1.13 ng
 RT: 15.36 min Scan# 929
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
166	100		
165	113.3	87.8	131.8
167	12.2	10.0	15.0




Manual Integrations
 APPROVED


apatel
 6/10/2015 1:05:59 PM



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.09 min Scan# 980
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
188	100		
94	1.3	1.0	1.6
80	1.3	1.0	1.4






#19
4-Bromophenyl-phenylether
Concen: 1.06 ng
RT: 16.27 min Scan# 956
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC




Tgt Ion	Ratio	Lower	Upper
248	100		
250	98.8	68.6	103.0
141	0.0	25.5	38.3

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM






Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#20
Hexachlorobenzene
Concen: 1.29 ng
RT: 16.38 min Scan# 959
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.7	47.3	70.9
249	80.1	75.8	113.6

Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E





#21

Pentachlorophenol

Concen: 1.24 ng

RT: 16.72 min Scan# 969

Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	266	Resp:	954
Ion Ratio	Lower	Upper	
266	100		
268	55.7	45.8	68.8
264	85.2	70.6	106.0

Instrument :





BNA_M

ClientSampleId :

SSTDCCC001EC

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#22

Phenanthrene

Concen: 0.95 ng


RT: 17.12 min Scan# 981


Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	178	Resp:	15048
Ion Ratio	Lower	Upper	
178	100		
176	16.1	13.2	19.8
179	18.0	13.8	20.8





#23

Anthracene

Concen: 1.36 ng m

RT: 17.19 min Scan# 983

Delta R.T. 0.00 min

Lab File: BM001623.D


Acq: 09 Jun 2015 20:11

Instrument :

BNA_M

ClientSampleId :



SSTDCCC001EC



Tgt Ion:	178	Resp:	12793
Ion Ratio	Lower	Upper	
178	100		
176	18.9	21.9	32.9#
179	34.1	23.0	34.4

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#24

Fluoranthene

Concen: 1.09 ng




RT: 19.14 min Scan# 1088


Delta R.T. -0.01 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	202	Resp:	18766
Ion Ratio	Lower	Upper	
202	100		
101	13.1	10.4	15.6
203	17.0	13.8	20.6









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.27 min Scan# 1251
Delta R.T. -0.02 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
240	100		
120	22.8	2.2	3.4#
236	33.0	19.6	29.4#




Manual Integrations
APPROVED


apatel
6/10/2015 1:05:59 PM



#26
Pyrene
Concen: 0.96 ng
RT: 19.50 min Scan# 1118
Delta R.T. -0.01 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.7	25.1
203	17.3	14.0	21.0





#27

Terphenyl-d14

Concen: 0.98 ng

RT: 19.72 min Scan# 1136

Delta R.T. 0.00 min

Lab File: BM001623.D


Acq: 09 Jun 2015 20:11

Instrument :

BNA_M

ClientSampleId :



SSTDCCC001EC



Tgt Ion:	244	Resp:	8155
Ion Ratio	Lower	Upper	
244	100		
212	10.2	9.2	13.8
122	12.1	14.5	21.7#

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM



#28

Benzo(a)anthracene

Concen: 0.96 ng




RT: 21.25 min Scan# 1250


Delta R.T. 0.00 min

Lab File: BM001623.D

Acq: 09 Jun 2015 20:11

Tgt Ion:	228	Resp:	17971
Ion Ratio	Lower	Upper	
228	100		
226	28.9	24.6	37.0
229	18.1	13.8	20.8









#29
Chrysene
 Concen: 0.98 ng
 RT: 21.31 min Scan# 1254
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.0	20.6	31.0
229	23.0	17.4	26.2




Manual Integrations
 APPROVED


apatel
 6/10/2015 1:05:59 PM



#30
Bis(2-ethylhexyl)phthalate
 Concen: 0.89 ng
 RT: 21.19 min Scan# 1246
 Delta R.T. 0.00 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.8	20.2	30.4
279	2.0	1.7	2.5






#31
 Indeno(1,2,3-cd)pyrene
 Concen: 0.99 ng
 RT: 25.76 min Scan# 1659
 Delta R.T. -0.01 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC




Tgt Ion	Ratio	Lower	Upper
276	100		
138	35.9	28.2	42.2
277	24.4	19.4	29.2

Manual Integrations
 APPROVED

apatel
 6/10/2015 1:05:59 PM






Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E




#32
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.51 min Scan# 1443
 Delta R.T. -0.01 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.1	21.8	32.6
265	22.0	16.8	25.2

Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E






#33
Benzo(b)fluoranthene
Concen: 1.00 ng
RT: 22.83 min Scan# 1378
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.6	26.4
125	14.4	12.4	18.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM






Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E




#34
Benzo(k)fluoranthene
Concen: 0.96 ng
RT: 22.88 min Scan# 1382
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.1	25.7
125	15.0	12.8	19.2

Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#35
Benzo(a)pyrene
Concen: 0.98 ng
RT: 23.41 min Scan# 1433
Delta R.T. -0.01 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.4	19.4	29.0
125	16.8	13.0	19.6

Manual Integrations
APPROVED

apatel
6/10/2015 1:05:59 PM






Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E




#36
Dibenzo(a,h)anthracene
Concen: 0.99 ng
RT: 25.78 min Scan# 1661
Delta R.T. 0.00 min
Lab File: BM001623.D
Acq: 09 Jun 2015 20:11

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.4	20.6	31.0
279	24.4	19.7	29.5


Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#37
 Benzo(g,h,i)perylene
 Concen: 0.97 ng
 RT: 26.45 min Scan# 1725
 Delta R.T. -0.01 min
 Lab File: BM001623.D
 Acq: 09 Jun 2015 20:11

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC



Tgt Ion:	276	Resp:	14746
Ion Ratio	Lower	Upper	
276	100		
277	24.2	19.4	29.0
138	32.2	24.5	36.7

Manual Integrations
 APPROVED

apatel
 6/10/2015 1:05:59 PM

