

Data Path : Z:\HPCHEM1\BNA_M\Data\BM060817\
 Data File : BM010440.D
 Acq On : 07 Jun 2017 20:59
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.422

Quant Time: Jun 08 07:09:00 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM053017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 06 01:03:41 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	1.055	1.013	4.0	99	0.00
4 SURR	2-Methylnaphthalene-d10	0.581	0.619	-6.5	109	0.00
5	2-Methylnaphthalene	0.744	0.773	-3.9	107	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
7	Acenaphthylene	1.997	2.026	-1.5	112	0.00
8 C	Acenaphthene	1.510	1.498	0.8	110	0.00
9	Fluorene	1.808	1.776	1.8	109	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
11	Pentachlorophenol	0.153	0.151	1.3	106	0.00
12	Phenanthrene	1.286	1.223	4.9	103	0.00
13	Anthracene	1.204	1.184	1.7	107	0.00
14 SURR	Fluoranthene-d10	1.061	0.981	7.5	97	0.00
15 C	Fluoranthene	1.595	1.456	8.7	98	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
17	Pyrene	1.704	1.805	-5.9	94	0.00
18	Benzo(a)anthracene	1.626	1.642	-1.0	90	0.00
19	Chrysene	1.573	1.520	3.4	84	0.00
20 I	Perylene-d12	1.000	1.000	0.0	84	0.00
21	Benzo(b)fluoranthene	1.751	1.681	4.0	82	0.00
22	Benzo(k)fluoranthene	1.665	1.621	2.6	83	0.00
23 C	Benzo(a)pyrene	1.657	1.624	2.0	83	0.00
24	Indeno(1,2,3-cd)pyrene	2.178	2.128	2.3	83	0.00
25	Dibenzo(a,h)anthracene	1.735	1.699	2.1	83	0.00
26	Benzo(g,h,i)perylene	1.878	1.799	4.2	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0