

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060820\
 Data File : BM026291.D
 Acq On : 08 Jun 2020 17:44
 Operator : JU/CG
 Sample : L2890-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CHRT-26160

Manual Integrations
 APPROVED

mohammad
 6/9/2020 12:21:46 PM

Quant Time: Jun 08 18:46:57 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 15:00:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	233120	20.00	ng	0.00
21) Naphthalene-d8	10.20	136	996422	20.00	ng	0.00
39) Acenaphthene-d10	14.10	164	630062	20.00	ng	0.00
64) Phenanthrene-d10	16.87	188	1307170	20.00	ng	0.03
76) Chrysene-d12	21.06	240	924488	20.00	ng	0.00
86) Perylene-d12	23.16	264	905500	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.08	112	1021663	77.49	ng	0.00
7) Phenol-d6	6.65	99	1538873	80.60	ng	0.00
23) Nitrobenzene-d5	8.59	82	1144265	56.38	ng	0.00
42) 2,4,6-Tribromophenol	15.61	330	701135	91.86	ng	0.02
45) 2-Fluorobiphenyl	12.71	172	2463990	59.36	ng	0.00
79) Terphenyl-d14	19.52	244	3297513	69.24	ng	0.02
Target Compounds						
50) Dimethylphthalate	13.57	163	436617	9.474	ng	# 96
71) Phenanthrene	16.92	178	411102	6.039	ng	# 20
75) Fluoranthene	18.94	202	861087	11.033	ng	93
78) Pyrene	19.30	202	669891	10.967	ng	# 93
81) Benzo(a)anthracene	21.04	228	233127	4.201	ng	98
83) Chrysene	21.09	228	366883	6.938	ng	99
84) Bis(2-ethylhexyl)phthalate	21.00	149	638505	18.128	ng	97
87) Indeno(1,2,3-cd)pyrene	25.21	276	228788	4.368	ng	96
88) Benzo(b)fluoranthene	22.54	252	503203	9.238	ng	# 95
89) Benzo(k)fluoranthene	22.57	252	134272m	2.537	ng	
90) Benzo(a)pyrene	23.06	252	269626	5.573	ng	96
92) Benzo(a,h,i)perylene	25.83	276	233892	5.621	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM060820\
Data File : BM026291.D
Acq On : 08 Jun 2020 17:44
Operator : JU/CG
Sample : L2890-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CHRT-26160

Manual Integrations
APPROVED

mohammad
6/9/2020 12:21:46 PM

Quant Time: Jun 08 18:46:57 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 15:00:18 2020
Response via : Initial Calibration

