

Data Path : Z:\HPCHEM1\BNA M\DATA\BM060916\
 Data File : BM005716.D
 Acq On : 09 Jun 2016 16:59
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02051

Quant Time: Jun 10 01:25:31 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM060216.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 10 01:23:18 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2 S	1,4-Dioxane-d8	0.476	0.444	6.7	90	0.00
3 S	Phenol-d5	1.758	1.707	2.9	90	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.070	1.046	2.2	89	0.00
5 S	2-Chlorophenol-d4	1.451	1.422	2.0	93	0.00
6 S	4-Methylphenol-d8	1.391	1.425	-2.4	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.154	0.148	3.9	99	0.00
9 S	2-Nitrophenol-d4	0.164	0.165	-0.6	102	0.00
10 S	2,4-Dichlorophenol-d3	0.292	0.299	-2.4	106	0.00
11	Naphthalene	1.017	0.991	2.6	101	0.00
12 S	4-Chloroaniline-d4	0.347	0.417	-20.2#	111	0.00
13	2-Methylnaphthalene	0.707	0.736	-4.1	107	0.00
14	1-Methylnaphthalene	0.695	0.728	-4.7	108	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
16 S	Dimethylphthalate-d6	1.538	1.582	-2.9	125	0.00
17 S	Acenaphthylene-d8	2.000	1.992	0.4	119	0.00
18	Acenaphthylene	2.069	2.046	1.1	118	0.00
19 C	Acenaphthene	1.343	1.326	1.3	119	0.00
20 S	4-Nitrophenol-d4	0.220	0.224	-1.8	125	0.00
21 S	Fluorene-d10	1.321	1.373	-3.9	125	0.00
22	Fluorene	1.453	1.540	-6.0	125	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.094	0.093	1.1	149	0.00
25	Phenanthrene	1.104	1.089	1.4	138	0.00
26 S	Anthracene-d10	0.950	0.952	-0.2	139	0.00
27	Anthracene	1.122	1.136	-1.2	140	0.00
28 C	Fluoranthene	1.128	1.261	-11.8	154#	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	156#	0.00
30 S	Pyrene-d10	1.005	0.953	5.2	149	0.00
31	Pyrene	1.262	1.219	3.4	152#	0.00
32	Benzo(a)anthracene	1.173	1.161	1.0	154#	0.00
33	Chrysene	1.108	1.111	-0.3	158#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	152#	0.00
35	Benzo(b)fluoranthene	1.206	1.201	0.4	155#	0.00
36	Benzo(k)fluoranthene	1.132	1.182	-4.4	158#	0.00
37 S	Benzo(a)pyrene-d12	0.920	0.918	0.2	153#	0.00
38 C	Benzo(a)pyrene	1.161	1.139	1.9	149	0.00
39	Indeno(1,2,3-cd)pyrene	1.374	1.117	18.7	124	0.00
40	Dibenzo(a,h)anthracene	1.153	0.939	18.6	123	0.00
41	Benzo(g,h,i)perylene	1.152	0.914	20.7#	123	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM060916\
Data File : BM005716.D
Acq On : 09 Jun 2016 16:59
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02051

Quant Time: Jun 10 01:25:31 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM060216.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 10 01:23:18 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0