


Data Path : Z:\HPCHEM1\BNA M\DATA\BM060916\  
 Data File : BM005706.D  
 Acq On : 09 Jun 2016 09:30  
 Operator : UM/SJ  
 Sample : DFTPP51  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP51

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM041416.M  
 Title : SVOA CALIBRATION  
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2547, 2548, 2549; Background Corrected with Scan 2542

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 36.9         | 38669      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 703        | PASS                |
| 69             | 198             | 0.00            | 100             | 36.2         | 37931      | PASS                |
| 70             | 69              | 0.00            | 2               | 1.0          | 376        | PASS                |
| 127            | 198             | 10              | 80              | 48.8         | 51242      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.6          | 584        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 104898     | PASS                |
| 199            | 198             | 5               | 9               | 7.0          | 7300       | PASS                |
| 275            | 198             | 10              | 60              | 25.7         | 27008      | PASS                |
| 365            | 198             | 1               | 100             | 3.3          | 3464       | PASS                |
| 441            | 443             | 0.01            | 100             | 79.9         | 12855      | PASS                |
| 442            | 198             | 50              | 100             | 80.4         | 84341      | PASS                |
| 443            | 442             | 15              | 24              | 19.1         | 16080      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 39.00  | 398.0     |
| 39.90  | 353.0     |
| 41.00  | 824.0     |
| 42.00  | 386.0     |
| 43.00  | 1339.0    |
| 44.00  | 786.0     |
| 44.90  | 311.0     |
| 55.10  | 951.0     |
| 57.00  | 673.0     |
| 60.00  | 455.0     |
| 67.00  | 306.0     |
| 69.00  | 541.0     |
| 73.00  | 874.0     |
| 83.00  | 406.0     |
| 84.00  | 350.0     |
| 97.00  | 314.0     |
| 207.00 | 489.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 475.0     |
| 39.00  | 2951.0    |
| 40.00  | 441.0     |
| 41.00  | 851.0     |
| 43.00  | 1082.0    |
| 44.00  | 911.0     |
| 45.00  | 337.0     |
| 48.90  | 309.0     |
| 50.00  | 9171.0    |
| 51.00  | 37120.0   |
| 52.00  | 2143.0    |
| 53.00  | 342.0     |
| 55.00  | 1025.0    |
| 56.00  | 1153.0    |
| 57.00  | 2930.0    |
| 60.10  | 422.0     |
| 61.00  | 491.0     |
| 62.00  | 511.0     |
| 63.00  | 1532.0    |
| 65.00  | 763.0     |
| 68.00  | 578.0     |
| 69.00  | 36664.0   |
| 70.00  | 359.0     |
| 71.00  | 313.0     |
| 73.00  | 929.0     |
| 74.00  | 3731.0    |
| 75.00  | 5845.0    |
| 76.00  | 2229.0    |
| 77.00  | 46344.0   |
| 78.00  | 3335.0    |
| 79.00  | 2800.0    |
| 80.00  | 2211.0    |
| 81.00  | 3208.0    |
| 82.00  | 897.0     |
| 83.00  | 955.0     |
| 84.00  | 439.0     |
| 85.00  | 801.0     |
| 86.00  | 896.0     |
| 87.00  | 592.0     |
| 90.90  | 808.0     |
| 91.90  | 685.0     |
| 93.00  | 4461.0    |
| 94.00  | 361.0     |
| 95.00  | 380.0     |
| 95.90  | 382.0     |
| 98.00  | 3380.0    |
| 99.00  | 3005.0    |
| 101.00 | 1873.0    |
| 102.90 | 448.0     |
| 104.00 | 1004.0    |
| 105.00 | 1036.0    |
| 106.00 | 347.0     |
| 107.00 | 13697.0   |
| 108.00 | 2352.0    |
| 109.00 | 467.0     |
| 110.00 | 24064.0   |
| 111.00 | 3766.0    |
| 111.90 | 420.0     |
| 115.90 | 688.0     |
| 117.00 | 10286.0   |
| 118.00 | 804.0     |
| 122.00 | 936.0     |
| 123.00 | 1250.0    |
| 124.00 | 659.0     |
| 125.00 | 620.0     |
| 127.00 | 46496.0   |
| 128.00 | 3585.0    |
| 129.00 | 17360.0   |
| 129.90 | 1546.0    |
| 130.90 | 358.0     |
| 134.00 | 523.0     |
| 135.00 | 1506.0    |
| 135.90 | 695.0     |
| 137.00 | 768.0     |
| 140.90 | 2108.0    |
| 141.90 | 702.0     |
| 143.00 | 577.0     |
| 146.00 | 454.0     |
| 147.00 | 1093.0    |
| 147.90 | 2575.0    |
| 149.00 | 619.0     |
| 153.00 | 747.0     |
| 154.00 | 528.0     |
| 155.00 | 1231.0    |
| 156.00 | 1919.0    |
| 157.00 | 439.0     |
| 157.90 | 351.0     |
| 159.00 | 403.0     |
| 160.00 | 725.0     |
| 161.00 | 1177.0    |
| 165.00 | 894.0     |
| 166.00 | 664.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |         |
|--------|---------|
| 167.00 | 4809.0  |
| 168.00 | 2718.0  |
| 168.90 | 427.0   |
| 171.90 | 400.0   |
| 173.00 | 501.0   |
| 174.00 | 946.0   |
| 175.00 | 2003.0  |
| 176.00 | 505.0   |
| 177.00 | 760.0   |
| 178.90 | 3397.0  |
| 180.00 | 2396.0  |
| 181.00 | 1127.0  |
| 183.90 | 336.0   |
| 185.00 | 1693.0  |
| 186.00 | 12470.0 |
| 187.00 | 3574.0  |
| 187.90 | 391.0   |
| 189.00 | 722.0   |
| 190.90 | 449.0   |
| 192.00 | 986.0   |
| 193.00 | 1312.0  |
| 194.00 | 308.0   |
| 196.00 | 2513.0  |
| 196.80 | 540.0   |
| 197.90 | 87608.0 |
| 199.00 | 6022.0  |
| 200.00 | 469.0   |
| 201.40 | 411.0   |
| 203.00 | 554.0   |
| 204.00 | 3242.0  |
| 205.00 | 5676.0  |
| 206.00 | 24144.0 |
| 207.00 | 3456.0  |
| 208.00 | 991.0   |
| 208.90 | 349.0   |
| 210.10 | 465.0   |
| 211.00 | 868.0   |
| 216.00 | 506.0   |
| 217.00 | 6258.0  |
| 218.00 | 854.0   |
| 221.00 | 3606.0  |
| 221.80 | 848.0   |
| 222.90 | 1369.0  |
| 224.00 | 12532.0 |
| 225.00 | 3158.0  |
| 226.00 | 339.0   |
| 227.00 | 5223.0  |
| 227.90 | 752.0   |
| 229.00 | 1157.0  |
| 231.00 | 506.0   |
| 233.90 | 368.0   |
| 235.00 | 409.0   |
| 236.90 | 382.0   |
| 241.00 | 370.0   |
| 242.00 | 805.0   |
| 243.00 | 739.0   |
| 244.00 | 9774.0  |
| 245.00 | 1482.0  |
| 246.00 | 1937.0  |
| 247.00 | 407.0   |
| 248.90 | 313.0   |
| 254.00 | 371.0   |
| 255.00 | 49624.0 |
| 256.00 | 7534.0  |
| 257.00 | 545.0   |
| 258.00 | 2751.0  |
| 259.00 | 487.0   |
| 265.00 | 1091.0  |
| 273.00 | 1391.0  |
| 274.00 | 3984.0  |
| 275.00 | 20680.0 |
| 276.00 | 2790.0  |
| 277.00 | 1844.0  |
| 277.90 | 301.0   |
| 292.90 | 365.0   |
| 296.00 | 6146.0  |
| 297.00 | 743.0   |
| 303.00 | 593.0   |
| 313.90 | 334.0   |
| 314.90 | 704.0   |
| 315.90 | 327.0   |
| 323.00 | 2013.0  |
| 324.00 | 325.0   |
| 326.90 | 397.0   |
| 334.00 | 1172.0  |
| 346.00 | 445.0   |
| 352.00 | 500.0   |
| 353.00 | 439.0   |
| 354.00 | 557.0   |
| 365.00 | 2421.0  |
| 366.00 | 369.0   |
| 372.00 | 798.0   |
| 401.90 | 377.0   |
| 403.00 | 493.0   |
| 421.00 | 366.0   |
| 422.00 | 396.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |         |
|--------|---------|
| 423.00 | 2828.0  |
| 424.00 | 731.0   |
| 441.00 | 8366.0  |
| 442.00 | 53808.0 |
| 443.00 | 10578.0 |
| 444.00 | 1112.0  |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP51

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 520.0     |
| 39.00  | 3300.0    |
| 39.90  | 449.0     |
| 41.00  | 950.0     |
| 43.00  | 1100.0    |
| 44.00  | 979.0     |
| 45.00  | 406.0     |
| 48.90  | 393.0     |
| 50.00  | 10940.0   |
| 51.00  | 44688.0   |
| 52.00  | 2308.0    |
| 55.00  | 1043.0    |
| 56.00  | 1445.0    |
| 57.00  | 3430.0    |
| 60.00  | 414.0     |
| 60.90  | 576.0     |
| 62.00  | 626.0     |
| 63.00  | 1832.0    |
| 65.00  | 842.0     |
| 67.00  | 319.0     |
| 68.00  | 812.0     |
| 69.00  | 44480.0   |
| 70.00  | 445.0     |
| 73.00  | 1008.0    |
| 74.00  | 4882.0    |
| 75.00  | 7422.0    |
| 76.00  | 2552.0    |
| 77.00  | 58144.0   |
| 78.00  | 4286.0    |
| 79.00  | 3362.0    |
| 80.00  | 2833.0    |
| 81.00  | 4144.0    |
| 82.00  | 1092.0    |
| 83.00  | 1370.0    |
| 84.00  | 333.0     |
| 85.00  | 1040.0    |
| 86.00  | 1069.0    |
| 87.00  | 637.0     |
| 91.00  | 867.0     |
| 92.00  | 900.0     |
| 93.00  | 5799.0    |
| 93.90  | 387.0     |
| 95.00  | 367.0     |
| 96.00  | 557.0     |
| 97.00  | 339.0     |
| 98.00  | 4607.0    |
| 99.00  | 3835.0    |
| 99.90  | 327.0     |
| 101.00 | 2140.0    |
| 103.00 | 737.0     |
| 104.00 | 1221.0    |
| 105.00 | 1377.0    |
| 106.00 | 409.0     |
| 107.00 | 17136.0   |
| 108.00 | 3052.0    |
| 109.00 | 508.0     |
| 110.00 | 30312.0   |
| 111.00 | 4642.0    |
| 112.00 | 570.0     |
| 116.00 | 984.0     |
| 117.00 | 13575.0   |
| 118.00 | 962.0     |
| 121.90 | 963.0     |
| 123.00 | 1829.0    |
| 123.90 | 739.0     |
| 125.00 | 755.0     |
| 126.10 | 308.0     |
| 127.00 | 60048.0   |
| 128.00 | 4668.0    |
| 129.00 | 22552.0   |
| 130.00 | 1925.0    |
| 131.00 | 513.0     |
| 134.00 | 607.0     |
| 135.00 | 1729.0    |
| 135.90 | 713.0     |
| 137.00 | 1099.0    |
| 139.90 | 339.0     |
| 141.00 | 2981.0    |
| 141.90 | 1042.0    |
| 143.00 | 740.0     |
| 146.00 | 605.0     |
| 147.00 | 1672.0    |
| 148.00 | 3568.0    |
| 149.00 | 934.0     |
| 150.90 | 444.0     |
| 151.80 | 324.0     |
| 152.90 | 940.0     |
| 154.00 | 671.0     |
| 155.00 | 1588.0    |
| 156.00 | 2660.0    |
| 157.00 | 541.0     |
| 158.00 | 508.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |          |
|--------|----------|
| 158.90 | 374.0    |
| 160.00 | 969.0    |
| 161.00 | 1447.0   |
| 161.90 | 384.0    |
| 164.90 | 1142.0   |
| 166.00 | 789.0    |
| 167.00 | 6633.0   |
| 168.00 | 3565.0   |
| 169.00 | 594.0    |
| 170.90 | 366.0    |
| 171.90 | 624.0    |
| 172.90 | 669.0    |
| 174.00 | 1390.0   |
| 175.00 | 2594.0   |
| 176.00 | 689.0    |
| 177.00 | 1081.0   |
| 177.90 | 387.0    |
| 179.00 | 4220.0   |
| 180.00 | 3174.0   |
| 181.00 | 1422.0   |
| 182.00 | 333.0    |
| 183.90 | 333.0    |
| 185.00 | 2308.0   |
| 186.00 | 17592.0  |
| 187.00 | 5319.0   |
| 188.10 | 533.0    |
| 188.90 | 1139.0   |
| 191.00 | 609.0    |
| 192.00 | 1486.0   |
| 193.00 | 1819.0   |
| 194.00 | 458.0    |
| 196.00 | 3681.0   |
| 196.90 | 569.0    |
| 198.00 | 122632.0 |
| 199.00 | 8876.0   |
| 199.90 | 654.0    |
| 201.50 | 532.0    |
| 203.00 | 813.0    |
| 204.00 | 4661.0   |
| 205.00 | 8098.0   |
| 206.00 | 33384.0  |
| 207.10 | 4811.0   |
| 208.00 | 1162.0   |
| 209.00 | 479.0    |
| 209.90 | 452.0    |
| 211.00 | 1311.0   |
| 214.90 | 341.0    |
| 216.00 | 718.0    |
| 217.00 | 8729.0   |
| 218.00 | 1042.0   |
| 221.00 | 4464.0   |
| 222.00 | 1097.0   |
| 223.00 | 2095.0   |
| 224.00 | 18656.0  |
| 225.00 | 4880.0   |
| 226.00 | 484.0    |
| 227.00 | 7643.0   |
| 228.00 | 915.0    |
| 229.00 | 1589.0   |
| 231.00 | 568.0    |
| 234.00 | 447.0    |
| 235.00 | 450.0    |
| 235.90 | 412.0    |
| 237.00 | 696.0    |
| 241.00 | 458.0    |
| 241.90 | 998.0    |
| 243.00 | 1229.0   |
| 244.10 | 13777.0  |
| 245.00 | 1800.0   |
| 246.00 | 2554.0   |
| 247.00 | 572.0    |
| 249.00 | 482.0    |
| 252.90 | 340.0    |
| 254.10 | 523.0    |
| 255.00 | 73664.0  |
| 256.00 | 10252.0  |
| 257.00 | 884.0    |
| 258.00 | 4496.0   |
| 258.90 | 590.0    |
| 265.00 | 1687.0   |
| 273.00 | 2104.0   |
| 274.00 | 5758.0   |
| 275.00 | 31968.0  |
| 276.00 | 4190.0   |
| 277.00 | 2470.0   |
| 278.00 | 440.0    |
| 283.00 | 345.0    |
| 285.00 | 441.0    |
| 292.90 | 777.0    |
| 296.00 | 9459.0   |
| 297.00 | 1450.0   |
| 303.00 | 945.0    |
| 314.00 | 446.0    |
| 315.00 | 1037.0   |
| 315.90 | 509.0    |
| 323.00 | 2834.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |         |
|--------|---------|
| 324.00 | 455.0   |
| 326.90 | 487.0   |
| 327.90 | 301.0   |
| 334.00 | 1830.0  |
| 335.00 | 523.0   |
| 341.10 | 329.0   |
| 346.00 | 672.0   |
| 352.00 | 921.0   |
| 353.00 | 550.0   |
| 354.00 | 927.0   |
| 365.00 | 3966.0  |
| 366.00 | 644.0   |
| 372.00 | 1435.0  |
| 373.00 | 315.0   |
| 382.90 | 406.0   |
| 402.00 | 546.0   |
| 403.00 | 904.0   |
| 421.00 | 733.0   |
| 422.00 | 575.0   |
| 423.00 | 5203.0  |
| 424.00 | 951.0   |
| 441.10 | 14916.0 |
| 442.00 | 98384.0 |
| 443.00 | 18392.0 |
| 444.00 | 1878.0  |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP51

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 325.0     |
| 39.00  | 2584.0    |
| 40.00  | 468.0     |
| 41.00  | 828.0     |
| 43.00  | 1223.0    |
| 44.00  | 917.0     |
| 45.00  | 414.0     |
| 49.00  | 370.0     |
| 50.00  | 8618.0    |
| 51.00  | 34200.0   |
| 52.00  | 1760.0    |
| 55.00  | 936.0     |
| 56.00  | 1161.0    |
| 57.00  | 3043.0    |
| 60.00  | 454.0     |
| 61.00  | 479.0     |
| 61.90  | 493.0     |
| 63.00  | 1569.0    |
| 65.00  | 706.0     |
| 67.00  | 396.0     |
| 68.00  | 720.0     |
| 69.00  | 34272.0   |
| 70.00  | 326.0     |
| 71.00  | 357.0     |
| 73.00  | 871.0     |
| 74.00  | 3544.0    |
| 75.00  | 5531.0    |
| 76.00  | 1933.0    |
| 77.00  | 43856.0   |
| 78.00  | 3041.0    |
| 79.00  | 2741.0    |
| 80.00  | 1891.0    |
| 81.00  | 3254.0    |
| 82.00  | 827.0     |
| 83.00  | 1017.0    |
| 85.00  | 732.0     |
| 86.00  | 871.0     |
| 87.00  | 505.0     |
| 91.00  | 769.0     |
| 92.00  | 715.0     |
| 93.00  | 4589.0    |
| 94.00  | 357.0     |
| 95.00  | 325.0     |
| 98.00  | 3519.0    |
| 98.90  | 2867.0    |
| 101.00 | 1816.0    |
| 103.00 | 451.0     |
| 104.00 | 1030.0    |
| 105.00 | 1065.0    |
| 105.90 | 406.0     |
| 107.00 | 13796.0   |
| 108.00 | 2271.0    |
| 109.00 | 424.0     |
| 110.00 | 23536.0   |
| 111.00 | 3705.0    |
| 112.00 | 575.0     |
| 115.90 | 628.0     |
| 117.00 | 10629.0   |
| 118.00 | 906.0     |
| 122.00 | 997.0     |
| 123.00 | 1587.0    |
| 124.00 | 641.0     |
| 124.90 | 596.0     |
| 127.00 | 47184.0   |
| 128.00 | 3564.0    |
| 129.00 | 17920.0   |
| 130.00 | 1472.0    |
| 131.00 | 480.0     |
| 134.00 | 557.0     |
| 135.00 | 1624.0    |
| 136.00 | 661.0     |
| 137.00 | 1065.0    |
| 140.90 | 2345.0    |
| 142.00 | 712.0     |
| 143.00 | 514.0     |
| 145.90 | 349.0     |
| 146.90 | 1325.0    |
| 148.00 | 2745.0    |
| 148.90 | 678.0     |
| 151.00 | 364.0     |
| 153.00 | 816.0     |
| 153.90 | 617.0     |
| 155.00 | 1433.0    |
| 156.00 | 2252.0    |
| 157.00 | 365.0     |
| 157.90 | 412.0     |
| 159.00 | 385.0     |
| 159.90 | 755.0     |
| 161.00 | 1290.0    |
| 162.00 | 345.0     |
| 164.90 | 852.0     |
| 166.00 | 723.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |          |
|--------|----------|
| 167.00 | 5238.0   |
| 168.00 | 2719.0   |
| 168.90 | 435.0    |
| 172.00 | 463.0    |
| 173.00 | 590.0    |
| 174.00 | 1088.0   |
| 175.00 | 2141.0   |
| 176.00 | 516.0    |
| 176.90 | 887.0    |
| 178.90 | 3603.0   |
| 180.00 | 2554.0   |
| 181.00 | 1207.0   |
| 181.90 | 306.0    |
| 184.00 | 306.0    |
| 185.00 | 1994.0   |
| 186.00 | 14696.0  |
| 187.00 | 4498.0   |
| 188.00 | 484.0    |
| 188.90 | 794.0    |
| 190.90 | 478.0    |
| 192.00 | 1315.0   |
| 193.00 | 1406.0   |
| 193.90 | 336.0    |
| 196.00 | 2724.0   |
| 196.70 | 645.0    |
| 197.90 | 104456.0 |
| 198.90 | 7003.0   |
| 200.00 | 586.0    |
| 201.40 | 449.0    |
| 203.00 | 714.0    |
| 204.00 | 3936.0   |
| 205.00 | 6762.0   |
| 206.00 | 28208.0  |
| 207.00 | 4374.0   |
| 208.00 | 1098.0   |
| 209.00 | 325.0    |
| 211.00 | 1145.0   |
| 215.00 | 364.0    |
| 215.90 | 516.0    |
| 217.00 | 7408.0   |
| 218.00 | 1009.0   |
| 221.00 | 4282.0   |
| 221.70 | 991.0    |
| 223.00 | 1702.0   |
| 224.00 | 15813.0  |
| 225.00 | 4163.0   |
| 226.00 | 456.0    |
| 227.00 | 6659.0   |
| 228.00 | 981.0    |
| 229.00 | 1435.0   |
| 231.00 | 528.0    |
| 233.90 | 443.0    |
| 235.00 | 443.0    |
| 236.00 | 336.0    |
| 237.00 | 584.0    |
| 240.90 | 422.0    |
| 241.90 | 879.0    |
| 243.00 | 878.0    |
| 244.00 | 12830.0  |
| 245.00 | 1834.0   |
| 246.00 | 2559.0   |
| 247.00 | 552.0    |
| 248.90 | 509.0    |
| 254.00 | 440.0    |
| 255.00 | 65592.0  |
| 256.00 | 9196.0   |
| 257.00 | 706.0    |
| 258.00 | 4006.0   |
| 259.00 | 589.0    |
| 265.00 | 1623.0   |
| 273.00 | 2071.0   |
| 274.00 | 5521.0   |
| 275.00 | 28376.0  |
| 276.00 | 3832.0   |
| 277.00 | 2318.0   |
| 277.90 | 454.0    |
| 285.00 | 475.0    |
| 293.00 | 475.0    |
| 296.00 | 8591.0   |
| 296.90 | 1086.0   |
| 303.00 | 929.0    |
| 314.00 | 420.0    |
| 314.90 | 835.0    |
| 316.00 | 522.0    |
| 323.00 | 3039.0   |
| 324.00 | 427.0    |
| 326.90 | 494.0    |
| 334.00 | 1712.0   |
| 335.00 | 404.0    |
| 341.00 | 442.0    |
| 345.90 | 674.0    |
| 352.00 | 795.0    |
| 353.00 | 613.0    |
| 354.00 | 860.0    |
| 365.00 | 4007.0   |
| 365.90 | 519.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP51

|        |          |
|--------|----------|
| 372.00 | 1555.0   |
| 373.00 | 359.0    |
| 383.00 | 405.0    |
| 402.00 | 666.0    |
| 403.00 | 836.0    |
| 421.00 | 731.0    |
| 422.00 | 678.0    |
| 423.00 | 5213.0   |
| 424.00 | 1022.0   |
| 441.00 | 15284.0  |
| 442.10 | 100832.0 |
| 443.00 | 19272.0  |
| 444.00 | 1881.0   |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP51