

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060920\
 Data File : BM026301.D
 Acq On : 09 Jun 2020 13:27
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 14:48:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 09 14:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2	1,4-Dioxane	0.510	0.482	5.5	58	0.00
3	Pyridine	1.498	1.403	6.3	58	0.00
4	n-Nitrosodimethylamine	0.742	0.628	15.4	51	0.00
5 S	2-Fluorophenol	1.131	1.093	3.4	60	0.00
6	Aniline	2.149	1.972	8.2	57	0.00
7 S	Phenol-d6	1.638	1.510	7.8	57	0.00
8	2-Chlorophenol	1.305	1.235	5.4	59	0.00
9	Benzaldehyde	1.044	1.016	2.7	78	0.00
10 C	Phenol	1.744	1.617	7.3	58	0.00
11	bis(2-Chloroethyl)ether	1.405	1.287	8.4	57	0.00
12	1,3-Dichlorobenzene	1.445	1.402	3.0	60	0.00
13 C	1,4-Dichlorobenzene	1.471	1.424	3.2	60	0.00
14	1,2-Dichlorobenzene	1.437	1.375	4.3	59	0.00
15	Benzyl Alcohol	1.391	1.258	9.6	56	0.00
16	2,2'-oxybis(1-Chloropropane	2.624	2.247	14.4	52	0.00
17	2-Methylphenol	1.275	1.180	7.5	58	0.00
18	Hexachloroethane	0.557	0.535	3.9	59	0.00
19 P	n-Nitroso-di-n-propylamine	1.276	1.106	13.3	53	0.00
20	3+4-Methylphenols	1.738	1.582	9.0	56	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
22	Acetophenone	0.514	0.493	4.1	55	0.00
23 S	Nitrobenzene-d5	0.407	0.384	5.7	54	0.00
24	Nitrobenzene	0.426	0.401	5.9	55	0.00
25	Isophorone	0.765	0.724	5.4	55	0.00
26 C	2-Nitrophenol	0.169	0.167	1.2	57	0.00
27	2,4-Dimethylphenol	0.266	0.258	3.0	56	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.408	6.0	54	0.00
29 C	2,4-Dichlorophenol	0.292	0.283	3.1	56	0.00
30	1,2,4-Trichlorobenzene	0.321	0.315	1.9	56	0.00
31	Naphthalene	1.008	0.974	3.4	56	0.00
32	Benzoic acid	0.198	0.151	23.7	42#	0.00
33	4-Chloroaniline	0.440	0.410	6.8	54	0.00
34 C	Hexachlorobutadiene	0.202	0.202	0.0	57	0.00
35	Caprolactam	0.103	0.100	2.9	57	0.00
36 C	4-Chloro-3-methylphenol	0.356	0.328	7.9	54	0.00
37	2-Methylnaphthalene	0.718	0.672	6.4	55	0.00
38	1-Methylnaphthalene	0.680	0.635	6.6	54	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	58	0.00
40	1,2,4,5-Tetrachlorobenzene	0.569	0.577	-1.4	55	0.00
41 P	Hexachlorocyclopentadiene	0.335	0.265	20.9	43#	0.00
42 S	2,4,6-Tribromophenol	0.242	0.240	0.8	56	0.00
43 C	2,4,6-Trichlorophenol	0.385	0.390	-1.3	55	0.00
44	2,4,5-Trichlorophenol	0.447	0.444	0.7	54	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060920\
 Data File : BM026301.D
 Acq On : 09 Jun 2020 13:27
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 14:48:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 09 14:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.318	1.320	-0.2	55	0.00
46	1,1'-Biphenyl	1.402	1.373	2.1	54	0.00
47	2-Chloronaphthalene	1.139	1.115	2.1	54	0.00
48	2-Nitroaniline	0.451	0.424	6.0	51	0.00
49	Acenaphthylene	1.822	1.790	1.8	54	0.00
50	Dimethylphthalate	1.463	1.419	3.0	54	0.00
51	2,6-Dinitrotoluene	0.321	0.315	1.9	54	0.00
52 C	Acenaphthene	1.094	1.067	2.5	53	0.00
53	3-Nitroaniline	0.357	0.343	3.9	53	0.00
54 P	2,4-Dinitrophenol	0.195	0.164	15.9	47#	0.00
55	Dibenzofuran	1.762	1.706	3.2	53	0.00
56 P	4-Nitrophenol	0.283	0.270	4.6	53	0.00
57	2,4-Dinitrotoluene	0.446	0.432	3.1	54	0.00
58	Fluorene	1.431	1.376	3.8	53	0.00
59	2,3,4,6-Tetrachlorophenol	0.383	0.373	2.6	54	0.00
60	Diethylphthalate	1.490	1.447	2.9	54	0.00
61	4-Chlorophenyl-phenylether	0.759	0.738	2.8	54	0.00
62	4-Nitroaniline	0.381	0.366	3.9	54	0.00
63	Azobenzene	1.630	1.535	5.8	52	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.110	9.8	51	0.00
66 c	n-Nitrosodiphenylamine	0.579	0.556	4.0	54	0.00
67	4-Bromophenyl-phenylether	0.220	0.214	2.7	54	0.00
68	Hexachlorobenzene	0.230	0.224	2.6	55	0.00
69	Atrazine	0.173	0.180	-4.0	55	0.00
70 C	Pentachlorophenol	0.138	0.127	8.0	51	0.00
71	Phenanthrene	1.042	1.002	3.8	55	0.00
72	Anthracene	1.039	1.004	3.4	55	0.00
73	Carbazole	0.986	0.968	1.8	57	0.00
74	Di-n-butylphthalate	1.162	1.175	-1.1	59	0.00
75 C	Fluoranthene	1.194	1.208	-1.2	60	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
77	Benzidine	0.415	0.405	2.4	66	0.00
78	Pyrene	1.321	1.179	10.7	61	0.00
79 S	Terphenyl-d14	1.030	0.935	9.2	63	0.00
80	Butylbenzylphthalate	0.534	0.528	1.1	72	0.00
81	Benzo(a)anthracene	1.200	1.168	2.7	71	0.00
82	3,3'-Dichlorobenzidine	0.371	0.431	-16.2	84	0.00
83	Chrysene	1.144	1.123	1.8	71	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.780	-2.4	77	0.00
85 c	Di-n-octyl phthalate	1.175	1.341	-14.1	88	0.00
86 I	Perylene-d12	1.000	1.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	1.157	1.279	-10.5	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM060920\
Data File : BM026301.D
Acq On : 09 Jun 2020 13:27
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jun 09 14:48:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM060520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 09 14:46:47 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.203	1.116	7.2	83	0.00
89	Benzo(k)fluoranthene	1.169	1.042	10.9	80	0.00
90 C	Benzo(a)pyrene	1.069	1.027	3.9	84	0.00
91	Dibenzo(a,h)anthracene	0.966	1.067	-10.5	91	0.00
92	Benzo(q,h,i)perylene	0.919	1.032	-12.3	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0