




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050235.D
Acq On : 09 Jun 2025 10:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 09 13:03:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050235.D\data.ms

(70) Pentachlorophenol (C)




16.786min (-0.000) 0.00 ng

response 616214

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	63.75
264.00	63.50	62.57
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM060925\
Data File : BM050235.D
Acq On : 09 Jun 2025 10:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 09 13:03:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 09 11:54:38 2025
Response via : Initial Calibration



TIC: BM050235.D\data.ms

(77) Benzidine**19.374min (-0.000) 298883.97 ng****response 3092988**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	14.66
183.00	11.20	11.14
0.00	0.00	0.00