

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061021\
 Data File : BM030385.D
 Acq On : 11 Jun 2021 00:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Quant Time: Jun 11 07:53:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 08 20:51:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	139971	20.00	ng	-0.01
21) Naphthalene-d8	10.634	136	607488	20.00	ng	0.00
39) Acenaphthene-d10	14.463	164	449136	20.00	ng	-0.01
64) Phenanthrene-d10	17.198	188	1075503	20.00	ng	-0.01
76) Chrysene-d12	21.362	240	1495352	20.00	ng	0.00
86) Perylene-d12	23.639	264	1692151	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	659478	76.59	ng	0.00
7) Phenol-d6	7.005	99	995307	79.85	ng	0.00
23) Nitrobenzene-d5	8.999	82	919044	73.70	ng	-0.01
42) 2,4,6-Tribromophenol	15.951	330	168595	29.86	ng	0.00
45) 2-Fluorobiphenyl	13.086	172	2450225	71.74	ng	-0.01
79) Terphenyl-d14	19.815	244	5325052	62.55	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.346	88	127451	34.24	ng	93
3) Pyridine	3.752	79	366312	38.55	ng	88
4) n-Nitrosodimethylamine	3.652	42	167077	34.64	ng	# 85
6) Aniline	7.175	93	527378	37.08	ng	95
8) 2-Chlorophenol	7.405	128	399216	39.50	ng	98
9) Benzaldehyde	6.987	77	243483	46.77	ng	92
10) Phenol	7.028	94	492010	39.07	ng	92
11) bis(2-Chloroethyl)ether	7.269	93	365036	36.79	ng	99
12) 1,3-Dichlorobenzene	7.734	146	411143	36.93	ng	97
13) 1,4-Dichlorobenzene	7.881	146	419059	36.93	ng	99
14) 1,2-Dichlorobenzene	8.193	146	411816	37.31	ng	100
15) Benzyl Alcohol	8.081	79	343969	39.32	ng	93
16) 2,2'-oxybis(1-Chloropr...	8.369	45	514435	32.22	ng	92
17) 2-Methylphenol	8.275	107	325185	39.85	ng	94
18) Hexachloroethane	8.922	117	141764	34.76	ng	99
19) n-Nitroso-di-n-propyla...	8.646	70	310344	37.50	ng	98
20) 3+4-Methylphenols	8.604	107	466885	42.05	ng	96
22) Acetophenone	8.663	105	615601	37.93	ng	# 99
24) Nitrobenzene	9.040	77	463126	35.60	ng	94
25) Isophorone	9.563	82	885185	37.04	ng	99
26) 2-Nitrophenol	9.746	139	88115	17.66	ng	91
27) 2,4-Dimethylphenol	9.804	122	348738	37.70	ng	96
28) bis(2-Chloroethoxy)met...	10.040	93	468888	34.51	ng	100
29) 2,4-Dichlorophenol	10.275	162	414020	39.92	ng	99
30) 1,2,4-Trichlorobenzene	10.493	180	445887	37.67	ng	99
31) Naphthalene	10.681	128	1288084	36.40	ng	99
32) Benzoic acid	9.804	122	348738	48.49	ng	# 10
33) 4-Chloroaniline	10.793	127	554544	38.51	ng	95
34) Hexachlorobutadiene	10.969	225	265661	37.60	ng	96
35) Caprolactam	11.569	113	174045	48.63	ng	95
36) 4-Chloro-3-methylphenol	11.910	107	446749	41.13	ng	99
37) 2-Methylnaphthalene	12.287	142	997751	38.43	ng	98
38) 1-Methylnaphthalene	12.510	142	940829	38.04	ng	100
40) 1,2,4,5-Tetrachloroben...	12.657	216	544100	37.58	ng	98
41) Hexachlorocyclopentadiene	12.639	237	176256	22.20	ng	99
43) 2,4,6-Trichlorophenol	12.898	196	214146	21.88	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061021\
 Data File : BM030385.D
 Acq On : 11 Jun 2021 00:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Quant Time: Jun 11 07:53:57 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 08 20:51:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.969	196	334672	31.17	ng	97
46) 1,1'-Biphenyl	13.298	154	1299810	35.74	ng	98
47) 2-Chloronaphthalene	13.339	162	1050911	35.75	ng	98
48) 2-Nitroaniline	13.545	65	319240	36.67	ng	96
49) Acenaphthylene	14.186	152	1645283	36.53	ng	99
50) Dimethylphthalate	13.922	163	1349218	37.75	ng	99
51) 2,6-Dinitrotoluene	14.039	165	300269	38.11	ng	93
52) Acenaphthene	14.528	154	1021566	35.51	ng	99
53) 3-Nitroaniline	14.369	138	325553	37.93	ng	94
55) Dibenzofuran	14.863	168	1713264	37.43	ng	98
56) 4-Nitrophenol	14.675	139	24841	3.71	ng	96
57) 2,4-Dinitrotoluene	14.822	165	365393	33.75	ng	97
58) Fluorene	15.510	166	1441290	38.17	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.086	232	98193	10.57	ng	95
60) Diethylphthalate	15.280	149	1358861	38.03	ng	99
61) 4-Chlorophenyl-phenyle...	15.504	204	712623	38.29	ng	98
62) 4-Nitroaniline	15.528	138	381883	41.72	ng	87
63) Azobenzene	15.792	77	1255137	34.78	ng	97
66) n-Nitrosodiphenylamine	15.716	169	1268873	33.93	ng	100
67) 4-Bromophenyl-phenylether	16.398	248	439814	35.09	ng	95
68) Hexachlorobenzene	16.510	284	521677	36.71	ng	96
69) Atrazine	16.663	200	391298	39.06	ng	95
70) Pentachlorophenol	16.857	266	24536	2.95	ng	94
71) Phenanthrene	17.245	178	2440438	36.06	ng	99
72) Anthracene	17.333	178	2444391	37.20	ng	100
73) Carbazole	17.604	167	2479591	39.57	ng	100
74) Di-n-butylphthalate	18.163	149	2501637	38.29	ng	99
75) Fluoranthene	19.251	202	3247088	42.27	ng	99
77) Benzidine	19.433	184	1264313	47.79	ng	99
78) Pyrene	19.610	202	3480416	30.72	ng	99
80) Butylbenzylphthalate	20.504	149	1262557	28.79	ng	92
81) Benzo(a)anthracene	21.345	228	3999816	36.75	ng	99
82) 3,3'-Dichlorobenzidine	21.280	252	1373366	41.68	ng	99
83) Chrysene	21.398	228	3874742	36.63	ng	98
84) Bis(2-ethylhexyl)phtha...	21.274	149	2036322	30.55	ng	99
85) Di-n-octyl phthalate	22.157	149	3617210	34.53	ng	99
87) Indeno(1,2,3-cd)pyrene	25.962	276	4815634	37.42	ng	99
88) Benzo(b)fluoranthene	22.956	252	4455521	34.79	ng	99
89) Benzo(k)fluoranthene	22.998	252	4345707	34.74	ng	98
90) Benzo(a)pyrene	23.545	252	4158256	36.72	ng	100
91) Dibenzo(a,h)anthracene	25.974	278	4098419	36.93	ng	99
92) Benzo(g,h,i)perylene	26.668	276	3888389	36.84	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 11 07:53:57 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM060921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 08 20:51:08 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.840 min Scan# 808
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 139971
Ion Ratio Lower Upper
152 100
150 154.9 125.9 188.9
115 56.5 44.6 67.0



#2
1,4-Dioxane
Concen: 34.24 ng
RT: 3.346 min Scan# 44
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45


Tgt Ion: 88 Resp: 127451
Ion Ratio Lower Upper
88 100
58 75.6 53.6 80.4
43 32.2 25.8 38.6








#3
Pyridine
Concen: 38.55 ng
RT: 3.752 min Scan# 113
Delta R.T. 0.000 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040







Tgt Ion: 79 Resp: 366312
Ion Ratio Lower Upper
79 100
52 72.0 48.2 72.2
51 37.7 27.9 41.9



#4
n-Nitrosodimethylamine
Concen: 34.64 ng
RT: 3.652 min Scan# 96
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 42 Resp: 167077
Ion Ratio Lower Upper
42 100
74 121.0 84.1 126.1
44 8.1 4.5 6.7









#5
2-Fluorophenol
Concen: 76.59 ng
RT: 5.422 min Scan# 397
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 659478
Ion Ratio Lower Upper
112 100
64 56.9 47.1 70.7
63 29.8 27.3 40.9



#6
Aniline
Concen: 37.08 ng
RT: 7.175 min Scan# 695
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 93 Resp: 527378
Ion Ratio Lower Upper
93 100
66 37.1 32.7 49.1
65 19.7 16.8 25.2









#7
Phenol-d6
Concen: 79.85 ng
RT: 7.005 min Scan# 666
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	99	Resp:	995307
Ion	Ratio	Lower	Upper
99	100		
42	18.5	18.0	27.0
71	32.0	30.9	46.3



#8
2-Chlorophenol
Concen: 39.50 ng
RT: 7.405 min Scan# 734
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:	128	Resp:	399216
Ion	Ratio	Lower	Upper
128	100		
130	31.6	13.5	53.5
64	45.3	25.7	65.7









#9
Benzaldehyde
Concen: 46.77 ng
RT: 6.987 min Scan# 663
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 243483
Ion Ratio Lower Upper
77 100
105 99.6 71.2 111.2
106 95.9 68.6 108.6



#10
Phenol
Concen: 39.07 ng
RT: 7.028 min Scan# 670
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 94 Resp: 492010
Ion Ratio Lower Upper
94 100
65 28.8 11.9 51.9
66 36.5 22.6 62.6









#11
bis(2-Chloroethyl)ether
Concen: 36.79 ng
RT: 7.269 min Scan# 711
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 365036
Ion Ratio Lower Upper
93 100
63 75.3 54.3 94.3
95 32.3 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 36.93 ng
RT: 7.734 min Scan# 790
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:146 Resp: 411143
Ion Ratio Lower Upper
146 100
148 64.0 52.2 78.4
75 28.3 25.8 38.8









#13
1,4-Dichlorobenzene
Concen: 36.93 ng
RT: 7.881 min Scan# 815
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 419059
Ion Ratio Lower Upper
146 100
148 64.0 51.8 77.8
111 38.1 31.0 46.6



#14
1,2-Dichlorobenzene
Concen: 37.31 ng
RT: 8.193 min Scan# 868
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:146 Resp: 411816
Ion Ratio Lower Upper
146 100
148 64.6 51.7 77.5
111 40.9 32.6 49.0









#15
Benzyl Alcohol
Concen: 39.32 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 343969
Ion Ratio Lower Upper
79 100
108 79.8 57.0 85.4
77 65.1 54.6 81.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 32.22 ng
RT: 8.369 min Scan# 898
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 45 Resp: 514435
Ion Ratio Lower Upper
45 100
77 14.0 0.0 38.0
79 11.0 0.0 34.1









#17
2-Methylphenol
Concen: 39.85 ng
RT: 8.275 min Scan# 882
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:107 Resp: 325185
Ion Ratio Lower Upper
107 100
108 109.3 85.0 127.6
77 45.8 41.3 61.9
79 43.8 40.7 61.1



#18
Hexachloroethane
Concen: 34.76 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45


Tgt Ion:117 Resp: 141764
Ion Ratio Lower Upper
117 100
119 96.3 76.2 114.2
201 112.7 89.8 134.6








#19
n-Nitroso-di-n-propylamine
Concen: 37.50 ng
RT: 8.646 min Scan# 945
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040







Tgt Ion	Ratio	Lower	Upper
70	100		
42	62.5	51.4	77.0
101	10.0	8.1	12.1
130	23.0	17.8	26.6



#20
3+4-Methylphenols
Concen: 42.05 ng
RT: 8.604 min Scan# 938
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion	Ratio	Lower	Upper
107	100		
108	90.9	69.9	109.9
77	33.5	17.3	57.3
79	28.2	13.3	53.3









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.634 min Scan# 1283
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:136 Resp: 607488
Ion Ratio Lower Upper
136 100
137 10.6 8.4 12.6
54 7.7 5.5 8.3
68 5.7 4.2 6.2



#22
Acetophenone
Concen: 37.93 ng
RT: 8.663 min Scan# 948
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:105 Resp: 615601
Ion Ratio Lower Upper
105 100
71 1.9 3.4 5.0#
51 26.9 22.2 33.4
120 22.0 17.6 26.4









#23
Nitrobenzene-d5
Concen: 73.70 ng
RT: 8.999 min Scan# 1005
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 919044
Ion Ratio Lower Upper
82 100
128 43.2 31.1 46.7
54 50.8 37.3 55.9



#24
Nitrobenzene
Concen: 35.60 ng
RT: 9.040 min Scan# 1012
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 77 Resp: 463126
Ion Ratio Lower Upper
77 100
123 47.1 34.0 51.0
65 13.2 11.0 16.4









#25
Isophorone
Concen: 37.04 ng
RT: 9.563 min Scan# 1101
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 885185
Ion Ratio Lower Upper
82 100
95 7.2 6.3 9.5
138 18.4 14.5 21.7



#26
2-Nitrophenol
Concen: 17.66 ng
RT: 9.746 min Scan# 1132
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 139 Resp: 88115
Ion Ratio Lower Upper
139 100
109 28.1 26.7 40.1
65 46.7 41.9 62.9









#27
2,4-Dimethylphenol
Concen: 37.70 ng
RT: 9.804 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 348738
Ion Ratio Lower Upper
122 100
107 113.4 95.0 142.6
121 58.6 48.1 72.1



#28
bis(2-Chloroethoxy)methane
Concen: 34.51 ng
RT: 10.040 min Scan# 1182
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 93 Resp: 468888
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.2
123 11.2 9.3 13.9









#29
2,4-Dichlorophenol
Concen: 39.92 ng
RT: 10.275 min Scan# 1222
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 414020
Ion Ratio Lower Upper
162 100
164 65.5 46.5 86.5
98 33.1 12.1 52.1



#30
1,2,4-Trichlorobenzene
Concen: 37.67 ng
RT: 10.493 min Scan# 1259
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:180 Resp: 445887
Ion Ratio Lower Upper
180 100
182 96.0 76.3 114.5
145 27.7 22.3 33.5









#31
Naphthalene
Concen: 36.40 ng
RT: 10.681 min Scan# 1291
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1288084
Ion Ratio Lower Upper
128 100
129 11.2 9.0 13.6
127 13.0 10.8 16.2



#32
Benzoic acid
Concen: 48.49 ng
RT: 9.804 min Scan# 1142
Delta R.T. -0.141 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:122 Resp: 348738
Ion Ratio Lower Upper
122 100
105 3.9 107.3 147.3#
77 34.9 88.0 128.0#









#33
4-Chloroaniline
Concen: 38.51 ng
RT: 10.793 min Scan# 1310
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	554544
Ion Ratio	Lower	Upper	
127	100		
129	31.9	26.8	40.2
65	27.6	25.2	37.8
92	17.2	14.8	22.2



#34
Hexachlorobutadiene
Concen: 37.60 ng
RT: 10.969 min Scan# 1340
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:	225	Resp:	265661
Ion Ratio	Lower	Upper	
225	100		
223	62.8	52.1	78.1
227	63.5	53.8	80.6









#35
Caprolactam
Concen: 48.63 ng
RT: 11.569 min Scan# 1442
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 174045
Ion Ratio Lower Upper
113 100
55 161.0 132.0 172.0
56 118.0 96.3 136.3



#36
4-Chloro-3-methylphenol
Concen: 41.13 ng
RT: 11.910 min Scan# 1500
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:107 Resp: 446749
Ion Ratio Lower Upper
107 100
144 28.1 22.3 33.5
142 87.3 69.3 103.9









#37
2-Methylnaphthalene
Concen: 38.43 ng
RT: 12.287 min Scan# 1564
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 997751
Ion Ratio Lower Upper
142 100
141 88.2 71.7 107.5
115 33.9 29.1 43.7



#38
1-Methylnaphthalene
Concen: 38.04 ng
RT: 12.510 min Scan# 1602
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:142 Resp: 940829
Ion Ratio Lower Upper
142 100
141 92.4 74.2 111.4
116 3.7 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.463 min Scan# 1934
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 449136
Ion Ratio Lower Upper
164 100
162 101.2 83.4 125.0
160 45.0 37.2 55.8



#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.58 ng
RT: 12.657 min Scan# 1627
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:216 Resp: 544100
Ion Ratio Lower Upper
216 100
214 78.2 63.0 94.4
179 20.1 17.6 26.4
108 16.1 14.9 22.3









#41
Hexachlorocyclopentadiene
Concen: 22.20 ng
RT: 12.639 min Scan# 1624
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:237	Resp: 176256
Ion Ratio	Lower Upper
237	100
235	61.8 42.4 82.4
272	12.5 0.0 31.7



#42
2,4,6-Tribromophenol
Concen: 29.86 ng
RT: 15.951 min Scan# 2187
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:330	Resp: 168595
Ion Ratio	Lower Upper
330	100
332	96.8 77.0 115.4
141	34.5 30.2 45.2









#43
2,4,6-Trichlorophenol
Concen: 21.88 ng
RT: 12.898 min Scan# 1668
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 214146
Ion Ratio Lower Upper
196 100
198 94.1 77.6 116.4
200 31.4 24.6 37.0



#44
2,4,5-Trichlorophenol
Concen: 31.17 ng
RT: 12.969 min Scan# 1680
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:196 Resp: 334672
Ion Ratio Lower Upper
196 100
198 94.2 77.0 115.4
97 43.1 36.6 55.0
132 23.7 18.2 27.2









#45
2-Fluorobiphenyl
Concen: 71.74 ng
RT: 13.086 min Scan# 1700
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2450225
Ion Ratio Lower Upper
172 100
171 35.6 28.6 43.0
170 23.4 18.4 27.6



#46
1,1'-Biphenyl
Concen: 35.74 ng
RT: 13.298 min Scan# 1736
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:154 Resp: 1299810
Ion Ratio Lower Upper
154 100
153 39.8 20.7 60.7
76 11.7 0.0 31.0









#47
2-Chloronaphthalene
Concen: 35.75 ng
RT: 13.339 min Scan# 1743
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 1050911
Ion Ratio Lower Upper
162 100
127 35.6 27.1 40.7
164 33.1 26.1 39.1



#48
2-Nitroaniline
Concen: 36.67 ng
RT: 13.545 min Scan# 1778
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 65 Resp: 319240
Ion Ratio Lower Upper
65 100
92 70.9 55.5 83.3
138 113.0 85.7 128.5









#49
Acenaphthylene
Concen: 36.53 ng
RT: 14.186 min Scan# 1887
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1645283
Ion Ratio Lower Upper
152 100
151 19.6 16.3 24.5
153 12.9 10.2 15.4



#50
Dimethylphthalate
Concen: 37.75 ng
RT: 13.922 min Scan# 1842
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:163 Resp: 1349218
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 10.3 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 38.11 ng
RT: 14.039 min Scan# 1862
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 300269
Ion Ratio Lower Upper
165 100
63 50.6 45.3 67.9
89 53.0 46.3 69.5



#52
Acenaphthene
Concen: 35.51 ng
RT: 14.528 min Scan# 1945
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:154 Resp: 1021566
Ion Ratio Lower Upper
154 100
153 113.5 92.2 138.2
152 53.5 43.4 65.2









#53
3-Nitroaniline
Concen: 37.93 ng
RT: 14.369 min Scan# 1918
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 325553
Ion Ratio Lower Upper
138 100
108 9.2 7.7 11.5
92 109.2 92.6 138.8



#55
Dibenzofuran
Concen: 37.43 ng
RT: 14.863 min Scan# 2002
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:168 Resp: 1713264
Ion Ratio Lower Upper
168 100
139 35.0 29.4 44.0
169 13.3 10.8 16.2









#56
4-Nitrophenol
Concen: 3.71 ng
RT: 14.675 min Scan# 1970
Delta R.T. 0.001 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:139 Resp: 24841
Ion Ratio Lower Upper
139 100
109 78.9 64.5 104.5
65 96.0 78.9 118.9



#57
2,4-Dinitrotoluene
Concen: 33.75 ng
RT: 14.822 min Scan# 1995
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:165 Resp: 365393
Ion Ratio Lower Upper
165 100
63 40.2 35.0 52.4
89 67.7 55.2 82.8
182 2.7 2.6 3.8









#58
Fluorene
Concen: 38.17 ng
RT: 15.510 min Scan# 2112
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:166 Resp: 1441290
Ion Ratio Lower Upper
166 100
165 96.3 79.0 118.6
167 13.4 10.8 16.2



#59
2,3,4,6-Tetrachlorophenol
Concen: 10.57 ng
RT: 15.086 min Scan# 2040
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:232 Resp: 98193
Ion Ratio Lower Upper
232 100
131 41.7 34.4 51.6
130 2.2 2.2 3.2
166 27.1 26.2 39.4









#60
Diethylphthalate
Concen: 38.03 ng
RT: 15.280 min Scan# 2073
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 1358861
Ion Ratio Lower Upper
149 100
177 22.5 17.9 26.9
150 12.4 10.3 15.5



#61
4-Chlorophenyl-phenylether
Concen: 38.29 ng
RT: 15.504 min Scan# 2111
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:204 Resp: 712623
Ion Ratio Lower Upper
204 100
206 32.5 25.5 38.3
141 52.6 40.3 60.5









#62
4-Nitroaniline
Concen: 41.72 ng
RT: 15.528 min Scan# 2115
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 381883
Ion Ratio Lower Upper
138 100
92 49.4 33.4 73.4
108 78.6 75.5 115.5



#63
Azobenzene
Concen: 34.78 ng
RT: 15.792 min Scan# 2160
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion: 77 Resp: 1255137
Ion Ratio Lower Upper
77 100
182 27.3 6.8 46.8
105 19.6 0.0 36.6
51 29.8 11.5 51.5









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.198 min Scan# 2399
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:188 Resp: 1075503
Ion Ratio Lower Upper
188 100
94 8.4 6.0 9.0
80 9.1 6.6 9.8



#66
n-Nitrosodiphenylamine
Concen: 33.93 ng
RT: 15.716 min Scan# 2147
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:169 Resp: 1268873
Ion Ratio Lower Upper
169 100
168 67.3 54.2 81.2
167 36.1 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 35.09 ng
RT: 16.398 min Scan# 2263
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 439814
Ion Ratio Lower Upper
248 100
250 97.3 77.9 116.9
141 60.5 56.2 84.2



#68
Hexachlorobenzene
Concen: 36.71 ng
RT: 16.510 min Scan# 2282
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:284 Resp: 521677
Ion Ratio Lower Upper
284 100
142 33.5 29.0 43.4
249 30.8 26.0 39.0









#69
Atrazine
Concen: 39.06 ng
RT: 16.663 min Scan# 2308
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 391298
Ion Ratio Lower Upper
200 100
173 25.2 3.8 43.8
215 48.7 24.8 64.8



#70
Pentachlorophenol
Concen: 2.95 ng
RT: 16.857 min Scan# 2341
Delta R.T. 0.001 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:266 Resp: 24536
Ion Ratio Lower Upper
266 100
268 62.9 52.8 79.2
264 58.4 51.9 77.9









#71
Phenanthrene
Concen: 36.06 ng
RT: 17.245 min Scan# 2407
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 2440438
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.0
179 15.6 12.2 18.4



#72
Anthracene
Concen: 37.20 ng
RT: 17.333 min Scan# 2422
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:178 Resp: 2444391
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.5 12.5 18.7









#73
 Carbazole
 Concen: 39.57 ng
 RT: 17.604 min Scan# 2468
 Delta R.T. -0.005 min
 Lab File: BM030385.D
 Acq: 11 Jun 2021 00:45




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:167 Resp: 2479591
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.0
 139 12.2 9.8 14.6



#74
 Di-n-butylphthalate
 Concen: 38.29 ng
 RT: 18.163 min Scan# 2563
 Delta R.T. -0.005 min
 Lab File: BM030385.D
 Acq: 11 Jun 2021 00:45

Tgt Ion:149 Resp: 2501637
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.2 10.8
 104 4.9 4.3 6.5









#75
Fluoranthene
Concen: 42.27 ng
RT: 19.251 min Scan# 2748
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 3247088
Ion Ratio Lower Upper
202 100
101 9.7 0.0 28.6
203 17.6 0.0 37.3



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.362 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:240 Resp: 1495352
Ion Ratio Lower Upper
240 100
120 9.1 7.0 10.6
236 24.9 21.0 31.4









#77
Benzidine
Concen: 47.79 ng
RT: 19.433 min Scan# 2779
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 1264313
Ion Ratio Lower Upper
184 100
185 14.1 10.9 16.3
183 11.9 9.8 14.8



#78
Pyrene
Concen: 30.72 ng
RT: 19.610 min Scan# 2809
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:202 Resp: 3480416
Ion Ratio Lower Upper
202 100
200 21.7 17.3 25.9
203 17.9 13.9 20.9









#79
 Terphenyl-d14
 Concen: 62.55 ng
 RT: 19.815 min Scan# 2844
 Delta R.T. -0.006 min
 Lab File: BM030385.D
 Acq: 11 Jun 2021 00:45




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040


Tgt Ion:244 Resp: 5325052
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.3 9.5
 122 9.6 6.9 10.3



#80
 Butylbenzylphthalate
 Concen: 28.79 ng
 RT: 20.504 min Scan# 2961
 Delta R.T. -0.005 min
 Lab File: BM030385.D
 Acq: 11 Jun 2021 00:45

Tgt Ion:149 Resp: 1262557
 Ion Ratio Lower Upper
 149 100
 91 64.7 57.4 86.0
 206 23.5 16.6 25.0









#81
Benzo(a)anthracene
Concen: 36.75 ng
RT: 21.345 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3999816
Ion Ratio Lower Upper
228 100
226 28.3 22.3 33.5
229 20.3 15.7 23.5



#82
3,3'-Dichlorobenzidine
Concen: 41.68 ng
RT: 21.280 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:252 Resp: 1373366
Ion Ratio Lower Upper
252 100
254 64.5 51.5 77.3
126 9.3 8.6 13.0









#83
Chrysene
Concen: 36.63 ng
RT: 21.398 min Scan# 3113
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 3874742
Ion Ratio Lower Upper
228 100
226 31.5 24.6 36.8
229 20.4 15.5 23.3



#84
Bis(2-ethylhexyl)phthalate
Concen: 30.55 ng
RT: 21.274 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:149 Resp: 2036322
Ion Ratio Lower Upper
149 100
167 27.7 22.6 33.8
279 4.8 3.8 5.8









#85
Di-n-octyl phthalate
Concen: 34.53 ng
RT: 22.157 min Scan# 3242
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 3617210
Ion Ratio Lower Upper
149 100
167 1.7 1.3 1.9
43 8.7 7.4 11.0



#86
Perylene-d12
Concen: 20.00 ng
RT: 23.639 min Scan# 3494
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:264 Resp: 1692151
Ion Ratio Lower Upper
264 100
260 23.5 19.8 29.8
265 22.0 17.7 26.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 37.42 ng
RT: 25.962 min Scan# 3889
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 4815634
Ion Ratio Lower Upper
276 100
138 22.3 17.1 25.7
277 25.2 20.3 30.5



#88
Benzo(b)fluoranthene
Concen: 34.79 ng
RT: 22.956 min Scan# 3378
Delta R.T. -0.005 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:252 Resp: 4455521
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 8.6 6.8 10.2









#89
Benzo(k)fluoranthene
Concen: 34.74 ng
RT: 22.998 min Scan# 3385
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 4345707
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.2
125 9.1 6.2 9.2



#90
Benzo(a)pyrene
Concen: 36.72 ng
RT: 23.545 min Scan# 3478
Delta R.T. -0.006 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:252 Resp: 4158256
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 9.1 7.3 10.9









#91
Dibenzo(a,h)anthracene
Concen: 36.93 ng
RT: 25.974 min Scan# 3891
Delta R.T. -0.011 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 4098419
Ion Ratio Lower Upper
278 100
139 14.2 11.0 16.4
279 24.0 18.7 28.1



#92
Benzo(g,h,i)perylene
Concen: 36.84 ng
RT: 26.668 min Scan# 4009
Delta R.T. -0.017 min
Lab File: BM030385.D
Acq: 11 Jun 2021 00:45

Tgt Ion:276 Resp: 3888389
Ion Ratio Lower Upper
276 100
277 23.9 19.4 29.2
138 19.2 14.4 21.6

