

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061024\
 Data File : BM045943.D
 Acq On : 10 Jun 2024 10:27
 Operator : MA/JU
 Sample : PB161249BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK249

Quant Time: Jun 10 23:17:33 2024
 Quant Title :
 QLast Update : Mon Jun 10 22:47:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.459	152	4126	0.400	ng/ul	-0.03
4) Naphthalene-d8	10.226	136	13109	0.400	ng/ul	#-0.03
9) Acenaphthene-d10	14.108	164	6821	0.400	ng/ul	-0.01
13) Phenanthrene-d10	16.849	188	14590m	0.400	ng/ul	-0.02
17) Chrysene-d12	21.053	240	10513m	0.400	ng/ul	0.00
23) Perylene-d12	23.157	264	10988m	0.400	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.050	96	27615	5.876	ng/ul	-0.03
6) 2-Methylnaphthalene-d10	11.848	152	4676	0.248	ng/ul	0.00
18) Fluoranthene-d10	18.885	212	9318	0.296	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061024\
Data File : BM045943.D
Acq On : 10 Jun 2024 10:27
Operator : MA/JU
Sample : PB161249BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK249

Quant Time: Jun 10 23:17:33 2024
Quant Title :
QLast Update : Mon Jun 10 22:47:12 2024
Response via : Initial Calibration

