

Data Path : S:\HPCHEM1\BNA_M\DATA\BM061115\
 Data File : BM001645.D
 Acq On : 11 Jun 2015 12:59
 Operator : TP/IZ
 Sample : G2414-08
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 X1T62

Quant Time: Jun 11 15:28:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM061115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 11 07:52:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	1612	0.40	ng/ul	0.00
4) Naphthalene-d8	10.48	136	6033	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.34	164	2737	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.09	188	5431	0.40	ng/ul	0.00
18) Chrysene-d12	21.27	240	4479	0.40	ng/ul	0.00
22) Perylene-d12	23.52	264	3995	0.40	ng/ul	0.01
System Monitoring Compounds						
2) 1,4-Dioxane-d8	0.00	96	0d	0.00	ng/uL	
6) 2-Methylnaphthalene-d10	12.08	152	1944	0.27	ng/ul	0.00
16) Fluoranthene-d10	19.12	212	3864	0.35	ng/ul	0.00
Target Compounds				Qvalue		
5) Naphthalene	10.53	128	8232	0.52	ng/ul	99
7) 2-Methylnaphthalene	12.14	142	5036	0.51	ng/ul	99
11) Fluorene	15.38	166	6966	0.66	ng/ul	97
13) Pentachlorophenol	16.73	266	2105	1.94	ng/ul	95
14) Phenanthrene	17.12	178	11567	0.74	ng/ul	99
17) Fluoranthene	19.14	202	10905	0.65	ng/ul	98
20) Benzo(a)anthracene	21.26	228	11277	0.78	ng/ul	91
24) Benzo(k)fluoranthene	22.88	252	10855	0.70	ng/ul#	92
26) Indeno(1,2,3-cd)pyrene	25.77	276	8505	0.52	ng/ul	96
28) Benzo(g,h,i)perylene	26.45	276	9511	0.63	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : S:\HPCHEM1\BNA_M\DATA\BM061115\
Data File : BM001645.D
Acq On : 11 Jun 2015 12:59
Operator : TP/IZ
Sample : G2414-08
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
X1T62

Quant Time: Jun 11 15:28:49 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM061115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 11 07:52:35 2015
Response via : Initial Calibration

