


Data Path : S:\HPCHEM1\BNA_M\DATA\BM061115\
Data File : BM001641.D
Acq On : 11 Jun 2015 09:34
Operator : TP/IZ
Sample : DFTPP58
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP58

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-SIM-BM061115.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jun 11 07:52:35 2015



AutoFind: Scans 2522, 2523, 2524; Background Corrected with Scan 2516

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	32416	PASS
68	69	0.00	2	1.9	624	PASS
69	198	0.00	100	35.0	33248	PASS
70	69	0.00	2	0.3	112	PASS
127	198	10	80	41.0	38898	PASS
197	198	0.00	1	0.6	604	PASS
198	198	100	100	100.0	94976	PASS
199	198	5	9	7.3	6905	PASS
275	198	10	60	28.5	27032	PASS
365	198	1	100	4.5	4313	PASS
441	443	0.01	100	80.2	13052	PASS
442	198	50	100	91.9	87301	PASS
443	442	15	24	18.6	16266	PASS

m/z	Abundance
38.00	477.0
39.10	2529.0
40.00	548.0
44.00	1306.0
49.10	395.0
50.10	10263.0
51.10	38200.0
52.10	2038.0
56.00	1161.0
57.00	2801.0
60.90	441.0
62.00	579.0
63.00	1769.0
65.00	608.0
68.10	747.0
69.00	39288.0
70.10	338.0
72.90	607.0
74.10	4493.0
75.00	6960.0
76.00	2394.0
77.10	41240.0
78.10	2958.0
79.00	3551.0
80.00	2545.0
81.00	3804.0
82.00	811.0
83.00	605.0
84.90	621.0
86.00	1200.0
86.90	470.0
91.00	1116.0
92.00	972.0
93.00	5986.0
93.90	307.0
98.00	5065.0
99.00	3935.0
99.90	389.0
101.00	1753.0
103.00	530.0
104.00	1325.0
105.00	1387.0
105.90	386.0
107.00	15468.0
108.00	2073.0
109.20	405.0
110.00	23096.0
111.00	4442.0
112.10	517.0
116.00	1083.0
117.00	14320.0
118.00	1108.0
122.00	1330.0
123.00	1654.0
123.90	863.0
124.90	659.0
127.00	42840.0
128.00	3248.0
129.00	19464.0
129.90	1814.0
131.00	410.0
134.00	609.0
135.00	1684.0
136.00	706.0
137.10	1072.0
140.00	348.0
141.00	3217.0
141.90	486.0
142.90	731.0
145.90	631.0
147.00	1535.0
148.00	3673.0
149.00	742.0
151.20	384.0
152.90	869.0
154.00	582.0
155.00	1444.0
156.00	2228.0
157.00	454.0
158.00	595.0
158.90	411.0
160.00	1159.0
161.00	1234.0
161.90	525.0
164.90	917.0
165.90	639.0
167.00	4879.0
168.00	2976.0
169.00	356.0
170.80	307.0
172.00	518.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

173.00	675.0
174.00	1295.0
175.00	2084.0
176.10	608.0
176.90	1122.0
177.90	323.0
179.00	4041.0
180.00	2638.0
181.00	1325.0
184.00	389.0
185.00	1842.0
186.00	15941.0
187.00	3959.0
188.00	645.0
189.00	1070.0
190.90	422.0
192.00	1495.0
193.10	1488.0
194.10	366.0
195.90	2741.0
196.80	630.0
197.10	592.0
198.00	97128.0
198.90	7307.0
200.00	637.0
201.60	380.0
203.00	979.0
204.00	3945.0
205.00	7462.0
206.00	27984.0
207.00	4645.0
208.00	1302.0
208.90	380.0
210.00	437.0
211.00	983.0
215.00	404.0
216.00	726.0
216.90	8277.0
217.90	1019.0
221.10	3621.0
221.80	898.0
223.00	2065.0
224.00	15578.0
225.10	4538.0
226.10	583.0
227.00	8194.0
228.10	1042.0
229.00	1709.0
230.90	380.0
233.90	440.0
234.90	377.0
235.90	309.0
237.00	946.0
239.00	406.0
240.90	351.0
241.90	1109.0
243.00	969.0
244.00	11487.0
245.10	1570.0
246.00	2499.0
247.00	713.0
248.90	359.0
253.00	419.0
253.90	417.0
255.00	61816.0
256.00	8710.0
256.90	854.0
258.00	4224.0
258.90	587.0
264.90	1659.0
273.00	1451.0
274.00	4669.0
275.00	26616.0
276.10	3893.0
277.00	2866.0
277.90	440.0
283.00	326.0
285.00	679.0
293.10	682.0
296.00	9207.0
297.00	1443.0
303.00	843.0
315.10	804.0
316.10	552.0
323.00	2651.0
324.00	486.0
326.90	304.0
334.00	1598.0
335.00	439.0
345.90	470.0
352.00	941.0
353.10	502.0
354.10	782.0
365.00	3356.0
365.90	520.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

372.00	1278.0
373.10	427.0
402.10	549.0
403.00	705.0
421.00	705.0
422.00	737.0
423.10	3915.0
424.00	1324.0
425.00	308.0
441.00	10097.0
442.10	68704.0
443.10	12900.0
444.00	1244.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	573.0
39.10	2902.0
40.00	506.0
44.00	1011.0
50.10	10273.0
51.00	38368.0
52.10	1856.0
56.00	1060.0
57.00	2552.0
61.00	621.0
61.90	529.0
63.00	1728.0
65.00	798.0
68.00	713.0
69.00	40144.0
73.00	799.0
74.00	4402.0
75.00	6916.0
76.10	2105.0
77.10	40464.0
78.00	2520.0
79.00	3754.0
80.00	2780.0
81.00	3566.0
82.10	937.0
82.90	566.0
85.00	588.0
85.90	1130.0
87.10	417.0
91.00	779.0
91.90	1168.0
93.00	6427.0
98.00	5555.0
99.00	3800.0
101.00	2372.0
103.00	764.0
103.90	1288.0
105.00	1306.0
105.90	398.0
107.00	14946.0
108.00	2407.0
109.00	506.0
110.00	24976.0
111.00	4141.0
112.00	627.0
116.10	1041.0
117.00	15125.0
118.00	986.0
122.00	1119.0
123.00	1908.0
123.90	592.0
125.00	612.0
127.00	47160.0
128.00	3462.0
129.00	19760.0
130.00	1792.0
133.80	807.0
135.00	2011.0
135.90	714.0
137.00	1057.0
141.00	2721.0
141.90	875.0
143.00	485.0
146.00	532.0
147.00	1704.0
148.00	3418.0
148.90	773.0
151.00	486.0
152.90	903.0
154.00	752.0
154.90	1719.0
156.00	2446.0
157.10	406.0
157.90	556.0
158.90	340.0
159.90	876.0
160.90	1113.0
165.00	1269.0
166.10	761.0
167.00	6265.0
168.00	3520.0
168.90	525.0
170.90	380.0
171.90	602.0
173.00	651.0
174.00	1285.0
175.00	2371.0
176.10	644.0
177.00	1359.0
177.90	343.0
179.00	5542.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

180.00	2995.0
181.00	1515.0
183.90	377.0
185.00	2307.0
186.00	17384.0
187.00	5081.0
187.90	900.0
189.00	1276.0
190.90	598.0
192.00	1780.0
193.00	1494.0
194.00	406.0
195.90	2776.0
197.10	763.0
198.00	115792.0
198.90	8336.0
200.00	786.0
201.30	414.0
201.60	439.0
203.00	862.0
204.00	4602.0
205.00	7805.0
206.00	32688.0
207.00	5181.0
208.00	1564.0
209.00	699.0
210.00	631.0
211.00	1453.0
215.10	420.0
216.00	919.0
217.00	10034.0
217.90	1151.0
221.00	4233.0
221.80	1087.0
222.90	1973.0
224.00	18800.0
225.00	4947.0
226.00	619.0
227.00	9800.0
228.10	1248.0
229.00	2235.0
230.90	627.0
233.90	545.0
234.90	758.0
236.00	482.0
237.00	907.0
239.00	386.0
241.00	412.0
242.00	981.0
242.90	1133.0
244.00	13803.0
245.10	2415.0
246.00	3288.0
247.00	703.0
249.10	598.0
253.00	373.0
254.20	645.0
255.00	77528.0
256.00	11926.0
257.00	930.0
258.00	5780.0
259.00	832.0
265.00	2032.0
272.20	309.0
273.00	2169.0
274.00	6282.0
275.00	31328.0
276.00	4869.0
277.00	3422.0
277.90	479.0
282.90	346.0
285.00	507.0
292.90	589.0
296.00	11363.0
297.00	1432.0
303.00	1279.0
314.00	487.0
315.00	1278.0
315.90	939.0
320.80	403.0
323.00	3321.0
324.00	430.0
326.90	751.0
327.90	337.0
332.00	392.0
333.00	330.0
334.10	2118.0
335.00	546.0
341.10	348.0
345.90	573.0
352.00	770.0
353.00	753.0
354.00	1075.0
365.00	5200.0
366.00	703.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

372.00	1828.0
373.00	418.0
383.10	447.0
402.00	668.0
403.00	864.0
404.10	414.0
421.00	911.0
422.00	837.0
423.00	5431.0
424.00	1808.0
441.10	15691.0
442.10	103776.0
443.10	19536.0
444.00	1922.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
39.10	1609.0
40.00	590.0
44.00	1175.0
50.10	5965.0
51.10	20680.0
52.00	1084.0
56.00	678.0
57.00	1213.0
62.10	314.0
63.10	872.0
65.00	446.0
67.90	414.0
69.00	20312.0
72.90	373.0
74.00	2505.0
75.00	3824.0
76.10	1279.0
77.10	23048.0
78.00	1321.0
79.00	2173.0
80.00	1495.0
81.00	1903.0
82.00	418.0
86.00	650.0
91.10	386.0
91.90	514.0
93.00	3631.0
94.00	363.0
98.00	3345.0
99.00	1911.0
101.00	1140.0
103.00	495.0
104.00	664.0
104.90	665.0
107.00	9388.0
108.00	1286.0
110.00	14033.0
111.00	2474.0
116.10	486.0
117.00	9257.0
117.90	444.0
122.00	544.0
122.90	1184.0
123.90	509.0
124.90	511.0
127.00	26696.0
128.00	2048.0
129.00	12041.0
130.00	1006.0
134.00	534.0
135.00	994.0
135.90	364.0
136.90	483.0
141.00	1878.0
141.90	592.0
143.00	354.0
146.00	395.0
147.00	1111.0
148.00	2360.0
148.90	493.0
153.00	638.0
154.00	413.0
155.00	920.0
156.00	1294.0
157.20	347.0
157.90	443.0
159.90	577.0
160.90	826.0
164.90	676.0
166.00	790.0
167.00	3166.0
168.00	1868.0
168.90	430.0
173.10	320.0
173.90	1039.0
175.00	1479.0
177.00	603.0
178.00	307.0
179.00	2668.0
180.10	1984.0
181.00	846.0
185.10	1208.0
186.00	10633.0
187.00	3016.0
187.90	382.0
188.90	767.0
190.90	428.0
192.00	998.0
193.00	1248.0
196.00	1940.0
197.10	459.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

198.00	72008.0
198.90	5074.0
200.00	377.0
201.60	409.0
202.80	665.0
204.00	3317.0
205.00	5310.0
206.00	20728.0
207.10	3523.0
208.00	741.0
210.00	460.0
211.00	816.0
216.00	702.0
217.00	6831.0
218.00	652.0
221.00	2807.0
221.90	709.0
222.90	1410.0
224.00	12135.0
225.10	3028.0
226.00	402.0
227.00	6198.0
228.00	834.0
228.90	1133.0
231.00	630.0
234.00	459.0
235.00	552.0
236.00	356.0
237.00	594.0
242.00	539.0
243.00	771.0
244.00	8613.0
245.00	1228.0
246.00	2500.0
247.00	359.0
249.00	433.0
253.00	310.0
253.90	401.0
255.00	53696.0
256.00	7415.0
257.00	479.0
258.00	3555.0
259.00	591.0
265.10	1618.0
272.90	1533.0
274.00	4171.0
275.00	23152.0
276.00	2927.0
277.00	2620.0
278.00	384.0
285.00	474.0
292.90	521.0
296.00	7536.0
297.10	1024.0
303.00	902.0
314.10	355.0
315.00	969.0
316.00	647.0
323.10	2498.0
323.90	419.0
326.90	306.0
334.10	1589.0
335.00	543.0
345.90	344.0
351.90	669.0
353.10	525.0
354.10	897.0
365.00	4385.0
366.00	435.0
372.10	1384.0
401.90	556.0
402.90	570.0
404.00	307.0
421.10	727.0
422.10	868.0
423.00	4699.0
424.10	1417.0
441.10	13370.0
442.10	89424.0
443.10	16363.0
444.00	1636.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

Scan 2516 (17.486 min): BM001641.D
DFTPP58

m/z	Abundance
39.90	488.0
44.00	1244.0
207.00	597.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58