

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061116\
 Data File : BM005739.D
 Acq On : 12 Jun 2016 02:14
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02054

Quant Time: Jun 12 02:45:08 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Jun 11 23:08:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2 S	1,4-Dioxane-d8	0.461	0.368	20.2#	112	0.00
3 S	Phenol-d5	1.726	1.555	9.9	122	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	1.037	0.953	8.1	122	0.00
5 S	2-Chlorophenol-d4	1.395	1.341	3.9	128	0.00
6 S	4-Methylphenol-d8	1.458	1.339	8.2	123	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.144	0.145	-0.7	122	0.00
9 S	2-Nitrophenol-d4	0.163	0.164	-0.6	119	0.00
10 S	2,4-Dichlorophenol-d3	0.296	0.299	-1.0	124	0.00
11	Naphthalene	1.002	0.994	0.8	124	0.00
12 S	4-Chloroaniline-d4	0.357	0.387	-8.4	112	0.00
13	2-Methylnaphthalene	0.754	0.717	4.9	118	0.00
14	1-Methylnaphthalene	0.745	0.714	4.2	119	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
16 S	Dimethylphthalate-d6	1.650	1.524	7.6	103	0.00
17 S	Acenaphthylene-d8	2.018	2.016	0.1	111	0.00
18	Acenaphthylene	2.066	2.033	1.6	110	0.00
19 C	Acenaphthene	1.344	1.330	1.0	111	0.00
20 S	4-Nitrophenol-d4	0.241	0.187	22.4#	86	0.00
21 S	Fluorene-d10	1.397	1.337	4.3	106	0.00
22	Fluorene	1.534	1.456	5.1	104	0.00
23 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
24 S	4,6-Dinitro-2-methylphenol-	0.103	0.041	60.2#	38	0.00
25	Phenanthrene	1.095	1.084	1.0	91	0.00
26 S	Anthracene-d10	0.942	0.924	1.9	90	0.00
27	Anthracene	1.111	1.095	1.4	89	0.00
28 C	Fluoranthene	1.286	0.999	22.3	71	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
30 S	Pyrene-d10	0.961	1.052	-9.5	68	0.00
31	Pyrene	1.217	1.341	-10.2	67	0.00
32	Benzo(a)anthracene	1.155	1.145	0.9	60	0.00
33	Chrysene	1.139	1.099	3.5	58	0.00
34 I	Perylene-d12	1.000	1.000	0.0	67	0.00
35	Benzo(b)fluoranthene	1.188	1.150	3.2	66	0.00
36	Benzo(k)fluoranthene	1.201	1.084	9.7	59	0.00
37 S	Benzo(a)pyrene-d12	0.912	0.898	1.5	66	0.00
38 C	Benzo(a)pyrene	1.128	1.109	1.7	66	0.00
39	Indeno(1,2,3-cd)pyrene	1.101	1.307	-18.7	79	0.00
40	Dibenzo(a,h)anthracene	0.921	1.096	-19.0	80	0.00
41	Benzo(g,h,i)perylene	0.921	1.112	-20.7#	83	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM061116\
Data File : BM005739.D
Acq On : 12 Jun 2016 02:14
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02054

Quant Time: Jun 12 02:45:08 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM061016.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Jun 11 23:08:56 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0